

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: DESL

RQ: 1023-07-2449

Project: DR Lake Algae

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 05/23/2023

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	K. Halbert	8-1-23	757	23.2	761.0	8.5	-	100.1	-	1.011	(P) F	(L) F
ICV	K. Halbert	8-1-23	806	22.9	761.1	8.6	8.61	100.1	-	1.011	(P) F	(L) F
CCV	K. Villings	↓	1442	24.4	760.7	8.4	8.36	100.0	-	1.011	(P) F	(L) F
CCV											P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard umhos/cm	Meter Reading umhos/cm	Pass / Fail	Lab / Field
Calibr.	K. Halbert	8-1-23	808	230131B	02/2024	1000	1000	(P) F	(L) F
ICV	K. Halbert	8-1-23	810	230131A	02/2024	100	99.7	(P) F	(L) F
CCV	K. Villings	↓	1444	230131D	2/24	15000	15091	(P) F	(L) F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	K. Halbert	8-1-23	812	M241-15	9/1/24	7.01	23.6	7.01	-34.4	(P) F	(L) F
Calibr.	K. Halbert	8-1-23	813	M221-22	8/10/24	4.00	23.7	4.00	142.5	(P) F	(L) F
Calibr.	K. Halbert	8-1-23	814	M203-05	7/27/24	10.02	23.6	10.02	-203.6	(P) F	(L) F
ICV	K. Halbert	8-1-23	815	M221-22	8/10/24	4.00	23.7	4.00	142.3	(P) F	(L) F
CCV	K. Villings	↓	1446	M203-05	7/27/24	10.01	24.0	10.03	-205.9	(P) F	(L) F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 169.2, slope from 4 to 7 176.9 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 05/30/2023

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: —; Date of last verification: —

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	K. Halbert	8-1-23	818	0.000	0.000	(P) F	(L) F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:



Cyanobacteria Bloom Response Field Data Sheet (2023-06-12 v1)

Email scanned submittal forms to lab.receiving@FloridaDEP.gov

RQ - 2023-07-24-49				Collected By:		Katrin Villinger, Katherine Halbert					
Customer:		NE-DIST		Field Report Prepared By:		KATHERINE HALBERT					
Project ID:		BLOOM_RESP		Sampling Agency:		NEROC					
Location: Swimming Pen Creek - Whiteys Fish Camp				Comments: Odor present at sample site location.							
Field ID: NEHAB0074				Coordinates:		30.10095, -81.74671					
WIN Secondary Resource Type: River/Stream				Collection Equip:		WaterBottle					
WIN ID: NEHAB0074				High Tide Time:		1213		Low Tide Time:		0730	
Matrix:		Fresh		Grab		Tide:		☒ Rising ☐ Falling ☐ Slack ☐ N/A			
Water Quality											
Date	Time ☒ EST ☐ CEN	D.O. (%)	D.O. (mg/L)	Temp (°C)	pH (SU)	Sample Depth (m)	Total Depth (m)	Secchi Disk (m)	Sp. Cond. (µmho/cm)	Salinity (PPTH)	
08/01/2023	11:15	48.3	3.6	30.1	7.15	0.31	1.9	1.1	4467	2.36	
	11:11	41.5	3.11	29.7	7.01	1.4			4684	2.48	
Bloom Observations											
Algal Bloom Observed? ☒ Yes ☐ No											
Water Clarity		☐ Clear		☒ Slightly Turbid		☐ Turbid		☐ Opaque			
Bloom Color		☒ Green		☐ Red		☐ Brown		☐ Other:			
Algae Type		☒ Planktonic		☐ Filamentous		☐ Other:					
Algae Accumulation		☐ Mat		☒ Specks		☐ Scum		☒ Glob			
Algae Distribution		☒ Surface-Entire		☐ Surf-Localized		☐ Surf-Windblown		☒ Suspended		☐ Benthic ☐ Epiphytic	
Parameter Suite		Parameter Methods				Preservation		pH Check	#Bottles Sent to Lab	Bottle Group	
Algae (PT-50mL)		☐ ALGAL-ID ☒ BLOOM-ID				☒ Ice			1	A	
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	C	
Toxins (P-125/BG-250)		☒ W-MCYST-AA ☒ W-SAXTN-MS ☐ OV-TOXN				☒ Ice			1	C	
Physical/Aggregate		☒ W-COLOR				☒ Ice			1	C	
Preserved Nutrients (500mL HDPE)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-S-A-TP				☒ Ice ☒ H2SO4		☒ <2	1	C	
		Acid Lot:		SA3013020		Exp:		01/23/2024			
Filtered Nutrients (P125mL)		☒ W-PO4-F		☒ Field Filtered 0.45um Filter		☒ Ice			1	C	
		Filter Lot #		0000175533							
Quantitative Algae		☐ DTY-QN-C / PKD-QN-C				☐ Ice ☐ Lugols					
Relinquished By:		Date/Time		Number of Coolers Shipped		Received By:		Date/Time			
KATHERINE HALBERT		08/01/2023 16:30		2							



Cyanobacteria Bloom Response Field Data Sheet (2023-06-12 v1)

Email scanned submittal forms to lab.receiving@FloridaDEP.gov

RQ - 2023-07-24-49				Collected By:		Katrin Villinger, Katherine Halbert					
Customer:		NE-DIST		Field Report Prepared By:		Katherine Halbert					
Project ID:		BLOOM_RESP		Sampling Agency:		NEROC					
Location: Doctor's Lake -Catfish Point				Comments: Odor present at sample site location.							
Field ID: NAEHAB0110				Coordinates:		30.11996, -81.74755					
WIN Secondary Resource Type: Lake				Collection Equip:		WaterBottle					
WIN ID: NEHAB0110				High Tide Time:		1141		Low Tide Time:		0712	
Matrix:		Fresh		Grab		Tide:		☒ Rising ☐ Falling ☐ Slack ☐ N/A			
Water Quality											
Date	Time ☒ EST ☐ CEN	D.O. (%)	D.O. (mg/L)	Temp (°C)	pH (SU)	Sample Depth (m)	Total Depth (m)	Secchi Disk (m)	Sp. Cond. (µmho/cm)	Salinity (PPTH)	
08/01/2023	10:50	57.9	4.33	29.7	7.26	0.3	3.12	0.8	4501	2.38	
	10:53	33.2	2.52	29.1	7.1	2			4477	2.38	
	10:49	5.5	0.43	29.2	7.13	3.07			4549	2.41	
Bloom Observations											
Algal Bloom Observed? ☒ Yes ☐ No											
Water Clarity		☐ Clear		☐ Slightly Turbid		☒ Turbid		☐ Opaque			
Bloom Color		☒ Green		☐ Red		☐ Brown		☐ Other:			
Algae Type		☒ Planktonic		☐ Filamentous		☐ Other:					
Algae Accumulation		☐ Mat		☒ Specks		☐ Scum		☒ Glob			
Algae Distribution		☒ Surface-Entire		☐ Surf-Localized		☐ Surf-Windblown		☒ Suspended		☐ Benthic ☐ Epiphytic	
Parameter Suite		Parameter Methods				Preservation		pH Check	#Bottles Sent to Lab	Bottle Group	
Algae (PT-50mL)		☐ ALGAL-ID ☒ BLOOM-ID				☒ Ice			1	A	
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	C	
Toxins (P-125/BG-250)		☒ W-MCYST-AA ☒ W-SAXTN-MS ☐ OV-TOXN				☒ Ice			1	C	
Physical/Aggregate		☒ W-COLOR				☒ Ice			1	C	
Preserved Nutrients (500mL HDPE)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-S-A-TP				☒ Ice ☒ H2SO4		☒ <2	1	C	
		Acid Lot:		SA3013020		Exp:		01/25/2024			
Filtered Nutrients (P125mL)		☒ W-PO4-F		☒ Field Filtered 0.45um Filter		☒ Ice			1	C	
				Filter Lot # 0000175533							
Quantitative Algae		☐ DTY-QN-C / PKD-QN-C				☐ Ice ☐ Lugols					
Relinquished By:		Date/Time		Number of Coolers Shipped		Received By:		Date/Time			
Katherine Halbert		08/01/2023 16:30		2							



Cyanobacteria Bloom Response Field Data Sheet (2023-06-12 v1)

Email scanned submittal forms to lab.receiving@FloridaDEP.gov

RQ - 2023-07-24-49				Collected By:		Katrin Villinger, Katherine Halbert					
Customer:		NE-DIST		Field Report Prepared By:		Katherine Halbert					
Project ID:		BLOOM_RESP		Sampling Agency:		NEROC					
Location: Doctors Lake - 1915 Salt Myrtle Lane				Comments: Odor present at sample site location.							
Field ID: NEHAB0108				Coordinates:		30.12497, -81.71725					
WIN Secondary Resource Type: Lake				Collection Equip:		WaterBottle					
WIN ID: NEHAB0108				High Tide Time:		1141		Low Tide Time:		0712	
Matrix:		Fresh		Grab		Tide:		☒ Rising ☐ Falling ☐ Slack ☐ N/A			
Water Quality											
Date	Time ☒ EST ☐ CEN	D.O. (%)	D.O. (mg/L)	Temp (°C)	pH (SU)	Sample Depth (m)	Total Depth (m)	Secchi Disk (m)	Sp. Cond. (µmho/cm)	Salinity (PPTH)	
08/01/2023	10:25	51.2	3.84	29.7	7.03	0.3	1.49	1	3875	2.03	
	10:27	29.1	2.21	28.9	6.85	1			3645	1.9	
Bloom Observations											
Algal Bloom Observed? ☒ Yes ☐ No											
Water Clarity		☐ Clear		☒ Slightly Turbid		☐ Turbid		☐ Opaque			
Bloom Color		☒ Green		☐ Red		☐ Brown		☐ Other:			
Algae Type		☒ Planktonic		☐ Filamentous		☐ Other:					
Algae Accumulation		☐ Mat		☒ Specks		☐ Scum		☒ Glob			
Algae Distribution		☒ Surface-Entire		☐ Surf-Localized		☐ Surf-Windblown		☒ Suspended		☐ Benthic ☐ Epiphytic	
Parameter Suite		Parameter Methods				Preservation		pH Check	#Bottles Sent to Lab	Bottle Group	
Algae (PT-50mL)		☐ ALGAL-ID ☒ BLOOM-ID				☒ Ice			1	A	
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	C	
Toxins (P-125/BG-250)		☒ W-MCYST-AA ☒ W-SAXTN-MS ☐ OV-TOXN				☒ Ice			1	C	
Physical/Aggregate		☒ W-COLOR				☒ Ice			1	C	
Preserved Nutrients (500mL HDPE)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-S-A-TP				☒ Ice ☒ H2SO4		☒ <2	1	C	
		Acid Lot:		SA3013020		Exp:		01/25/2024			
Filtered Nutrients (P125mL)		☒ W-PO4-F		☒ Field Filtered 0.45um Filter		☒ Ice			1	C	
		Filter Lot #		0000175533							
Quantitative Algae		☐ DTY-QN-C / PKD-QN-C				☐ Ice ☐ Lugols					
Relinquished By:		Date/Time		Number of Coolers Shipped		Received By:		Date/Time			
Katherine Halbert		08/01/2023 10:35		2							



Cyanobacteria Bloom Response Field Data Sheet (2023-06-12 v1)

Email scanned submittal forms to lab.receiving@FloridaDEP.gov

RQ - 2023-07-24-49				Collected By:		Katrin Villinger, Katherine Halbert					
Customer:		NE-DIST		Field Report Prepared By:		Katherine Halbert					
Project ID:		BLOOM_RESP		Sampling Agency:		NEROC					
Location: Doctors Lake - Mill Cove				Comments: Odor present at sample site location.							
Field ID: NEHAB0057				Coordinates:		30.12509, -81.72749					
WIN Secondary Resource Type: Lake				Collection Equip:		WaterBottle					
WIN ID: NEHAB0057				High Tide Time:		11:41		Low Tide Time:		07:12	
Matrix:		Fresh		Grab		Tide:		☒ Rising ☐ Falling ☐ Slack ☐ N/A			
Water Quality											
Date	Time ☒ EST ☐ CEN	D.O. (%)	D.O. (mg/L)	Temp (°C)	pH (SU)	Sample Depth (m)	Total Depth (m)	Secchi Disk (m)	Sp. Cond. (µmho/cm)	Salinity (PPTH)	
08/01/2023	11:35	109.6	8.13	30.3	8.15	0.3	2.83	0.7	4425	2.34	
	11:39	78.6	5.9	29.7	7.64	2			4425	2.33	
	11:36	70.1	5.26	29.6	7.55	2.33			4446	2.35	
Bloom Observations											
Algal Bloom Observed? ☒ Yes ☐ No											
Water Clarity		☐ Clear		☐ Slightly Turbid		☒ Turbid		☐ Opaque			
Bloom Color		☒ Green		☐ Red		☐ Brown		☐ Other:			
Algae Type		☒ Planktonic		☐ Filamentous		☐ Other:					
Algae Accumulation		☐ Mat		☒ Specks		☐ Scum		☒ Glob			
Algae Distribution		☐ Surface-Entire		☐ Surf-Localized		☐ Surf-Windblown		☒ Suspended		☐ Benthic ☐ Epiphytic	
Parameter Suite		Parameter Methods				Preservation		pH Check	#Bottles Sent to Lab	Bottle Group	
Algae (PT-50mL)		☐ ALGAL-ID ☒ BLOOM-ID				☒ Ice			1	A	
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	C	
Toxins (P-125/BG-250)		☒ W-MCYST-AA ☒ W-SAXTN-MS ☐ OV-TOXN				☒ Ice			1	C	
Physical/Aggregate		☒ W-COLOR				☒ Ice			1	C	
Preserved Nutrients (500mL HDPE)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-S-A-TP				☒ Ice ☒ H2SO4		☒ <2	1	C	
		Acid Lot:		SA3013020		Exp:		01/25/2024			
Filtered Nutrients (P125mL)		☒ W-PO4-F		☒ Field Filtered 0.45um Filter		☒ Ice			1	C	
				Filter Lot # 0000175533							
Quantitative Algae		☐ DTY-QN-C / PKD-QN-C				☐ Ice ☐ Lugols					
Relinquished By:		Date/Time		Number of Coolers Shipped		Received By:		Date/Time			
Katherine Halbert		08/01/2023 16:30		2							