

Chemical Analysis Report

WQAP-2023-03-22-01

Florida DEP Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Algal Bloom Response - CEROC**
Request ID: **RQ-2023-03-06-43**
Customer: **WQAP**
Project ID: **BLOOM-RESP**

Send Reports to:
FL Dept. of Environmental Protection
3319 Maguire Blvd. Suite 232
Orlando, FL 32803
Attn: Jessie Atchison

For additional information please contact
Colin Wright, Ph.D.
Chris Bowen, B.S.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Dr. rer. nat. Bettina Steinbock, Environmental Administrator

Date Certified: 18-APR-2023 12:29



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical groups are included in this report: Nutrients and Pesticides.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Tiger Lake - Center

Collection Date/Time: 03/21/2023 12:19

Field ID: HABCE0185_TigerLake2

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2395429	EPA 8321B	Anatoxin-a	0.25	U	ug/L	P427536	
		Cylindrospermopsin	0.10	U	ug/L	P427444	
		Desmethyl microcystin LR	0.35	I	ug/L	P427444	
		Microcystin HILR	0.25	U	ug/L	P427444	
		Microcystin HtyR	0.25	U	ug/L	P427444	
		Microcystin LA	0.35	I	ug/L	P427444	
		Microcystin LF	0.23	I	ug/L	P427444	
		Microcystin LR	4.3		ug/L	P427444	
		Microcystin LW	0.43	I	ug/L	P427444	
		Microcystin LY	0.44		ug/L	P427444	
		Microcystin RR	0.36	I	ug/L	P427444	
		Microcystin WR	0.50	U	ug/L	P427444	
		Neosaxitoxin	0.50	U	ug/L	P427536	
		Microcystin YR	0.25	U	ug/L	P427444	
		Saxitoxin	0.50	U	ug/L	P427536	
		Nodularin-R	0.10	U	ug/L	P427444	
2395430	EPA 350.1 Rev. 2.0	Ammonia-N	0.007		mg N/L	P427633	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.9		mg N/L	P428038	MS
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P427864	
	EPA 365.1 Rev. 2.0	Total-P	0.12		mg P/L	P427596	
2395431	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P427518	

Sample Location: Tiger Lake -Center

Collection Date/Time: 03/21/2023 12:40

Field ID: FB_03212023

Matrix: W-FIELD-BK

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2395553	EPA 8321B	Anatoxin-a	0.50	U	ug/L	P427536	
		Cylindrospermopsin	0.10	U	ug/L	P427444	
		Desmethyl microcystin LR	0.25	U	ug/L	P427444	
		Microcystin HILR	0.25	U	ug/L	P427444	
		Microcystin HtyR	0.25	U	ug/L	P427444	
		Microcystin LA	0.10	U	ug/L	P427444	
		Microcystin LF	0.10	U	ug/L	P427444	
		Microcystin LR	0.25	U	ug/L	P427444	
		Microcystin LW	0.25	U	ug/L	P427444	
		Microcystin LY	0.10	U	ug/L	P427444	
		Microcystin RR	0.10	U	ug/L	P427444	
		Microcystin WR	0.50	U	ug/L	P427444	
		Neosaxitoxin	1.0	U	ug/L	P427536	
		Microcystin YR	0.25	U	ug/L	P427444	
		Saxitoxin	1.0	U	ug/L	P427536	
		Nodularin-R	0.10	U	ug/L	P427444	
2395554	EPA 350.1 Rev. 2.0	Ammonia-N	0.002	U	mg N/L	P427633	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.080	U	mg N/L	P428381	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P427816	
	EPA 365.1 Rev. 2.0	Total-P	0.002	U	mg P/L	P427598	
2395555	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P427518	

Ref. Method and Comment:

EPA 8321B: MDLs are elevated due to required dilution of sample matrix.

Quality Assurance Report Method Blank Results

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P427633

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P428038

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P428381

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P427816

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P427864

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P427596

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P427598

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P427518

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: EPA 8321B
Batch ID: P427444

Component	Result	Code	Units
Cylindrospermopsin	0.10	U	ug/L
Desmethyl microcystin LR	0.25	U	ug/L
Microcystin HILR	0.25	U	ug/L
Microcystin HtyR	0.25	U	ug/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 8321B

Batch ID: P427444

Component	Result	Code	Units
Microcystin LA	0.10	U	ug/L
Microcystin LF	0.10	U	ug/L
Microcystin LR	0.25	U	ug/L
Microcystin LW	0.25	U	ug/L
Microcystin LY	0.10	U	ug/L
Microcystin RR	0.10	U	ug/L
Microcystin WR	0.50	U	ug/L
Microcystin YR	0.25	U	ug/L
Nodularin-R	0.10	U	ug/L

Reference Method: EPA 8321B

Batch ID: P427536

Component	Result	Code	Units
Anatoxin-a	0.25	U	ug/L
Neosaxitoxin	0.50	U	ug/L
Saxitoxin	0.50	U	ug/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P427633

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	93.2		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P428038

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	106		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P428381

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	100		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P427816

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	103		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P427864

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	103		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P427596

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	101		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P427598

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	98.8		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P427518

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	101		P	90 - 110

Reference Method: EPA 8321B
Batch ID: P427444

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Cylindrospermopsin	96.2		P	40 - 150
Desmethyl microcystin LR	103		P	40 - 150
Microcystin HILR	103		P	40 - 150
Microcystin HtyR	104		P	40 - 150

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 8321B
Batch ID: P427444

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Microcystin LA	100		P	40 - 150
Microcystin LF	110		P	40 - 150
Microcystin LR	95.9		P	40 - 150
Microcystin LW	99.5		P	40 - 150
Microcystin LY	91.4		P	40 - 150
Microcystin RR	95.7		P	40 - 150
Microcystin WR	93.1		P	40 - 150
Microcystin YR	95.5		P	40 - 150
Nodularin-R	97.5		P	40 - 150

Reference Method: EPA 8321B
Batch ID: P427536

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Anatoxin-a	72.2		P	40 - 150
Neosaxitoxin	101		P	40 - 150
Saxitoxin	77.9		P	40 - 150

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P427633

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2395549	Ammonia-N	97.8	98.4	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P428038

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2395435	Kjeldahl Nitrogen	115	113	F/F	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P428381

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2396676	Kjeldahl Nitrogen	97.2	95.6	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P427816

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2395587	NO2NO3-N	105	104	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P427864

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2395395	NO2NO3-N	104	104	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P427596

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2395434	Total-P	97.7	98.6	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P427598

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2395639	Total-P	101	101	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P427518

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2395702	O-Phosphate-P	100	99.1	P/P	90 - 110

Reference Method: EPA 8321B
Batch ID: P427444

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2395004	Cylindrospermopsin	102	104	P/P	40 - 150
2395004	Desmethyl microcystin LR	105	115	P/P	40 - 150
2395004	Microcystin HtIR	106	113	P/P	40 - 150
2395004	Microcystin HtyR	109	115	P/P	40 - 150

Quality Assurance Report
Matrix Spike Accuracy

Reference Method: EPA 8321B
Batch ID: P427444

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2395004	Microcystin LA	95.7	126	P/P	40 - 150
2395004	Microcystin LF	103	112	P/P	40 - 150
2395004	Microcystin LR	99.8	111	P/P	40 - 150
2395004	Microcystin LW	95.0	96.3	P/P	40 - 150
2395004	Microcystin LY	94.0	95.4	P/P	40 - 150
2395004	Microcystin RR	103	109	P/P	40 - 150
2395004	Microcystin WR	98.8	93.5	P/P	40 - 150
2395004	Microcystin YR	101	106	P/P	40 - 150
2395004	Nodularin-R	104	110	P/P	40 - 150

Reference Method: EPA 8321B
Batch ID: P427536

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2395004	Anatoxin-a	70.6	67.4	P/P	40 - 150
2395004	Neosaxitoxin	82.8	78.0	P/P	40 - 150
2395004	Saxitoxin	67.8	64.0	P/P	40 - 150

Quality Assurance Report Precision

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P427633

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2395549	Ammonia-N	0.437	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P428038

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2395435	Kjeldahl Nitrogen	1.28	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P428381

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2396676	Kjeldahl Nitrogen	1.37	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P427816

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2395587	NO2NO3-N	0.957	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P427864

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2395395	NO2NO3-N	0.0	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P427596

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2395434	Total-P	0.613	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P427598

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2395639	Total-P	0.0	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P427518

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2395702	O-Phosphate-P	0.926	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 8321B
Batch ID: P427444

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2395004	Cylindrospermopsin	1.53	Spike	P	0 - 30
2395004	Desmethyl microcystin LR	9.21	Spike	P	0 - 30
2395004	Microcystin HilR	6.53	Spike	P	0 - 30
2395004	Microcystin HtyR	5.48	Spike	P	0 - 30
2395004	Microcystin LA	27.0	Spike	P	0 - 30
2395004	Microcystin LF	8.10	Spike	P	0 - 30
2395004	Microcystin LR	10.9	Spike	P	0 - 30
2395004	Microcystin LW	1.39	Spike	P	0 - 30
2395004	Microcystin LY	1.47	Spike	P	0 - 30
2395004	Microcystin RR	5.36	Spike	P	0 - 30
2395004	Microcystin WR	5.57	Spike	P	0 - 30
2395004	Microcystin YR	4.15	Spike	P	0 - 30
2395004	Nodularin-R	5.84	Spike	P	0 - 30

Reference Method: EPA 8321B
Batch ID: P427536

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2395004	Anatoxin-a	4.61	Spike	P	0 - 30
2395004	Neosaxitoxin	5.99	Spike	P	0 - 30
2395004	Saxitoxin	5.85	Spike	P	0 - 30

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Surrogates

Lab Sample ID: 2395429

Field Sample ID: HABCE0185_TigerLake2

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8321B	Microcystin LR ethyl-d5	118	P	30 - 160
EPA 8321B	Neosaxitoxin-15N7	69.6	P	30 - 160

Lab Sample ID: 2395553

Field Sample ID: FB_03212023

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8321B	Microcystin LR ethyl-d5	65.7	P	30 - 160
EPA 8321B	Neosaxitoxin-15N7	78.8	P	30 - 160

Quality Assurance Report Calibration Verification

Reference Method: EPA 8321B

Run ID: A118088

Included Lab Sample IDs: 2395429, 2395553

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Anatoxin-a	65.9	64.9	P/P	50 - 160
Anatoxin-a	70.7	65.9	P/P	50 - 160
Cylindrospermopsin	93.5	78.6	P/P	50 - 160
Cylindrospermopsin	99.6	93.5	P/P	50 - 160
Desmethyl microcystin LR	96.7	114	P/P	50 - 160
Desmethyl microcystin LR	97.8	96.7	P/P	50 - 160
Microcystin H1R	106	106	P/P	50 - 160
Microcystin H1R	99.8	106	P/P	50 - 160
Microcystin HtyR	102	106	P/P	50 - 160
Microcystin HtyR	97.3	102	P/P	50 - 160
Microcystin LA	102	105	P/P	50 - 160
Microcystin LA	81.2	102	P/P	50 - 160
Microcystin LF	80.8	99.5	P/P	50 - 160
Microcystin LF	99.4	80.8	P/P	50 - 160
Microcystin LR	92.6	101	P/P	50 - 160
Microcystin LR	94.5	92.6	P/P	50 - 160
Microcystin LW	79.2	73.1	P/P	50 - 160
Microcystin LW	94.5	79.2	P/P	50 - 160
Microcystin LY	82.6	98.1	P/P	50 - 160
Microcystin LY	98.1	86.8	P/P	50 - 160
Microcystin RR	92.1	97.0	P/P	50 - 160
Microcystin RR	97.0	94.6	P/P	50 - 160
Microcystin WR	76.4	85.5	P/P	50 - 160
Microcystin WR	85.5	83.7	P/P	50 - 160
Microcystin YR	98.3	98.9	P/P	50 - 160
Microcystin YR	98.9	94.8	P/P	50 - 160
Neosaxitoxin	101	92.9	P/P	50 - 160
Neosaxitoxin	92.9	90.4	P/P	50 - 160
Nodularin-R	96.1	95.6	P/P	50 - 160
Nodularin-R	96.5	96.1	P/P	50 - 160
Saxitoxin	74.0	72.4	P/P	50 - 160
Saxitoxin	77.4	74.0	P/P	50 - 160

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A118144

Included Lab Sample IDs: 2395431, 2395555

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	100	98.1	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A118183

Included Lab Sample IDs: 2395430, 2395554

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	101	101	P/P	90 - 110
Ammonia-N	99.4	101	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0
Run ID: A118197
Included Lab Sample IDs: 2395430

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	102	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Run ID: A118211
Included Lab Sample IDs: 2395554

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	98.2	96.8	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Run ID: A118262
Included Lab Sample IDs: 2395554

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	104	104	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Run ID: A118290
Included Lab Sample IDs: 2395430

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	104	104	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Run ID: A118440
Included Lab Sample IDs: 2395430

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	99.3	103	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Run ID: A118562
Included Lab Sample IDs: 2395554

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	92.8	97.3	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.
Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 350.1 Rev. 2.0	Ammonia-N	93.2	97.8	98.4			0.437
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	106	115	113			1.28
	Kjeldahl Nitrogen	100	97.2	95.6			1.37
EPA 353.2 Rev. 2.0	NO2NO3-N	103	105	104			0.957
	NO2NO3-N	103	104	104			0.0
EPA 365.1 Rev. 2.0	Total-P	101	97.7	98.6			0.613
	Total-P	98.8	101	101			0.0
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	101	100	99.1			0.926
EPA 8321B	Anatoxin-a	72.2	70.6	67.4			4.61
	Cylindrospermopsin	96.2	102	104			1.53
	Desmethyl microcystin LR	103	105	115			9.21
	Microcystin HiIR	103	106	113			6.53
	Microcystin HtyR	104	109	115			5.48
	Microcystin LA	100	95.7	126			27.0
	Microcystin LF	110	103	112			8.10
	Microcystin LR	95.9	99.8	111			10.9
	Microcystin LW	99.5	95.0	96.3			1.39
	Microcystin LY	91.4	94.0	95.4			1.47
	Microcystin RR	95.7	103	109			5.36
	Microcystin WR	93.1	98.8	93.5			5.57
	Microcystin YR	95.5	101	106			4.15
	Neosaxitoxin	101	82.8	78.0			5.99
	Nodularin-R	97.5	104	110			5.84
	Saxitoxin	77.9	67.8	64.0			5.85

Reference Method Descriptions

Method	Description	Associated Samples
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2395430, 2395554
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2395430, 2395554
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2395430, 2395554
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2395430, 2395554
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2395431, 2395555
EPA 8321B	Microcystins in water matrices by HPLC/MS/MS	2395429, 2395553
EPA 8321B	Saxitoxin in water matrices by HPLC/MS/MS	2395429, 2395553

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 350.1 Rev. 2.0	03/22/2023			03/24/2023 10:42	Ping Hua	2395430
	03/22/2023			03/24/2023 13:02	Ping Hua	2395554
EPA 351.2 Rev. 2.0	03/22/2023	04/05/2023 11:13	Dana G. Hughes	04/06/2023 14:14	Samantha L Bates	2395430
	03/22/2023	04/12/2023 13:25	Dana G. Hughes	04/13/2023 16:15	Samantha L Bates	2395554
EPA 353.2 Rev. 2.0	03/22/2023			03/29/2023 10:41	Beverly Y. Sanders	2395554
	03/22/2023			03/30/2023 09:27	Beverly Y. Sanders	2395430
EPA 365.1 Rev. 2.0	03/22/2023	03/24/2023 10:10	Alexis Price	03/27/2023 09:13	James L Waggerby	2395430
	03/22/2023	03/24/2023 10:10	Alexis Price	03/27/2023 13:20	James L Waggerby	2395554
EPA 365.1 Rev. 2.0 dissolved	03/22/2023			03/22/2023 16:25	Elise M. Gillis	2395431
	03/22/2023			03/22/2023 16:27	Elise M. Gillis	2395555
EPA 8321B	03/22/2023	03/23/2023 13:00	Manjita Shrestha	03/23/2023 19:59	Manjita Shrestha	2395429
	03/22/2023	03/23/2023 13:00	Manjita Shrestha	03/23/2023 20:33	Manjita Shrestha	2395553
	03/22/2023	03/23/2023 13:00	Manjita Shrestha	03/24/2023 01:43	Manjita Shrestha	2395429
	03/22/2023	03/23/2023 13:00	Manjita Shrestha	03/24/2023 03:21	Manjita Shrestha	2395553