



Cyanobacteria Bloom Response Field Data Sheet (2023-06-12 v1)

Email scanned submittal forms to lab.receiving@FloridaDEP.gov

RQ - 2024-03-04-72				Collected By:		Katrin Villinger, Haylee Brown					
Customer:		NE-DIST		Field Report Prepared By:		HAYLEE BROWN					
Project ID:		BLOOM_RESP		Sampling Agency:		NEROC					
Location: Crescent lake - Eagle Trail				Comments: Very windy							
Field ID: NEHAB0006				Coordinates:		29.38553, -81.47247					
WIN Secondary Resource Type: Lake				Collection Equip:		WaterBottle					
WIN ID: NEHAB0006				High Tide Time:				Low Tide Time:			
Matrix:		Fresh		Grab		Tide:		☐ Rising ☐ Falling ☐ Slack ☒ N/A			
Water Quality											
Date	Time ☒ EST ☐ CEN	D.O. (%)	D.O. (mg/L)	Temp (°C)	pH (SU)	Sample Depth (m)	Total Depth (m)	Secchi Disk (m)	Sp. Cond. (µmho/cm)	Salinity (PPTH)	
06/18/2024	11:25	104.3	7.93	29.6	8.1	0.3	1.48	0.6	482.1	0.23	
Bloom Observations											
Algal Bloom Observed? ☒ Yes ☐ No											
Water Clarity		☐ Clear		☐ Slightly Turbid		☒ Turbid		☐ Opaque			
Bloom Color		☒ Green		☐ Red		☐ Brown		☐ Other:			
Algae Type		☒ Planktonic		☐ Filamentous		☐ Other:					
Algae Accumulation		☐ Mat		☒ Specks		☐ Scum		☐ Glob			
Algae Distribution		☐ Surface-Entire		☐ Surf-Localized		☐ Surf-Windblown		☒ Suspended		☐ Benthic ☐ Epiphytic	
Parameter Suite		Parameter Methods				Preservation		pH Check	#Bottles Sent to Lab	Bottle Group	
Algae (PT-50mL)		☐ ALGAL-ID ☒ BLOOM-ID				☒ Ice			1	A	
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	C	
Toxins (P-125/BG-250)		☒ W-MCYST-AA ☒ W-SAXTN-MS ☐ OV-TOXN				☒ Ice			1	B	
Physical/Aggregate		☒ W-COLOR				☒ Ice			1	C	
Preserved Nutrients (500mL HDPE)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-S-A-TP				☒ Ice ☒ H2SO4		☒ <2	1	C	
		Acid Lot:		SA3340020		Exp:		12/11/2024			
Filtered Nutrients (P125mL)		☒ W-PO4-F		☒ Field Filtered 0.45um Filter		☒ Ice			1	C	
		Filter Lot #		R1EB01329							
Quantitative Algae		☐ DTY-QN-C / PKD-QN-C				☐ Ice ☐ Lugols					
Relinquished By:		Date/Time		Number of Coolers Shipped				Received By:		Date/Time	
HAYLEE BROWN		06/18/2024 11:25		1							



Cyanobacteria Bloom Response Field Data Sheet (2023-06-12 v1)

Email scanned submittal forms to lab.receiving@FloridaDEP.gov

RQ - 2024-03-04-72				Collected By:		Katrin Villinger, Haylee Brown					
Customer:		NE-DIST		Field Report Prepared By:		HAYLEE BROWN					
Project ID:		BLOOM_RESP		Sampling Agency:		NEROC					
Location: Crescent Lake - New City Boat Ramp				Comments: Very windy coming from N							
Field ID: NEHAB0118				Coordinates:		29.42829, -81.50675					
WIN Secondary Resource Type: Lake				Collection Equip:		WaterBottle					
WIN ID: NEHAB0118				High Tide Time:				Low Tide Time:			
Matrix:		Fresh		Grab		Tide:		☐ Rising ☐ Falling ☐ Slack ☒ N/A			
Water Quality											
Date	Time ☒ EST ☐ CEN	D.O. (%)	D.O. (mg/L)	Temp (°C)	pH (SU)	Sample Depth (m)	Total Depth (m)	Secchi Disk (m)	Sp. Cond. (µmho/cm)	Salinity (PPTH)	
06/18/2024	10:40	104.7	7.93	29.8	8.39	0.3	2.14	0.5	477.7	0.23	
	10:41	104.1	7.92	29.7	8.34	1.6				477.5	
Bloom Observations											
Algal Bloom Observed? ☒ Yes ☐ No											
Water Clarity		☐ Clear		☐ Slightly Turbid		☒ Turbid		☐ Opaque			
Bloom Color		☒ Green		☐ Red		☐ Brown		☐ Other:			
Algae Type		☒ Planktonic		☐ Filamentous		☐ Other:					
Algae Accumulation		☐ Mat		☒ Specks		☐ Scum		☐ Glob			
Algae Distribution		☐ Surface-Entire		☐ Surf-Localized		☐ Surf-Windblown		☒ Suspended		☐ Benthic ☐ Epiphytic	
Parameter Suite		Parameter Methods				Preservation		pH Check	#Bottles Sent to Lab	Bottle Group	
Algae (PT-50mL)		☐ ALGAL-ID ☒ BLOOM-ID				☒ Ice			1	A	
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	C	
Toxins (P-125/BG-250)		☒ W-MCYST-AA ☒ W-SAXTN-MS ☐ OV-TOXN				☒ Ice			1	B	
Physical/Aggregate		☒ W-COLOR				☒ Ice			1	C	
Preserved Nutrients (500mL HDPE)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-S-A-TP				☒ Ice ☒ H2SO4		☒ <2	1	C	
		Acid Lot:		SA3340020		Exp:		12/11/2024			
Filtered Nutrients (P125mL)		☒ W-PO4-F		☒ Field Filtered 0.45um Filter		☒ Ice			1	C	
				Filter Lot # R1EB01329							
Quantitative Algae		☐ DTY-QN-C / PKD-QN-C				☐ Ice ☐ Lugols					
Relinquished By:		Date/Time		Number of Coolers Shipped		Received By:		Date/Time			
HAYLEE BROWN		06/18/2024 10:50		1							

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Desi RQ: 2024-03-04-72 Project: Algae Response

Crescent Lake

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 4/1/24

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	H. Brown	6/18/24	800	22.7	761.7	8.7	—	100.2	—	1.016	(P) F	(L) F
ICV	H. Brown	6/18/24	805	22.6	761.8	8.6	8.67	100.3	—	1.016	(P) F	(L) F
CCV	H. Brown	6/18/24	1535	23.3	761.9	8.6	8.51	99.8	—	1.016	(P) F	(L) F
CCV											P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	H. Brown	6/18/24	807	240208B	02/2025	1000	1000	(P) F	(L) F
ICV	H. Brown	6/18/24	810	230612A	06/2024	100	99.7	(P) F	(L) F
CCV	H. Brown	6/18/24	1537	240208B	02/2025	1000	1003	(P) F	(L) F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	H. Brown	6/18/24	812	43080549	9/7/25	7.01	23	7.01	-30.4	(P) F	(L) F
Calibr.	H. Brown	6/18/24	814	44020143	2/24/26	4.00	23.1	4.00	142.7	(P) F	(L) F
Calibr.	H. Brown	6/18/24	816	44020286	2/27/26	10.02	23.2	10.02	-200	(P) F	(L) F
ICV	H. Brown	6/18/24	817	44020143	2/24/26	4.00	23	4.02	142.4	(P) F	(L) F
CCV	H. Brown	6/18/24	1538	44020286	2/27/26	10.02	23	10.05	-202.9	(P) F	(L) F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 169.6, slope from 4 to 7 173.1 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? (YES) / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 4/1/24

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	H. Brown	6/18/24	818	0.00	0.00	(P) F	(L) F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: