

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: 154179

RQ: 2024-02-19-59

Project: HAB

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 2/8/24

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	NM	2/20/24	0800	19.0	769	9.4	9.4	104.2			P / F	L / F
ICV	↓	↓	0802	↓	↓	↓	9.67	104.3			P / F	L / F
CCV	↓	↓	1430	19.0	767	9.4	9.59	106.4			P / F	L / F
CCV											P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	NM	2/24/24	0805	240110B	1/28	1000	1000	P / F	L / F
ICV	↓	↓	0808	23007A	7/24	100	104	P / F	L / F
CCV	↓	↓	1452	23007C	7/24	10,000	910	P / F	L / F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	NM	2/20/24	0810	1110690	8/24	7.	19.0	7.0		P / F	L / F
Calibr.	↓	↓	0812	111081de	5/24	4.	↓	4.0		P / F	L / F
Calibr.	↓	↓	0814	120057e	6/24	10.	↓	10.0		P / F	L / F
ICV	↓	↓	0816	1110690	8/24	7.0	↓	7.16	-41.8	P / F	L / F
CCV	↓	↓	1455	1110576	6/24	10.0	19.5	9.88	-71.8	P / F	L / F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 2/8/24

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	NM	2/20/24	0820	0.00	0.00	P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:



Cyanobacteria Bloom Response Field Data Sheet (2023-06-12 v1)

Email scanned submittal forms to lab.receiving@FloridaDEP.gov

RQ - 2024-02-19-59				Collected By:		Nathan Mulkey					
Customer:		NWROC		Field Report Prepared By:		Nathan Mulkey					
Project ID:		BLOOM_RESP		Sampling Agency:		NWROC					
Location: Unnamed ditch @ Legion Park				Comments: Ditch not flowing above ground. Pipe underground flowing with outfall in bay. Sample taken at end of pipe.							
Field ID: NWHAB0015				Coordinates:		30.38696, -86.35861					
WIN Secondary Resource Type: Estuary				Collection Equip:		WaterBottle					
WIN ID: NWHAB0015				High Tide Time:		19:39		Low Tide Time:		06:13	
Matrix:		Fresh		Grab		Tide:		☒ Rising ☐ Falling ☐ Slack ☐ N/A			
Water Quality											
Date	Time ☐ EST ☒ CEN	D.O. (%)	D.O. (mg/L)	Temp (°C)	pH (SU)	Sample Depth (m)	Total Depth (m)	Secchi Disk (m)	Sp. Cond. (µmho/cm)	Salinity (PPTH)	
02/20/2024	10:40	81.1	8.3	13.9	7.49	0.05	0.1	0.1	3000	1.6	
Bloom Observations											
Algal Bloom Observed? ☐ Yes ☒ No											
Water Clarity		☒ Clear		☐ Slightly Turbid		☐ Turbid		☐ Opaque			
Bloom Color		☐ Green		☐ Red		☐ Brown		☐ Other:			
Algae Type		☐ Planktonic		☐ Filamentous		☐ Other:					
Algae Accumulation		☐ Mat		☐ Specks		☐ Scum		☐ Glob			
Algae Distribution		☐ Surface-Entire		☐ Surf-Localized		☐ Surf-Windblown		☐ Suspended		☐ Benthic ☐ Epiphytic	
Parameter Suite		Parameter Methods				Preservation		pH Check	#Bottles Sent to Lab	Bottle Group	
Algae (PT-50mL)		☐ ALGAL-ID ☐ BLOOM-ID				☐ Ice					
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	C	
Toxins (P-125/BG-250)		☒ W-MCYST-AA ☒ W-SAXTN-MS ☐ OV-TOXN				☒ Ice			1	B	
Physical/Aggregate		☐ W-COLOR				☐ Ice					
Preserved Nutrients (500mL HDPE)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-S-A-TP				☒ Ice ☒ H2SO4		☐ <2	1	C	
		Acid Lot:		3193140		Exp:		07/31/2024			
Filtered Nutrients (P125mL)		☒ W-PO4-F		☒ Field Filtered 0.45um Filter		☒ Ice			1	C	
				Filter Lot # 17460495							
Quantitative Algae		☐ DTY-QN-C / PKD-QN-C				☐ Ice ☐ Lugols					
Relinquished By:		Date/Time		Number of Coolers Shipped		Received By:		Date/Time			
Nathan Mulkey		02/20/2024 11:19		1							