

Chemical Analysis Report

NW-DIST-2024-06-06-02

Florida DEP Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780


Event Description: **Algal Bloom Response - NW ROC**
Request ID: **RQ-2024-05-27-31**
Customer: **NW-DIST**
Project ID: **BLOOM-RESP**

Send Reports to:
FL Dept. of Environmental Protection
160 Government Center
Suite 308
Pensacola, FL 32502-5794
Attn: Myranda Butler

For additional information please contact
Colin Wright, Ph.D.
Chris Bowen, B.S.
Marek Topolski, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Chris Bowen, Environmental Administrator

Date Certified: 21-JUN-2024 08:58



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Lullwater Lake NE Lobe at dock

Collection Date/Time: 06/05/2024 09:55

Field ID: NWHAB0020					Matrix: W-SURF-FRH			
Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes	
2491927	DEP SOP: NU-094-1	Color (true)	80		PCU	P446148		
2491928	EPA 350.1 Rev. 2.0	Ammonia-N	0.008		mg N/L	P446280		
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.2		mg N/L	P446273		
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.31		mg N/L	P446300		
	EPA 365.1 Rev. 2.0	Total-P	0.16		mg P/L	P446375		
2491929	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.032		mg P/L	P446151		

Quality Assurance Report
Method Blank Results

Reference Method: DEP SOP: NU-094-1
Batch ID: P446148

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P446280

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P446273

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P446300

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P446375

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P446151

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Quality Assurance Report
Laboratory Control Sample Accuracy

Reference Method: DEP SOP: NU-094-1
Batch ID: P446148

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	100		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P446280

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	94.2		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P446273

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	98.5		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P446300

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	97.5		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P446375

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P446151

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	100		P	90 - 110

Quality Assurance Report

Matrix Spike Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P446280

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2491894	Ammonia-N	97.6	100	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P446273

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2491323	Kjeldahl Nitrogen	104	103	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P446300

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2491827	NO2NO3-N	96.5		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P446375

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2492798	Total-P	105	104	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P446151

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2491929	O-Phosphate-P	98.7	98.1	P/P	90 - 110

Quality Assurance Report Precision

Reference Method: DEP SOP: NU-094-1
Batch ID: P446148

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2492107	Color (true)	1.70	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P446280

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2491894	Ammonia-N	2.43	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P446273

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2491323	Kjeldahl Nitrogen	0.149	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P446300

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2491827	NO2NO3-N	0.715	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P446375

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2492798	Total-P	0.865	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P446151

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2491929	O-Phosphate-P	0.461	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report

Calibration Verification

Reference Method: DEP SOP: NU-094-1
Run ID: A126027
Included Lab Sample IDs: 2491927

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	98.9	100	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Run ID: A126031
Included Lab Sample IDs: 2491929

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	97.9	99.2	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Run ID: A126071
Included Lab Sample IDs: 2491928

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	97.2	97.0	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Run ID: A126076
Included Lab Sample IDs: 2491928

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	99.0	98.5	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Run ID: A126131
Included Lab Sample IDs: 2491928

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	102	102	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Run ID: A126167
Included Lab Sample IDs: 2491928

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	102	99.8	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.
Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report

Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		Precision	MS
					SMP	
DEP SOP: NU-094-1	Color (true)	100			1.70	
EPA 350.1 Rev. 2.0	Ammonia-N	94.2	97.6	100		2.43
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	98.5	104	103		0.149
EPA 353.2 Rev. 2.0	NO2NO3-N	97.5	96.5			0.715
EPA 365.1 Rev. 2.0	Total-P	102	105	104		0.865
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	100	98.7	98.1		0.461

Reference Method Descriptions

Method	Description	Associated Samples
DEP SOP: NU-094-1	True Color measured at 450 nm (not valid for NPDES monitoring)	2491927
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2491928
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2491928
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2491928
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2491928
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2491929

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
DEP SOP: NU-094-1	06/06/2024			06/06/2024 14:07	Jessica K Wilks	2491927
EPA 350.1 Rev. 2.0	06/06/2024			06/10/2024 12:35	Khloe J Berg	2491928
EPA 351.2 Rev. 2.0	06/06/2024	06/12/2024 12:53	Robyn Blevins	06/14/2024 12:31	Samantha L Bates	2491928
EPA 353.2 Rev. 2.0	06/06/2024			06/10/2024 13:59	Fadila Guebre	2491928
EPA 365.1 Rev. 2.0	06/06/2024	06/12/2024 11:25	Adam P Silver	06/13/2024 11:27	Simonne.V. Calhoun	2491928
EPA 365.1 Rev. 2.0 dissolved	06/06/2024			06/06/2024 14:55	Adam P Silver	2491929