

Quarterly Depth Verification Regional Operation Centers

SOP - S&T Sampling Manual and ROC Training Manual
Record depth reading with only one rounded decimal figure
Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up

Meter ID#: _____	Date: _____	Time: _____
Location: _____		
Date of Last Verification: _____		
Depth using verified weighted line marked in increments of 0.1 m: _____ m		
Depth read with the meter: _____ m		
Pass/Fail: _____ (acceptance Criteria 10%)		
Person Performing Verification: _____		

Secchi/Weighted Line ID#: <u>Knot head</u>	Date: <u>11/12/16</u>	Time: <u>1000</u>
Location: <u>Lab</u>	Date of Last Verification: <u>8/12/16</u>	
Incremental markings of 0.1 m checked <u>Yes</u> yes/no		
Pass/Fail: <u>Pass</u> (acceptable criteria 10%)	New increments done: <u>Yes</u> yes/no/NA	
Total Length of Line marked in 0.1 m increments: <u>10</u> m		
Reference meter stick or metal measuring tape: <u>5</u> m		
Pass/Fail: <u>Pass</u> (acceptable criteria of 5%)	New increments done: _____ yes/no/NA	
Person Performing Verification: <u>Mike Drennan</u>		

Secchi/Weighted Line ID#: <u>Lead Foot</u>	Date: <u>11/12/16</u>	Time: <u>1045</u>
Location: <u>Lab</u>	Date of Last Verification: <u>8/12/16</u>	
Incremental markings of 0.1 m checked <u>Yes</u> yes/no		
Pass/Fail: <u>Pass</u> (acceptable criteria 10%)	New increments done: <u>Darkened</u> yes/no/NA	
Total Length of Line marked in 0.1 m increments: <u>10</u> m		
Reference meter stick or metal measuring tape: <u>5</u> m		
Pass/Fail: _____ (acceptable criteria of 5%)	New increments done: _____ yes/no/NA	
Person Performing Verification: <u>Mike Drennan</u>		

Comments: _____
Blue-red-Blue @ 10m mark

Quarterly Depth Verification Regional Operation Centers

SOP - S&T Sampling Manual and ROC Training Manual
Record depth reading with only one rounded decimal figure
Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up

Meter ID#: _____	Date: _____	Time: _____
Location: _____		
Date of Last Verification: _____		
Depth using verified weighted line marked in increments of 0.1 m: _____ m		
Depth read with the meter: _____ m		
Pass/Fail: _____ (acceptance Criteria 10%)		
Person Performing Verification: _____		

Secchi/Weighted Line ID#: <u>Rusty</u>	Date: <u>11/12/16</u>	Time: <u>1130</u>
Location: <u>Lab</u>	Date of Last Verification: <u>8/12/16</u>	
Incremental markings of 0.1 m checked <u>Yes</u> yes/no		
Pass/Fail: <u>Pass</u> (acceptable criteria 10%)	New increments done: <u>Yes</u> yes/no/NA	
Total Length of Line marked in 0.1 m increments: <u>5.8</u> m		
Reference meter stick or metal measuring tape: <u>5</u> m		
Pass/Fail: <u>Pass</u> (acceptable criteria of 5%)	New increments done: _____ yes/no/NA	
Person Performing Verification: <u>Mike Drennan</u>		

Secchi/Weighted Line ID#: _____	Date: _____	Time: _____
Location: <u>Lab</u>	Date of Last Verification: _____	
Incremental markings of 0.1 m checked _____ yes/no		
Pass/Fail: _____ (acceptable criteria 10%)	New increments done: _____ yes/no/NA	
Total Length of Line marked in 0.1 m increments: _____ m		
Reference meter stick or metal measuring tape: _____ m		
Pass/Fail: _____ (acceptable criteria of 5%)	New increments done: _____ yes/no/NA	
Person Performing Verification: _____		
Comments: <u>Blue - Red - Blue @ 5 m mark</u>		