

Jackson Blue BMAP

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page.

Meter ID: **Biology #2**

RQ: **2017-12-11-19**

Project: **S BMAP**

- Notes:** (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site _____

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification _____

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Ben Ralys	12/18/17	7:31	20.02	9.092	9.09	100.0	72.7	—	P/F	L/F
ICV	Ben Ralys	12/18/17	7:33	20.02	9.092	9.08	99.9	73.7	—	P/F	L/F
CCV	Blake Spencer	12/19/17	13:42	23.00	8.528	8.34	92.5%	72.7		P/F	L/F
CCV										P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	Ben Ralys	12/18/17	7:35	171005D	10/2018	1,000	1,000	P/F	L/F
ICV	Ben Ralys	12/18/17	7:36	171005D	10/2018	1,000	1,000	P/F	L/F
CCV	Blake Spencer	12/19/17	13:44	171005E	10/2018	100	101	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Ben Ralys	12/18/17	07:41	170614B	12/2018	7.0	7.00	25.9	P/F	L/F
Calibr.	Ben Ralys	12/18/17	07:44	171005A	10/2018	4.0	4.00	190.9	P/F	L/F
Calibr.	Ben Ralys	12/18/17	07:49	171005C	04/2019	10.0	9.99	134.6	P/F	L/F
ICV	Ben Ralys	12/18/17	07:50	171005C	04/2019	10.0	10.00	135.2	P/F	L/F
CCV	Blake Spencer	12/19/17	14:47	171005C	04/2019	10.0	10.09	149.6	P/F	L/F
CCV									P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Is meter equipped with depth sensor? **Yes / No** If yes, list date of last quarterly depth verification _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value meters	Pass / Fail	Lab / Field
Pressure mode in air						P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: