

Depth Verification Regional Operation Centers

SOP - S&T Sampling Manual and ROC Training Manual.

Report two decimal places for electronic devices. Report one decimal place for manual devices.

Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up.

QUARTERLY VERIFICATION OF ELECTRONIC DEVICES (SONDE, SONAR DEVICE, ETC.)

Meter / Device ID#:	Date of Last Verification:		
Date:	Time:	ETZ / CTZ	Verification Location:
Person Performing Verification:			
Reference Device: Graduated Bucket / Metal Measuring Tape / Meter Stick / Other			
Depth measurements: Reference Device: m ; Device Being Tested: m			
Result: Pass / Fail (acceptance Criteria 10%)			

Meter / Device ID#:	Date of Last Verification:		
Date:	Time:	ETZ / CTZ	Verification Location:
Person Performing Verification:			
Reference Device: Graduated Bucket / Metal Measuring Tape / Meter Stick / Other			
Depth measurements: Reference Device: m ; Device Being Tested: m			
Result: Pass / Fail (acceptance Criteria 10%)			

Comments:

6 MONTH VERIFICATION OF MANUAL DEVICES (SECCHI DISK, WEIGHTED LINE, ETC.)

Secchi/Weighted Line ID#:	Date of Last Verification:		
Date:	Time:	ETZ / CTZ	Verification Location:
Person Performing Verification:			
Reference Device: Metal Measuring Tape / Meter Stick / Other			
Incremental markings of 0.1 m checked: YES / NO			
Result: Pass / Fail (acceptable criteria 10%)			
Total length of line (up to anticipated depth encountered in field) checked: YES / NO			
Total Length: indicated by line markings m ; measured by reference device m			
Result: Pass / Fail (acceptable criteria of 5%)			
Markings redone: YES / NO			

Secchi/Weighted Line ID#:	Date of Last Verification:		
Date:	Time:	ETZ / CTZ	Verification Location:
Person Performing Verification:			
Reference Device: Metal Measuring Tape / Meter Stick / Other			
Incremental markings of 0.1 m checked: YES / NO			
Result: Pass / Fail (acceptable criteria 10%)			
Total length of line (up to anticipated depth encountered in field) checked: YES / NO			
Total Length: indicated by line markings m ; measured by reference device m			
Result: Pass / Fail (acceptable criteria of 5%)			
Markings redone: YES / NO			

Comments: