

Quarterly Temperature Verification Regional Operation Centers

DEP SOP FT 1400; Acceptance Criteria for Temp. $\pm 0.5^{\circ}\text{C}$.

Correction factor needed if acceptance criteria are exceeded.

Record all digits displayed for temperature readings. Do not round before reporting measurements.

Meter ID#:	<u>120840</u>	Date of Last Verification:	<u>12/01/17</u>
Date:	<u>8/8/18</u>	Time:	<u>750</u> ETZ / CTZ
NIST Reference Device ID#:		<u>844016</u>	
Cold Bath (between 0 – 10°C): Sonde:		<u>6.47</u> °C; NIST:	<u>6.4</u> °C; Result: <u>Pass</u> / Fail
Warm Bath (between 30 – 40°C): Sonde:		<u>37.4</u> °C; NIST:	<u>37.1</u> °C; Result: <u>Pass</u> / Fail
Is Correction Factor Needed? Y / <u>(N)</u> If yes, describe: _____			
Person Performing Verification: <u>Mike Drennan</u>			

Comments: _____

Meter ID#:	<u>SCUD</u>	Date of Last Verification:	<u>N/A (new instrument)</u>
Date:	<u>4/6/18</u>	Time:	<u>930</u> <u>(ETZ)</u> / CTZ
NIST Reference Device ID#:		<u>844016</u>	
Cold Bath (between 0 – 10°C): Sonde:		<u>2.8</u> °C; NIST:	<u>2.7</u> °C; Result: <u>Pass</u> / Fail
Warm Bath (between 30 – 40°C): Sonde:		<u>37.5</u> °C; NIST:	<u>37.3</u> °C; Result: <u>Pass</u> / Fail
Is Correction Factor Needed? Y / <u>(N)</u> If yes, describe: _____			
Person Performing Verification: <u>Mike Drennan</u>			

Comments: _____

Meter ID#:	<u>SCUD</u>	Date of Last Verification:	<u>4/6/18</u>
Date:	<u>7/6/18</u>	Time:	<u>1040</u> <u>(ETZ)</u> / CTZ
NIST Reference Device ID#:		<u>878708</u>	
Cold Bath (between 0 – 10°C): Sonde:		<u>7.1</u> °C; NIST:	<u>7.3</u> °C; Result: <u>Pass</u> / Fail
Warm Bath (between 30 – 40°C): Sonde:		<u>35.0</u> °C; NIST:	<u>34.7</u> °C; Result: <u>Pass</u> / Fail
Is Correction Factor Needed? Y / <u>(N)</u> If yes, describe: _____			
Person Performing Verification: <u>Mike Drennan</u>			

Comments: _____