

Chemical Analysis Report

NE-DIST-2020-09-23-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780


Event Description: **Bmap 2 2020 SMP**
Request ID: **RQ-2020-09-07-28**
Customer: **NE-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
8800 Baymeadows Way West
Suite 100
Jacksonville, FL 32256
Attn: Jeremy Parrish

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Dr. rer. nat. Bettina Steinbock, Environmental Administrator

Date Certified: 09-OCT-2020 11:12



NON-CONFORMANCE REPORT INCLUDED

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical groups are included in this report: Metals, Nutrients and Pesticides.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Moncrief Cr At Moncrief Road

Collection Date/Time: 09/22/2020 09:50

Field ID: 20030726

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2197720	EPA 180.1 Rev. 2.0	Turbidity	4.7		NTU	P387753	
	EPA 300.0 Rev. 2.1	Chloride	17	A	mg Cl/L	P388092	
		Sulfate	31	A	mg SO4/L	P388096	
	SM 2120 B-2011	Color (true)	44		PCU	P387764	
	SM 2320 B-2011	Total Alkalinity	145	A	mg CaCO3/L	P387829	
	SM 2540 C-2011	TDS	233		mg/L	P388197	
	SM 2540 D-2011	TSS	2	I	mg/L	P388200	
	SM 4500 F-C-2011	Fluoride	0.16		mg F/L	P387832	
2197722	EPA 350.1 Rev. 2.0	Ammonia-N	0.18		mg N/L	P387888	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.77		mg N/L	P387922	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.30		mg N/L	P387789	
	EPA 365.1 Rev. 2.0	Total-P	0.064		mg P/L	P387967	MS
	SM 5310 B-2011	Organic Carbon	8.1		mg C/L	P387982	
2197724	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.029		mg P/L	P387761	

Ref. Method and Comment:

EPA 351.2 Rev. 2.0: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Sample Location: Trout River at Old Kings Rd

Collection Date/Time: 09/22/2020 10:35

Field ID: 20030753

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2197721	EPA 180.1 Rev. 2.0	Turbidity	2.4		NTU	P387753	
	EPA 300.0 Rev. 2.1	Chloride	13		mg Cl/L	P388092	
		Sulfate	1.7		mg SO4/L	P388096	
	SM 2120 B-2011	Color (true)	440		PCU	P387764	
	SM 2320 B-2011	Total Alkalinity	23		mg CaCO3/L	P387829	
	SM 2540 C-2011	TDS	126		mg/L	P388197	
	SM 2540 D-2011	TSS	3	I	mg/L	P388200	
	SM 4500 F-C-2011	Fluoride	0.069	I	mg F/L	P387832	
2197723	EPA 350.1 Rev. 2.0	Ammonia-N	0.013		mg N/L	P387888	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.3		mg N/L	P387922	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.014		mg N/L	P387789	
	EPA 365.1 Rev. 2.0	Total-P	0.12		mg P/L	P387967	MS
	SM 5310 B-2011	Organic Carbon	39		mg C/L	P387982	
2197725	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.078		mg P/L	P387761	

Ref. Method and Comment:

EPA 351.2 Rev. 2.0: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Sample Location: Nw Tributary to Open Cr At Hodges

Collection Date/Time: 09/22/2020 11:50

Field ID: 20030848

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2197717	EPA 180.1 Rev. 2.0	Turbidity	2.7		NTU	P387753	
	EPA 300.0 Rev. 2.1	Chloride	42		mg Cl/L	P388030	
		Sulfate	61		mg SO4/L	P388032	
	SM 2120 B-2011	Color (true)	82	A	PCU	P387764	
	SM 2320 B-2011	Total Alkalinity	59	A	mg CaCO3/L	P387828	
	SM 2540 C-2011	TDS	232		mg/L	P388197	
	SM 2540 D-2011	TSS	2	I	mg/L	P388200	
2197718	SM 4500 F-C-2011	Fluoride	0.093	I	mg F/L	P387943	
	EPA 350.1 Rev. 2.0	Ammonia-N	0.11		mg N/L	P387888	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.88		mg N/L	P387922	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.29		mg N/L	P387789	
	EPA 365.1 Rev. 2.0	Total-P	0.068		mg P/L	P387967	MS
2197719	SM 5310 B-2011	Organic Carbon	15		mg C/L	P387982	
	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.030		mg P/L	P387761	
2197732	EPA 8321B	2,4-D	0.027		ug/L	P387668	
		Acesulfame K**	0.10	U	ug/L	P387746	
		AMPA**	0.22	I	ug/L	P387746	
		Sucralose**	0.54		ug/L	P387746	
		Acetaminophen**	0.0080	U	ug/L	P387668	
		Acetamiprid**	0.0020	U	ug/L	P387668	
		Endothall**	0.25	U	ug/L	P387746	
		Glufosinate**	0.10	U	ug/L	P387746	
		Glyphosate**	1.5		ug/L	P387746	
		Afidopyrophen**	0.0020	U	ug/L	P387668	
		Bentazon	0.044		ug/L	P387668	
		Benzovindiflupyr**	0.0020	U	ug/L	P387668	
		Carbamazepine**	8.0E-04	U	ug/L	P387668	
		Clothianidin**	0.030		ug/L	P387668	
		Dinotefuran**	0.0020	U	ug/L	P387668	
		Diuron	0.0020	U	ug/L	P387668	
		Fenuron	0.021	I	ug/L	P387668	
		Fluridone**	0.071		ug/L	P387668	
		Imazapyr**	0.62		ug/L	P387668	
		Hydrocodone**	0.0020	U	ug/L	P387668	
		Ibuprofen**	0.020	U	ug/L	P387668	
		Imidacloprid	0.047		ug/L	P387668	
		Linuron	0.0040	U	ug/L	P387668	
		Mandestrobin**	0.0020	U	ug/L	P387668	
		MCP	0.0066	I	ug/L	P387668	
		Naproxen**	0.0080	U	ug/L	P387668	
		Primidone**	0.0040	U	ug/L	P387668	
		Pyraclostrobin**	0.0020	U	ug/L	P387668	
		Thiamethoxam**	0.0020	U	ug/L	P387668	

Field ID: 20030848

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2197732	EPA 8321B	Tolfenpyrad**	0.0020	U	ug/L	P387668	
		Triclopyr**	0.0040	U	ug/L	P387668	
2197733	EPA 200.7 Rev. 4.4	Barium	37.5		ug/L	P387926	
		Calcium	41.9		mg/L	P387926	
		Iron	500		ug/L	P387926	
		Magnesium	8.27		mg/L	P387926	
		Potassium	4.4		mg/L	P387926	
		Sodium	24.0		mg/L	P387926	
		Zinc	5.0	U	ug/L	P387926	
	EPA 200.8 Rev. 5.4	Aluminum	210		ug/L	P387926	
		Antimony	0.061	I	ug/L	P387926	
		Arsenic	1.18		ug/L	P387926	
		Boron**	32.1	J	ug/L	P387926	
		Cadmium	0.020	U	ug/L	P387926	
		Chromium	0.46	I	ug/L	P387926	
		Copper	4.77		ug/L	P387926	
		Lead	0.20	U	ug/L	P387926	
		Nickel	0.50	U	ug/L	P387926	
		Selenium	0.20	U	ug/L	P387926	
		Silver	0.010	U	ug/L	P387926	
		Thallium	0.10	U	ug/L	P387926	
	SM 2340B	Hardness	139		mg/L	P387926	

Ref. Method and Comment:

EPA 351.2 Rev. 2.0: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

EPA 8321B: MS accuracy for some analytes could not be assessed due to high concentration of parameter in the spiked sample.

EPA 200.8 Rev. 5.4: The sample was preserved improperly for boron.

Non-Conformance Report

NCR ID: 8440

Event(s)

NE-DIST-2020-09-23-01

Job(s)

TLH-2020-09-23-55

Sample(s)

2197733

Test(s)

NCR Type: SHIPPING/RECEIVING

NCR Category: Preservation Not Intact

Observation: Sample not preserved to pH < 2. (for metals analysis)

Resolution: Sample preserved in the laboratory with nitric acid 09/23/2020 at 12:00.
HNO3 Lot# NA9354060 | Expiration Date: 12/20/2020
Data will be qualified as appropriate.
Customer notified by email.
Supervisor notified by email.

Authorized by/Date: Joshua Ayres 9/29/2020

The Non-Conformance Report details exceptions or problems encountered with the events/jobs/samples/test.
Please address questions to:

Chemistry	Colin Wright	(850) 245-8085
Biology	Cheryl Swanson	(850) 245-8177

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P387753

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P387926

Component	Result	Code	Units
Barium	1.0	U	ug/L
Calcium	0.075	U	mg/L
Iron	30	U	ug/L
Magnesium	0.040	U	mg/L
Potassium	0.30	U	mg/L
Sodium	0.50	U	mg/L
Zinc	5.0	U	ug/L

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P387926

Component	Result	Code	Units
Aluminum	5.0	U	ug/L
Antimony	0.050	U	ug/L
Arsenic	0.050	U	ug/L
Boron	2.0	U	ug/L
Cadmium	0.020	U	ug/L
Chromium	0.40	U	ug/L
Copper	0.40	U	ug/L
Lead	0.20	U	ug/L
Nickel	0.50	U	ug/L
Selenium	0.20	U	ug/L
Silver	0.010	U	ug/L
Thallium	0.10	U	ug/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P388030

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P388032

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P388092

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P388096

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P387888

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P387922

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P387789

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P387967

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P387761

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: EPA 8321B
Batch ID: P387668

Component	Result	Code	Units
2,4-D	0.0020	U	ug/L
Acetaminophen	0.0080	U	ug/L
Acetamiprid	0.0020	U	ug/L
Afidopyropen	0.0020	U	ug/L
Bentazon	8.0E-04	U	ug/L
Benzovindiflupyr	0.0020	U	ug/L
Carbamazepine	8.0E-04	U	ug/L
Clothianidin	0.0020	U	ug/L
Dinotefuran	0.0020	U	ug/L
Diuron	0.0020	U	ug/L
Fenuron	0.016	U	ug/L
Fluridone	8.0E-04	U	ug/L
Hydrocodone	0.0020	U	ug/L
Ibuprofen	0.020	U	ug/L
Imazapyr	0.0040	U	ug/L
Imidacloprid	0.0020	U	ug/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 8321B
Batch ID: P387668

Component	Result	Code	Units
Linuron	0.0040	U	ug/L
Mandestrobin	0.0020	U	ug/L
MCP	0.0020	U	ug/L
Naproxen	0.0080	U	ug/L
Primidone	0.0040	U	ug/L
Pyraclostrobin	0.0020	U	ug/L
Thiamethoxam	0.0020	U	ug/L
Tolfenpyrad	0.0020	U	ug/L
Triclopyr	0.0040	U	ug/L

Reference Method: EPA 8321B
Batch ID: P387746

Component	Result	Code	Units
Acesulfame K	0.10	U	ug/L
AMPA	0.10	U	ug/L
Endothall	0.25	U	ug/L
Glufosinate	0.10	U	ug/L
Glyphosate	0.10	U	ug/L
Sucralose	0.010	U	ug/L

Reference Method: SM 2120 B-2011
Batch ID: P387764

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-2011
Batch ID: P387828

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2320 B-2011
Batch ID: P387829

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 C-2011
Batch ID: P388197

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-2011
Batch ID: P388200

Component	Result	Code	Units
TSS	2	U	mg/L

Quality Assurance Report

Method Blank Results

Reference Method: SM 4500 F-C-2011
Batch ID: P387832

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 4500 F-C-2011
Batch ID: P387943

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-2011
Batch ID: P387982

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P387926

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Barium	99.3		P	85 - 115
Calcium	107		P	85 - 115
Iron	106		P	85 - 115
Magnesium	105		P	85 - 115
Potassium	104		P	85 - 115
Sodium	102		P	85 - 115
Zinc	94.7		P	85 - 115

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P387926

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Aluminum	98.4		P	80 - 120
Antimony	102		P	80 - 120
Arsenic	100		P	80 - 120
Boron	96.2		P	80 - 120
Cadmium	96.8		P	80 - 120
Chromium	99.7		P	80 - 120
Copper	99.0		P	80 - 120
Lead	97.4		P	80 - 120
Nickel	99.2		P	80 - 120
Selenium	99.6		P	80 - 120
Silver	96.7		P	80 - 120
Thallium	97.9		P	80 - 120

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P388030

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	101		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P388032

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	98.8		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P388092

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	105		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P388096

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	103		P	90 - 110

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P387888

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	97.3		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P387922

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	101		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P387789

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	98.5		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P387967

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	109		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P387761

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	100		P	90 - 110

Reference Method: EPA 8321B
Batch ID: P387668

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2,4-D	97.8		P	30 - 160
Acetaminophen	70.1		P	30 - 160
Acetamiprid	67.5		P	30 - 160
Afidopyropen	55.4		P	30 - 160
Bentazon	53.6		P	30 - 160
Benzovindiflupyr	53.1		P	30 - 160
Carbamazepine	56.4		P	30 - 160
Clothianidin	69.8		P	30 - 160
Dinotefuran	80.4		P	30 - 160
Diuron	47.0		P	30 - 160
Fenuron	82.8		P	30 - 160
Fluridone	69.4		P	30 - 160
Hydrocodone	59.9		P	30 - 160
Ibuprofen	53.9		P	30 - 160
Imazapyr	75.6		P	30 - 160
Imidacloprid	86.5		P	30 - 160
Linuron	59.3		P	30 - 160
Mandestrobin	66.1		P	30 - 160
MCPP	97.5		P	30 - 160
Naproxen	75.2		P	30 - 160
Primidone	67.8		P	30 - 160
Pyraclostrobin	53.7		P	30 - 160

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 8321B
Batch ID: P387668

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Thiamethoxam	69.2		P	30 - 160
Tolfenpyrad	35.4		P	30 - 160
Triclopyr	70.8		P	30 - 160

Reference Method: EPA 8321B
Batch ID: P387746

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Acesulfame K	98.1		P	40 - 150
AMPA	82.5		P	40 - 150
Endothall	94.7		P	40 - 150
Glufosinate	89.8		P	40 - 150
Glyphosate	101		P	40 - 140
Sucralose	95.8		P	40 - 150

Reference Method: SM 2120 B-2011
Batch ID: P387764

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	98.7		P	85 - 115

Reference Method: SM 2320 B-2011
Batch ID: P387828

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	101		P	93 - 104

Reference Method: SM 2320 B-2011
Batch ID: P387829

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	98.8		P	93 - 104

Reference Method: SM 4500 F-C-2011
Batch ID: P387832

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	98.1		P	90 - 106

Reference Method: SM 4500 F-C-2011
Batch ID: P387943

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	100		P	90 - 106

Reference Method: SM 5310 B-2011
Batch ID: P387982

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	102		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P387926

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2197555	Barium	103	103	P/P	80 - 120
2197555	Calcium	110	103	P/P	80 - 120
2197555	Iron	111	108	P/P	80 - 120
2197555	Magnesium	109	106	P/P	80 - 120
2197555	Potassium	108	105	P/P	80 - 120
2197555	Sodium	113	107	P/P	80 - 120
2197555	Zinc	96.0	95.9	P/P	80 - 120
2197741	Barium	103		P	80 - 120
2197741	Calcium	105		P	80 - 120
2197741	Iron	113		P	80 - 120
2197741	Magnesium	111		P	80 - 120
2197741	Potassium	111		P	80 - 120
2197741	Sodium	119		P	80 - 120
2197741	Zinc	94.2		P	80 - 120

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P387926

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2197555	Aluminum	98.1	97.4	P/P	80 - 120
2197555	Antimony	101	101	P/P	80 - 120
2197555	Arsenic	99.3	98.9	P/P	80 - 120
2197555	Boron	95.7	95.2	P/P	80 - 120
2197555	Cadmium	96.9	96.6	P/P	80 - 120
2197555	Chromium	99.1	98.9	P/P	80 - 120
2197555	Copper	98.3	98.1	P/P	80 - 120
2197555	Lead	97.8	96.9	P/P	80 - 120
2197555	Nickel	97.8	97.7	P/P	80 - 120
2197555	Selenium	99.4	97.9	P/P	80 - 120
2197555	Silver	97.3	97.0	P/P	80 - 120
2197555	Thallium	98.5	99.4	P/P	80 - 120
2197741	Aluminum	93.9		P	80 - 120
2197741	Antimony	101		P	80 - 120
2197741	Arsenic	96.7		P	80 - 120
2197741	Boron	94.3		P	80 - 120
2197741	Cadmium	96.2		P	80 - 120
2197741	Chromium	96.3		P	80 - 120
2197741	Copper	95.4		P	80 - 120
2197741	Lead	94.1		P	80 - 120
2197741	Nickel	95.0		P	80 - 120
2197741	Selenium	95.8		P	80 - 120
2197741	Silver	94.8		P	80 - 120
2197741	Thallium	95.2		P	80 - 120

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P388030

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2197327	Chloride	103		P	90 - 110
2197587	Chloride	104		P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P388032

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2197327	Sulfate	100		P	90 - 110
2197587	Sulfate	98.4		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P388092

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2197720	Chloride	103		P	90 - 110
2197931	Chloride	103		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P388096

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2197720	Sulfate	98.9		P	90 - 110
2197931	Sulfate	100		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P387888

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2197738	Ammonia-N	97.8	97.1	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P387789

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2197709	NO2NO3-N	95.1	94.6	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P387967

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2197709	Total-P	113	111	F/F	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P387761

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2197719	O-Phosphate-P	99.8	98.3	P/P	90 - 110

Reference Method: EPA 8321B
Batch ID: P387668

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2197057	Acetaminophen	77.2	72.4	P/P	30 - 160
2197057	Acetamiprid	65.3	59.4	P/P	30 - 160
2197057	Afidopyropen	54.6	50.4	P/P	30 - 160
2197057	Benzovindiflupyr	60.5	57.8	P/P	30 - 160
2197057	Carbamazepine	49.7	46.7	P/P	30 - 160
2197057	Clothianidin	61.0	61.2	P/P	30 - 160
2197057	Dinotefuran	89.2	81.2	P/P	30 - 160

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 8321B
Batch ID: P387668

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2197057	Diuron	39.4	41.2	P/P	30 - 160
2197057	Fenuron	75.4	73.8	P/P	30 - 160
2197057	Hydrocodone	70.1	66.4	P/P	30 - 160
2197057	Ibuprofen	52.5	50.9	P/P	30 - 160
2197057	Linuron	50.4	54.0	P/P	30 - 160
2197057	Mandestrobin	65.5	66.3	P/P	30 - 160
2197057	MCCP	94.0	86.9	P/P	30 - 160
2197057	Naproxen	116	109	P/P	30 - 160
2197057	Primidone	75.0	71.4	P/P	30 - 160
2197057	Pyraclostrobin	59.9	54.1	P/P	30 - 160
2197057	Thiamethoxam	88.4	86.8	P/P	30 - 160
2197057	Tolfenpyrad	44.2	38.0	P/P	30 - 160
2197057	Triclopyr	72.9	78.7	P/P	30 - 160

Reference Method: EPA 8321B
Batch ID: P387746

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2197190	Acesulfame K	114	119	P/P	40 - 150
2197190	AMPA	114	117	P/P	40 - 150
2197190	Endothall	107	93.8	P/P	40 - 150
2197190	Glufosinate	83.8	94.5	P/P	40 - 150
2197190	Glyphosate	82.6	80.4	P/P	40 - 140
2197190	Sucralose	99.1	102	P/P	40 - 150

Reference Method: SM 4500 F-C-2011
Batch ID: P387832

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2197720	Fluoride	95.2	95.8	P/P	94 - 107

Reference Method: SM 4500 F-C-2011
Batch ID: P387943

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2197717	Fluoride	100	100	P/P	94 - 107

Reference Method: SM 5310 B-2011
Batch ID: P387982

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2197863	Organic Carbon	100	99.5	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P387753

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197686	Turbidity	6.87	Sample	P	0 - 20

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P387926

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197555	Barium	0.228	Spike	P	0 - 20
2197555	Calcium	4.51	Spike	P	0 - 20
2197555	Iron	1.97	Spike	P	0 - 20
2197555	Magnesium	1.97	Spike	P	0 - 20
2197555	Potassium	3.66	Spike	P	0 - 20
2197555	Sodium	4.15	Spike	P	0 - 20
2197555	Zinc	0.132	Spike	P	0 - 20

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P387926

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197555	Aluminum	0.688	Spike	P	0 - 20
2197555	Antimony	0.0840	Spike	P	0 - 20
2197555	Arsenic	0.402	Spike	P	0 - 20
2197555	Boron	0.479	Spike	P	0 - 20
2197555	Cadmium	0.379	Spike	P	0 - 20
2197555	Chromium	0.187	Spike	P	0 - 20
2197555	Copper	0.189	Spike	P	0 - 20
2197555	Lead	0.953	Spike	P	0 - 20
2197555	Nickel	0.111	Spike	P	0 - 20
2197555	Selenium	1.57	Spike	P	0 - 20
2197555	Silver	0.312	Spike	P	0 - 20
2197555	Thallium	0.952	Spike	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P388030

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197327	Chloride	0.509	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P388032

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197327	Sulfate	0.774	Sample	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P388092

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197720	Chloride	0.719	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P388096

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197720	Sulfate	0.320	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P387888

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197738	Ammonia-N	0.603	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P387922

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197710	Kjeldahl Nitrogen	1.28	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P387789

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197709	NO2NO3-N	0.482	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P387967

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197709	Total-P	1.46	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P387761

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197719	O-Phosphate-P	1.16	Spike	P	0 - 20

Reference Method: EPA 8321B
Batch ID: P387668

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197057	2,4-D	4.75	Spike	P	0 - 30
2197057	Acetaminophen	6.44	Spike	P	0 - 30

Quality Assurance Report Precision

Reference Method: EPA 8321B
Batch ID: P387668

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197057	Acetamiprid	9.37	Spike	P	0 - 30
2197057	Afidopyropen	8.08	Spike	P	0 - 30
2197057	Bentazon	11.1	Spike	P	0 - 30
2197057	Benzovindiflupyr	4.71	Spike	P	0 - 30
2197057	Carbamazepine	6.15	Spike	P	0 - 30
2197057	Clothianidin	0.184	Spike	P	0 - 30
2197057	Dinotefuran	9.31	Spike	P	0 - 30
2197057	Diuron	3.77	Spike	P	0 - 30
2197057	Fenuron	2.08	Spike	P	0 - 30
2197057	Fluridone	5.52	Spike	P	0 - 30
2197057	Hydrocodone	5.47	Spike	P	0 - 30
2197057	Ibuprofen	3.12	Spike	P	0 - 30
2197057	Imazapyr	7.01	Spike	P	0 - 30
2197057	Imidacloprid	3.58	Spike	P	0 - 30
2197057	Linuron	6.85	Spike	P	0 - 30
2197057	Mandestrobin	1.24	Spike	P	0 - 30
2197057	MCPP	7.90	Spike	P	0 - 30
2197057	Naproxen	6.51	Spike	P	0 - 30
2197057	Primidone	4.95	Spike	P	0 - 30
2197057	Pyraclostrobin	10.3	Spike	P	0 - 30
2197057	Thiamethoxam	1.64	Spike	P	0 - 30
2197057	Tolfenpyrad	15.3	Spike	P	0 - 30
2197057	Triclopyr	7.62	Spike	P	0 - 30

Reference Method: EPA 8321B
Batch ID: P387746

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197190	Acesulfame K	4.24	Spike	P	0 - 30
2197190	AMPA	1.86	Spike	P	0 - 30
2197190	Endothall	13.0	Spike	P	0 - 30
2197190	Glufosinate	12.0	Spike	P	0 - 30
2197190	Glyphosate	2.62	Spike	P	0 - 30
2197190	Sucralose	2.90	Spike	P	0 - 30

Reference Method: SM 2120 B-2011
Batch ID: P387764

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197717	Color (true)	1.42	Sample	P	0 - 20

Reference Method: SM 2320 B-2011
Batch ID: P387828

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197717	Total Alkalinity	0.488	Sample	P	0 - 8.1

Quality Assurance Report Precision

Reference Method: SM 2320 B-2011
 Batch ID: P387829

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197720	Total Alkalinity	0.385	Sample	P	0 - 8.1

Reference Method: SM 2540 C-2011
 Batch ID: P388197

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197862	TDS	2.11	Sample	P	0 - 10

Reference Method: SM 2540 D-2011
 Batch ID: P388200

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197862	TSS	2.41	Sample	P	0 - 10

Reference Method: SM 4500 F-C-2011
 Batch ID: P387832

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197720	Fluoride	0.467	Spike	P	0 - 2.8

Reference Method: SM 4500 F-C-2011
 Batch ID: P387943

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197717	Fluoride	0.0	Spike	P	0 - 2.8

Reference Method: SM 5310 B-2011
 Batch ID: P387982

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2197863	Organic Carbon	0.401	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Surrogates

Lab Sample ID: 2197732
Field Sample ID: 20030848

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8321B	2,4-D-d3	72.4	P	30 - 160
EPA 8321B	Acetaminophen-d4	64.9	P	30 - 160
EPA 8321B	Carbamazepine-d10	48.6	P	30 - 160
EPA 8321B	Diuron-d6	52.1	P	30 - 160
EPA 8321B	Endothall-d6	113	P	30 - 160
EPA 8321B	Endothall-d6	113	P	30 - 160
EPA 8321B	Glyphosate 13C 15N	84.8	P	30 - 160
EPA 8321B	Glyphosate 13C 15N	84.8	P	30 - 160
EPA 8321B	Ibuprofen-d3	72.4	P	30 - 160
EPA 8321B	Primidone-d5	75.2	P	30 - 160
EPA 8321B	Sucralose-d6	101	P	30 - 160
EPA 8321B	Sucralose-d6	101	P	30 - 160

Quality Assurance Report Calibration Verification

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A101160

Included Lab Sample IDs: 2197717, 2197720, 2197721

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	100	P/P	95 - 105

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A101162

Included Lab Sample IDs: 2197719, 2197724, 2197725

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	100	101	P/P	90 - 110
O-Phosphate-P	101	100	P/P	90 - 110

Reference Method: SM 2120 B-2011

Run ID: A101163

Included Lab Sample IDs: 2197717, 2197720, 2197721

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	99.4	99.7	P/P	90 - 115

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A101175

Included Lab Sample IDs: 2197718, 2197722, 2197723

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	97.5	97.5	P/P	90 - 110

Reference Method: EPA 8321B

Run ID: A101191

Included Lab Sample IDs: 2197732

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Acesulfame K	90.8	93.8	P/P	60 - 160
Endothall	112	109	P/P	60 - 160

Reference Method: SM 2320 B-2011

Run ID: A101193

Included Lab Sample IDs: 2197717, 2197720, 2197721

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	100	101	P/P	90 - 110
Total Alkalinity	98.0	99.7	P/P	90 - 110

Reference Method: SM 4500 F-C-2011

Run ID: A101193

Included Lab Sample IDs: 2197720, 2197721

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	100	100	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A101210

Included Lab Sample IDs: 2197718, 2197722, 2197723

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	98.6	99.0	P/P	90 - 110
Ammonia-N	99.0	95.9	P/P	90 - 110

Reference Method: EPA 8321B

Run ID: A101224

Included Lab Sample IDs: 2197732

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
AMPA	71.8	74.8	P/P	60 - 160
Glufosinate	94.0	99.4	P/P	60 - 160
Glyphosate	97.5	98.9	P/P	60 - 160

Reference Method: SM 4500 F-C-2011

Run ID: A101239

Included Lab Sample IDs: 2197717

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	104	104	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A101254

Included Lab Sample IDs: 2197718, 2197722, 2197723

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	104	100	P/P	90 - 110

Reference Method: SM 5310 B-2011

Run ID: A101262

Included Lab Sample IDs: 2197718, 2197722, 2197723

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	99.6	99.6	P/P	90 - 110

Reference Method: EPA 200.7 Rev. 4.4

Run ID: A101267

Included Lab Sample IDs: 2197733

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Barium	103	102	P/P	90 - 110
Calcium	104	104	P/P	90 - 110
Iron	107	107	P/P	90 - 110
Magnesium	104	104	P/P	90 - 110
Potassium	104	104	P/P	90 - 110
Sodium	108	108	P/P	90 - 110
Zinc	96.6	95.9	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A101281

Included Lab Sample IDs: 2197717, 2197720, 2197721

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
-----------	---------	---------	------------	----------------

Quality Assurance Report Calibration Verification

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A101281

Included Lab Sample IDs: 2197717, 2197720, 2197721

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	102	102	P/P	90 - 110
Chloride	106	107	P/P	90 - 110
Sulfate	100	100	P/P	90 - 110
Sulfate	104	105	P/P	90 - 110

Reference Method: EPA 8321B

Run ID: A101289

Included Lab Sample IDs: 2197732

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
2,4-D	123	134	P/P	50 - 150
Acetaminophen	125	124	P/P	60 - 160
Acetamiprid	122	126	P/P	50 - 150
Afidopyropen	95.6	102	P/P	50 - 150
Bentazon	106	102	P/P	50 - 160
Benzovindiflupyr	107	110	P/P	50 - 150
Carbamazepine	103	110	P/P	60 - 150
Clothianidin	117	124	P/P	60 - 150
Dinotefuran	111	122	P/P	50 - 150
Diuron	104	113	P/P	60 - 160
Fenuron	121	122	P/P	60 - 150
Fluridone	113	120	P/P	60 - 160
Hydrocodone	113	119	P/P	60 - 150
Ibuprofen	87.5	74.3	P/P	50 - 160
Imazapyr	130	115	P/P	50 - 160
Imidacloprid	109	113	P/P	60 - 150
Linuron	112	107	P/P	60 - 150
Mandestrobin	104	109	P/P	50 - 150
MCPP	136	133	P/P	50 - 150
Naproxen	102	79.5	P/P	50 - 160
Primidone	127	115	P/P	60 - 150
Pyraclostrobin	103	115	P/P	60 - 150
Thiamethoxam	137	138	P/P	50 - 150
Tolfenpyrad	106	119	P/P	60 - 150
Triclopyr	130	133	P/P	50 - 150

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A101308

Included Lab Sample IDs: 2197718, 2197722, 2197723

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	104	103	P/P	90 - 110

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A101347

Included Lab Sample IDs: 2197733

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Aluminum	97.2	99.9	P/P	90 - 110
Antimony	103	106	P/P	90 - 110
Arsenic	96.2	97.9	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A101347

Included Lab Sample IDs: 2197733

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Boron	99.1	102	P/P	90 - 110
Cadmium	96.0	97.7	P/P	90 - 110
Chromium	97.1	97.0	P/P	90 - 110
Copper	94.6	98.2	P/P	90 - 110
Lead	96.8	97.0	P/P	90 - 110
Nickel	95.2	98.5	P/P	90 - 110
Selenium	98.6	98.6	P/P	90 - 110
Silver	96.8	98.0	P/P	90 - 110
Thallium	93.7	92.0	P/P	90 - 110

Reference Method: EPA 8321B

Run ID: A101350

Included Lab Sample IDs: 2197732

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Sucralose	101	98.7	P/P	60 - 160

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery			LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity						6.87	
EPA 200.7 Rev. 4.4	Barium	99.3	103	103	103			0.228
	Calcium	107	110	103	105			4.51
	Iron	106	111	108	113			1.97
	Magnesium	105	109	106	111			1.97
	Potassium	104	108	105	111			3.66
	Sodium	102	113	107	119			4.15
	Zinc	94.7	96.0	95.9	94.2			0.132
EPA 200.8 Rev. 5.4	Aluminum	98.4	98.1	97.4	93.9			0.688
	Antimony	102	101	101	101			0.0840
	Arsenic	100	99.3	98.9	96.7			0.402
	Boron	96.2	95.7	95.2	94.3			0.479
	Cadmium	96.8	96.9	96.6	96.2			0.379
	Chromium	99.7	99.1	98.9	96.3			0.187
	Copper	99.0	98.3	98.1	95.4			0.189
	Lead	97.4	97.8	96.9	94.1			0.953
	Nickel	99.2	97.8	97.7	95.0			0.111
	Selenium	99.6	99.4	97.9	95.8			1.57
	Silver	96.7	97.3	97.0	94.8			0.312
	Thallium	97.9	98.5	99.4	95.2			0.952
EPA 300.0 Rev. 2.1	Chloride	101	103	104			0.509	
	Chloride	105	103	103			0.719	
	Sulfate	98.8	100	98.4			0.774	
	Sulfate	103	98.9	100			0.320	
EPA 350.1 Rev. 2.0	Ammonia-N	97.3	97.8	97.1				0.603
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	101						1.28
EPA 353.2 Rev. 2.0	NO2NO3-N	98.5	95.1	94.6				0.482
EPA 365.1 Rev. 2.0	Total-P	109	113	111				1.46
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	100	99.8	98.3				1.16
EPA 8321B	2,4-D	97.8						4.75
	Acesulfame K	98.1	114	119				4.24
	Acetaminophen	70.1	77.2	72.4				6.44
	Acetamiprid	67.5	65.3	59.4				9.37
	Afidopyropen	55.4	54.6	50.4				8.08
	AMPA	82.5	114	117				1.86
	Bentazon	53.6						11.1
	Benzovindiflupyr	53.1	60.5	57.8				4.71
	Carbamazepine	56.4	49.7	46.7				6.15
	Clothianidin	69.8	61.0	61.2				0.184
	Dinotefuran	80.4	89.2	81.2				9.31
	Diuron	47.0	39.4	41.2				3.77
	Endothall	94.7	107	93.8				13.0
	Fenuron	82.8	75.4	73.8				2.08
	Fluridone	69.4						5.52
	Glufosinate	89.8	83.8	94.5				12.0
	Glyphosate	101	82.6	80.4				2.62
	Hydrocodone	59.9	70.1	66.4				5.47
	Ibuprofen	53.9	52.5	50.9				3.12
	Imazapyr	75.6						7.01
	Imidacloprid	86.5						3.58
	Linuron	59.3	50.4	54.0				6.85
	Mandestrobin	66.1	65.5	66.3				1.24
	MCPD	97.5	94.0	86.9				7.90
	Naproxen	75.2	116	109				6.51

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		Precision	
					LCS	MS
EPA 8321B	Primidone	67.8	75.0	71.4		4.95
	Pyraclostrobin	53.7	59.9	54.1		10.3
	Sucralose	95.8	99.1	102		2.90
	Thiamethoxam	69.2	88.4	86.8		1.64
	Tolfenpyrad	35.4	44.2	38.0		15.3
	Triclopyr	70.8	72.9	78.7		7.62
	Color (true)	98.7			1.42	
SM 2120 B-2011	Total Alkalinity	101			0.488	
SM 2320 B-2011	Total Alkalinity	98.8			0.385	
SM 2540 C-2011	TDS				2.11	
SM 2540 D-2011	TSS				2.41	
SM 4500 F-C-2011	Fluoride	98.1	95.2	95.8		0.467
	Fluoride	100	100	100		0.0
SM 5310 B-2011	Organic Carbon	102	100	99.5		0.401

Reference Method Descriptions

Method	Description	Associated Samples
EPA 180.1 Rev. 2.0	Turbidity in aqueous matrices	2197717, 2197720, 2197721
EPA 200.7 Rev. 4.4	Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects	2197733
EPA 200.8 Rev. 5.4	Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects	2197733
EPA 300.0 Rev. 2.1	Chloride in aqueous matrices	2197717, 2197720, 2197721
EPA 300.0 Rev. 2.1	Sulfate in aqueous matrices	2197717, 2197720, 2197721
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2197718, 2197722, 2197723
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2197718, 2197722, 2197723
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2197718, 2197722, 2197723
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2197718, 2197722, 2197723
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2197719, 2197724, 2197725
EPA 8321B	Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS	2197732
EPA 8321B	Water soluble organic compounds by HPLC/MS/MS	2197732
SM 2120 B-2011	True Color measured at 450 nm (not valid for NPDES monitoring)	2197717, 2197720, 2197721
SM 2320 B-2011	Alkalinity in aqueous matrices as mg CaCO3/L	2197717, 2197720, 2197721
SM 2340B	Hardness by calculation, in aqueous matrices, reported as mg/L CaCO3	2197733
SM 2540 C-2011	Total Dissolved Solids in aqueous matrices	2197717, 2197720, 2197721
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2197717, 2197720, 2197721
SM 4500 F-C-2011	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2197717, 2197720, 2197721
SM 5310 B-2011	Nonpurgeable Organic Carbon in aqueous matrices	2197718, 2197722, 2197723

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	09/23/2020			09/23/2020 15:04	Yen Ta	2197717, 2197720, 2197721
EPA 200.7 Rev. 4.4	09/23/2020	09/28/2020 14:35	Elliott D. Healy	09/29/2020 16:18	Betina Topolski	2197733
EPA 200.8 Rev. 5.4	09/23/2020	09/28/2020 14:35	Elliott D. Healy	10/03/2020 18:14	Alexander Thompson	2197733
EPA 300.0 Rev. 2.1	09/23/2020			09/30/2020 03:45	Elliott Rodriguez	2197717
	09/23/2020			09/30/2020 19:10	Lipi Saha	2197720
	09/23/2020			09/30/2020 20:02	Lipi Saha	2197721
EPA 350.1 Rev. 2.0	09/23/2020			09/26/2020 14:26	Ping Hua	2197718
	09/23/2020			09/26/2020 14:33	Ping Hua	2197722
	09/23/2020			09/26/2020 14:35	Ping Hua	2197723
EPA 351.2 Rev. 2.0	09/23/2020	09/28/2020 01:45	Elliott Rodriguez	09/29/2020 11:31	Elliott Rodriguez	2197718
	09/23/2020	09/28/2020 01:45	Elliott Rodriguez	09/29/2020 11:32	Elliott Rodriguez	2197722
	09/23/2020	09/28/2020 01:45	Elliott Rodriguez	09/29/2020 11:34	Elliott Rodriguez	2197723
EPA 353.2 Rev. 2.0	09/23/2020			09/24/2020 11:04	Beverly Y. Sanders	2197718
	09/23/2020			09/24/2020 11:06	Beverly Y. Sanders	2197722
	09/23/2020			09/24/2020 11:07	Beverly Y. Sanders	2197723
EPA 365.1 Rev. 2.0	09/23/2020	09/29/2020 12:36	Yen Ta	09/30/2020 14:35	Benjamin T. Nordmann	2197718
	09/23/2020	09/29/2020 12:36	Yen Ta	09/30/2020 14:36	Benjamin T. Nordmann	2197722
	09/23/2020	09/29/2020 12:36	Yen Ta	09/30/2020 14:37	Benjamin T. Nordmann	2197723
EPA 365.1 Rev. 2.0 dissolved	09/23/2020			09/23/2020 15:56	Virginia P. Brown	2197719
	09/23/2020			09/23/2020 16:07	Virginia P. Brown	2197724, 2197725
EPA 8321B	09/23/2020	09/23/2020 13:00	Hoor Shaik	09/25/2020 18:45	Juyoung Kim	2197732
	09/23/2020	09/24/2020 09:45	Rebecca Ware	09/24/2020 17:38	Rebecca Ware	2197732
	09/23/2020	09/24/2020 09:45	Rebecca Ware	09/25/2020 14:47	Rebecca Ware	2197732
	09/23/2020	09/24/2020 09:45	Rebecca Ware	09/30/2020 14:13	Rebecca Ware	2197732
SM 2120 B-2011	09/23/2020			09/23/2020 16:52	Benjamin T. Nordmann	2197717
	09/23/2020			09/23/2020 16:53	Benjamin T. Nordmann	2197720
	09/23/2020			09/23/2020 16:56	Benjamin T. Nordmann	2197721
SM 2320 B-2011	09/23/2020			09/24/2020 15:30	Dale L. Simmons	2197717
	09/23/2020			09/24/2020 21:24	Dale L. Simmons	2197720
	09/23/2020			09/24/2020 21:45	Dale L. Simmons	2197721
SM 2340B	09/23/2020			09/29/2020 16:18	Betina Topolski	2197733
SM 2540 C-2011	09/23/2020			09/28/2020 15:25	Yijie Li	2197717, 2197720, 2197721
SM 2540 D-2011	09/23/2020			09/28/2020 15:25	Yijie Li	2197717, 2197720, 2197721
SM 4500 F-C-2011	09/23/2020			09/24/2020 21:13	Dale L. Simmons	2197720
	09/23/2020			09/24/2020 21:45	Dale L. Simmons	2197721
	09/23/2020			09/25/2020 19:50	Dale L. Simmons	2197717
SM 5310 B-2011	09/23/2020			09/29/2020 03:11	David Boose	2197718
	09/23/2020			09/29/2020 03:25	David Boose	2197722
	09/23/2020			09/29/2020 03:39	David Boose	2197723