



Method/Parameter		Technician	MMDDYY	24 HH:MM											Acceptance Criteria	
DEP SOP FT 1400		Initials	Date	Time	Previous CV		Bath Type	NIST Thermometer		Meter		± 0.5 °C				
Temperature								S/N		Temp °C						
CAL	ICV	CCV	MAINT	07.16.20										Pass	Fail	
DEP SOP FT 1500		Initials	Date	Time	Baro in Hg	Water Temp °C	Saturation mg/L	Meter				± 0.3 mg/l				
Dissolved Oxygen								Baro in Hg	DO mg/L	% DO	Temp °C					
CAL	ICV	CCV	MAINT	tm	10.19.20	08:26	30.06	23.1	8.60	30.07	8.43	98.20	23.13	Pass	Fail	
CAL	ICV	CCV	MAINT	tm	10.19.20	17:18	30.00	21.5	8.85	30.01	8.69	98.91	21.82	Pass	Fail	
CAL	ICV	CCV	MAINT											Pass	Fail	
CAL	ICV	CCV	MAINT											Pass	Fail	
CAL	ICV	CCV	MAINT											Pass	Fail	
DEP SOP FT 1200		Initials	Date	Time	Standard				Meter				± 5%			
Specific Conductance					µmhos/cm	Expiration	ID	Temp °C	µmhos/cm							
CAL	ICV	CCV	MAINT	tm	10.19.20	08:33	146.9	03.21	INV200683	24.89	140.25		Pass	Fail		
CAL	ICV	CCV	MAINT	tm	10.19.20	17:27	50000	10.21	INV200121	23.91	48535		Pass	Fail		
CAL	ICV	CCV	MAINT										Pass	Fail		
CAL	ICV	CCV	MAINT										Pass	Fail		
CAL	ICV	CCV	MAINT										Pass	Fail		
DEP SOP FT 1100		Initials	Date	Time	Standard				Meter				± 0.2 SU			
pH					SU	Expiration	ID	pH mV	SU							
CAL	ICV	CCV	MAINT	tm	10.19.20	08:37	4.00	03.22	INV200556	163.6	4.07		Pass	Fail		
CAL	ICV	CCV	MAINT	tm	10.19.20	17:32	7.00	11.20	INV190041	-16.7	7.13		Pass	Fail		
CAL	ICV	CCV	MAINT	tm	10.19.20	17:49	10.00	06.21	INV200099	-193.2	10.15		Pass	Fail		
CAL	ICV	CCV	MAINT										Pass	Fail		
DEP SOP FT 1000		Initials	Date	Time	Standard				Meter				± 20 mV			
ORP					mV	Expiration	ID	Temp °C	mV							
CAL	ICV	CCV	MAINT										Pass	Fail		
CAL	ICV	CCV	MAINT										Pass	Fail		
CAL	ICV	CCV	MAINT										Pass	Fail		
CAL	ICV	CCV	MAINT										Pass	Fail		

Notes: