



Meter ID: SMARTROLL
Project: BMAP

Method/Parameter		Technician	MMDDYY	24 HH:MM	Acceptance Criteria									
DEP SOP FT 1400 Temperature		Initials	Date	Time	Previous CV		Bath Type		NIST Thermometer		Meter		± 0.5 °C	
CAL	ICV CCV MAINT				4,28.20				S/N	Temp °C	Temp °C		Pass Fail	
DEP SOP FT 1500 Dissolved Oxygen		Initials	Date	Time	Baro Temp mm Hg	Water Temp °C	Saturation mg/L	Meter					± 0.3 mg/l	
CAL	ICV CCV MAINT	tm	5.20.20	08:44	29.80	17.8	9.47	—	9.51	100.6	17.85	Pass	Fail	
CAL	ICV CCV MAINT	tm	5.20.20	16:25	29.79	17.4	9.55		9.57	100.6	17.51	Pass	Fail	
CAL	ICV CCV MAINT											Pass	Fail	
CAL	ICV CCV MAINT											Pass	Fail	
CAL	ICV CCV MAINT											Pass	Fail	
DEP SOP FT 1200 Specific Conductance		Initials	Date	Time	Standard		ID	Meter		± 5%				
CAL	ICV CCV MAINT	tm	5.20.20	08:58	1000	01.22	INV2000144	19.5	1031	Pass	Fail			
CAL	ICV CCV MAINT	tm	5.20.20	09:06	146.9	10.20	INV2000311	19.72	149.4	Pass	Fail			
CAL	ICV CCV MAINT	tm	5.20.20	16:32	50000	06.21	INV1900620	19.70	49276.5	Pass	Fail			
CAL	ICV CCV MAINT											Pass	Fail	
CAL	ICV CCV MAINT											Pass	Fail	
CAL	ICV CCV MAINT											Pass	Fail	
DEP SOP FT 1100 pH		Initials	Date	Time	Standard		ID	Meter		± 0.2 SU				
CAL	ICV CCV MAINT	tm	5.20.20	09:16	4.00	08.21	INV2000098	—	4.14	Pass	Fail			
CAL	ICV CCV MAINT	tm	5.20.20	16:43	10.00	06.21	INV2000099	—	10.16	Pass	Fail			
CAL	ICV CCV MAINT											Pass	Fail	
CAL	ICV CCV MAINT											Pass	Fail	
CAL	ICV CCV MAINT											Pass	Fail	
DEP SOP FT 1000 ORP		Initials	Date	Time	Standard		ID	Meter		± 20 mV				
CAL	ICV CCV MAINT				mV	Expiration	ID	Temp °C	mV					
CAL	ICV CCV MAINT											Pass	Fail	
CAL	ICV CCV MAINT											Pass	Fail	
CAL	ICV CCV MAINT											Pass	Fail	
CAL	ICV CCV MAINT											Pass	Fail	

Notes: