



Method/Parameter				Technician	MMDDYY	24 HH:MM							Acceptance Criteria		
DEP SOP FT 1400 Temperature				Initials	Date	Time	Previous CV	Bath Type	NIST Thermometer		Meter		± 0.5 °C		
									S/N	Temp °C	Temp °C				
CAL	ICV	CCV	MAINT	4.28.20									Pass	Fail	
DEP SOP FT 1500 Dissolved Oxygen				Initials	Date	Time	Baro Temp mm Hg	Water Temp °C	Saturation mg/L	Meter				± 0.3 mg/l	
										Baro mmHg	DO mg/L	% DO	Temp °C		
CAL	ICV	CCV	MAINT	tm	06.08.20	08:15	29.71	18.9	9.22	—	9.26	100.6	19.00	Pass	Fail
CAL	ICV	CCV	MAINT	tm	06.08.20	15:57	29.76	19.4	9.15	—	9.15	100.5	19.67	Pass	Fail
CAL	ICV	CCV	MAINT											Pass	Fail
CAL	ICV	CCV	MAINT											Pass	Fail
CAL	ICV	CCV	MAINT											Pass	Fail
CAL	ICV	CCV	MAINT											Pass	Fail
DEP SOP FT 1200 Specific Conductance				Initials	Date	Time	Standard			Meter		± 5%			
							µmhos/cm	Expiration	ID	Temp °C	µmhos/cm				
CAL	ICV	CCV	MAINT	tm	06.08.20	08:19	1000	01.22	INV200144					Pass	Fail
CAL	ICV	CCV	MAINT	tm	06.08.20	08:28	146.9	10.20	INV200311	20.65	148.3			Pass	Fail
CAL	ICV	CCV	MAINT	tm	06.08.20	15:59	50000	10.21	INV200121	21.04	47757.7			Pass	Fail
CAL	ICV	CCV	MAINT											Pass	Fail
CAL	ICV	CCV	MAINT											Pass	Fail
CAL	ICV	CCV	MAINT											Pass	Fail
DEP SOP FT 1100 pH				Initials	Date	Time	Standard			Meter		± 0.2 SU			
							SU	Expiration	ID	pH mV	SU				
CAL	ICV	CCV	MAINT	tm	06.08.20	08:31	4.00	INV200098	08.21	—	4.00			Pass	Fail
CAL	ICV	CCV	MAINT	tm	06.08.20	16:05	10.00	INV200099	06.21	—	10.13			Pass	Fail
CAL	ICV	CCV	MAINT											Pass	Fail
CAL	ICV	CCV	MAINT											Pass	Fail
CAL	ICV	CCV	MAINT											Pass	Fail
DEP SOP FT 1000 ORP				Initials	Date	Time	Standard			Meter		± 20 mV			
							mV	Expiration	ID	Temp °C	mV				
CAL	ICV	CCV	MAINT											Pass	Fail
CAL	ICV	CCV	MAINT											Pass	Fail
CAL	ICV	CCV	MAINT											Pass	Fail
CAL	ICV	CCV	MAINT											Pass	Fail

Notes: