

537400

Method/Parameter		Technician	MMDDYY	24 HH:MM	Previous CV		Bath Type	NIST Thermometer		Acceptance Criteria					
DEP SOP FT 1400		Initials	Date	Time				S/N		Temp °C	± 0.5 °C				
Temperature															
CAL	ICV	CCV	MAINT									Pass	Fail		
DEP SOP FT 1500				Initials	Date	Time	Baro in Hg	Water Temp °C	Saturation mg/L	Meter					
Dissolved Oxygen															
CAL	ICV	CCV	MAINT	RC	7-13-20	9:15	29.85	19.8	9.11	29.89	29.91 87.93	19.92	Pass	Fail	
CAL	ICV	CCV	MAINT										Pass	Fail	
CAL	ICV	CCV	MAINT	RC	7-13-20	1657	29.81	20.1	9.04	29.84	8.75	97.98	20.78	Pass	Fail
CAL	ICV	CCV	MAINT											Pass	Fail
CAL	ICV	CCV	MAINT											Pass	Fail
DEP SOP FT 1200				Initials	Date	Time	Standard			Meter					
Specific Conductance															
CAL	ICV	CCV	MAINT	RC	7-13-20	9:20	146.9	10.20		INV2000311	22.12	141.34	Pass	Fail	
CAL	ICV	CCV	MAINT	RC	7-13-20	1702	50.000	10-21		INV200121	22.00	19.162	Pass	Fail	
CAL	ICV	CCV	MAINT											Pass	Fail
DEP SOP FT 1100				Initials	Date	Time	Standard			Meter					
pH															
CAL	ICV	CCV	MAINT	RC	7-13-20	9:27	4.00	08.21		INV200098	161.6	4.03	Pass	Fail	
CAL	ICV	CCV	MAINT	RC	7-13-20	1705	10.05	1-22		INV200394	-191.3	9.94	Pass	Fail	
CAL	ICV	CCV	MAINT											Pass	Fail
DEP SOP FT 1000				Initials	Date	Time	Standard			Meter					
ORP															
CAL	ICV	CCV	MAINT				mV	Expiration	ID	Temp °C		mV	Pass	Fail	
CAL	ICV	CCV	MAINT										Pass	Fail	
CAL	ICV	CCV	MAINT										Pass	Fail	
CAL	ICV	CCV	MAINT										Pass	Fail	
CAL	ICV	CCV	MAINT										Pass	Fail	

Notes: