

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Stitch

RQ- 2021-01-04-41

Project: RMAP 2 + SMC

2021-01-04-42

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 10/14/20

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Kim Appleby	1/7/21	941	21.0	764.3	8.9	-	100.6	-	1.0337	(P) F	(L) F
ICV	↓	↓	1000	21.4	764.2	8.9	8.92	100.7	-	1.0337	(P) F	(L) F
CCV	↓	↓	1534	19.5	759.6	9.2	4.34	101.5	-	1.0337	(P) F	(L) F
CCV											P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard µmhos/cm	Meter Reading µmhos/cm	Pass / Fail	Lab / Field
Calibr.	Kim Appleby	1/7/21	1006	201222B	1/2022	1000	1000	(P) F	(L) F
ICV	↓	↓	1007	200902A	9/2021	100	103.8	(P) F	(L) F
CCV	↓	↓	1536	200407B	4/2021	20.000	19196	(P) F	(L) F
CCV								P / F	L / F

Conductivity Acceptance criteria ± 5%

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Kim Appleby	1/7/21	1009	2006F52	6/2022	7.00	21.8	7.00	-10.4	(P) F	(L) F
Calibr.	↓	↓	1011	2006B06	5/2022	4.00	22.2	4.00	165.6	(P) F	(L) F
Calibr.	↓	↓	1013	2006E28	12/2021	10.00	21.7	10.00	-183.1	(P) F	(L) F
ICV	↓	↓	1014	2006F52	6/2022	7.00	21.8	7.03	-10.9	(P) F	(L) F
CCV	↓	↓	1538	2006E28	12/21	10.00	22.0	10.03	-185.4	(P) F	(L) F
CCV	↓	↓	1540	2006B06	5/22	4.00	22.9	4.07	163.8	(P) F	(L) F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50; mV pH 4 Range +180 ± 50; mV pH 10 Range -180 ± 50;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 10/14/20

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Kim Appleby	1/7/21	1015	0.00	0.00	(P) F	(L) F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria ± 5 % or ± 0.05m, whichever is greater.

COMMENTS:



Florida DEP North East District ROC BMAP Field sheet

 Data Pass CCV?
 Yes ☒ No ☐

 Date: 1/7/2021 RQ#: 2021-01-04-42

 Event Name: BMAP1 Meter ID: 51764

 Sampling Team Members: B Davis K Appleby

All bacteria samples taken are surface water grab samples.

WIN Number / Site Name <u>Butcher Pen Cr At Wescomett</u> <u>20030760</u>					WBID Number <u>2332</u>		Matrix: Surface ☒ Fresh ☒ Surface Salt ☐		Sample Type: E. coli ☒ Enterococcus ☐ Fecal ☐		Initials Sample & Preservation <u>BD</u>		Initials Documentation <u>KA</u>	
					Sample Access: Wading ☐ Dock ☐ From Shore ☒ Bridge ☐				Sample Ice Preserved in 15 minutes? Yes ☒ No ☐					
Time	Total depth(m)	Sample depth(m)	Temp °C	Conductivity µmhos	Salinity ppt	pH s.u.	D.O. mg/L	D.O. % sat	Secchi m	Tide Ebb/ Flood Slack				
<u>1025</u>	<u>1.1</u>	<u>0.3</u>	<u>12.7</u>	<u>358.8</u>	<u>0.17</u>	<u>7.14</u>	<u>7.22</u>	<u>68.1</u>	<u>1.1 vol</u>	<u>Ebb</u>				
Comments and Weather:														

WIN Number / Site Name <u>Bis fishweir ^{CR} Creek SF @</u> <u>Hamilton St</u> <u>20030801</u>					WBID Number <u>2280A</u>		Matrix: Surface ☒ Fresh ☒ Surface Salt ☐		Sample Type: E. coli ☒ Enterococcus ☐ Fecal ☐		Initials Sample & Preservation <u>BD</u>		Initials Documentation <u>KA</u>	
					Sample Access: Wading ☐ Dock ☐ From Shore ☒ Bridge ☐				Sample Ice Preserved in 15 minutes? Yes ☒ No ☐					
Time	Total depth(m)	Sample depth(m)	Temp °C	Conductivity µmhos	Salinity ppt	pH s.u.	D.O. mg/L	D.O. % sat	Secchi m	Tide Ebb/ Flood Slack				
<u>1025</u>	<u>0.57</u>	<u>0.25</u>	<u>13.7</u>	<u>571</u>	<u>0.28</u>	<u>7.58</u>	<u>8.42</u>	<u>81.3</u>	<u>VOB</u>					
Comments and Weather:														

WIN Number / Site Name <u>Little fishweir @ Oak St.</u> <u>20030953</u>					WBID Number <u>2280A</u>		Matrix: Surface ☒ Fresh ☒ Surface Salt ☐		Sample Type: E. coli ☒ Enterococcus ☐ Fecal ☐		Initials Sample & Preservation <u>BD</u>		Initials Documentation <u>KA</u>	
					Sample Access: Wading ☐ Dock ☐ From Shore ☒ Bridge ☐				Sample Ice Preserved in 15 minutes? Yes ☒ No ☐					
Time	Total depth(m)	Sample depth(m)	Temp °C	Conductivity µmhos	Salinity ppt	pH s.u.	D.O. mg/L	D.O. % sat	Secchi m	Tide Ebb/ Flood Slack				
<u>1140</u>	<u>0.65</u>	<u>0.3</u>	<u>14.4</u>	<u>451.2</u>	<u>0.22</u>	<u>7.14</u>	<u>4.72</u>	<u>46.9</u>	<u>VOB</u>	<u>Ebb</u>				
Comments and Weather:														

 Signature: B-D Date: 1/7/2021 BD 1/7/21

 Field Sheet Scanned (Initials/date): BD 1/7/21 Parameters uploaded to LIMS (Initials/date): _____

 Data QC'd in LIMS (Initials/date): BD 2/9/21 Data Migrated to WIN (Initials/date): _____

WIN Number / Site Name <i>Craig Cr @ Hendricks Ave</i>					WBID Number <i>2297B</i>		Matrix: Surface ☐ Fresh ☐ Surface Salt ☒		Sample Type: E. coli ☐ Enterococcus ☒ Fecal ☐		Initials Sample & Preservation <i>BD</i>		Initials Documentation <i>KA</i>	
Sample Access: Wading ☐ Dock ☐ From Shore ☒ Bridge ☐					Sample Ice Preserved in 15 minutes? Yes ☒ No ☐									
Time	Total depth(m)	Sample depth(m)	Temp °C	Conductivity µmhos	Salinity ppt	pH s.u.	D.O. mg/L	D.O. % sat	Secchi m	Tide Ebb/ Flood Slack				
<i>1250</i>	<i>0.5</i>	<i>0.3</i>	<i>15.4</i>	<i>5746</i>	<i>3.40</i>	<i>7.11</i>	<i>5.89</i>	<i>60.5</i>	<i>103</i>	<i>SLACK</i>				
Comments and Weather:														

WIN Number / Site Name <i>Goodbys Cr Above west Branch at San Clark</i>					WBID Number <i>2326A</i>		Matrix: Surface ☐ Fresh ☒ Surface Salt ☐		Sample Type: E. coli ☐ Enterococcus ☐ Fecal ☐		Initials Sample & Preservation <i>BD</i>		Initials Documentation <i>KA</i>	
Sample Access: Wading ☐ Dock ☐ From Shore ☒ Bridge ☐					Sample Ice Preserved in 15 minutes? Yes ☒ No ☐									
Time	Total depth(m)	Sample depth(m)	Temp °C	Conductivity µmhos	Salinity ppt	pH s.u.	D.O. mg/L	D.O. % sat	Secchi m	Tide Ebb/ Flood Slack				
<i>1335</i>	<i>0.53</i>	<i>0.3</i>	<i>14.0</i>	<i>400.0</i>	<i>0.19</i>	<i>7.53</i>	<i>7.93</i>	<i>77.1</i>	<i>103</i>	<i>EBB</i>				
Comments and Weather:														

WIN Number / Site Name <i>Craig Cr. 100m Down from Footbridge In Park</i>					WBID Number <i>2297B</i>		Matrix: Surface ☐ Fresh ☐ Surface Salt ☒		Sample Type: E. coli ☐ Enterococcus ☒ Fecal ☐		Initials Sample & Preservation <i>BD</i>		Initials Documentation <i>KA</i>	
Sample Access: Wading ☐ Dock ☐ From Shore ☒ Bridge ☐					Sample Ice Preserved in 15 minutes? Yes ☒ No ☐									
Time	Total depth(m)	Sample depth(m)	Temp °C	Conductivity µmhos	Salinity ppt	pH s.u.	D.O. mg/L	D.O. % sat	Secchi m	Tide Ebb/ Flood Slack				
<i>1300</i>	<i>0.7</i>	<i>0.3</i>	<i>15.6</i>	<i>14006</i>	<i>8.16</i>	<i>7.4</i>	<i>7.56</i>	<i>79.3</i>	<i>0.4</i>	<i>EBB</i>				
Comments and Weather:														

WIN Number / Site Name <i>Butcher Pen Cr @ Confederate Pt Rd.</i>					WBID Number <i>2322</i>		Matrix: Surface ☒ Fresh ☒ Surface Salt ☐		Sample Type: E. coli ☒ Enterococcus ☐ Fecal ☐		Initials Sample & Preservation <i>BD</i>		Initials Documentation <i>KA</i>	
Sample Access: Wading ☐ Dock ☐ From Shore ☒ Bridge ☐					Sample Ice Preserved in 15 minutes? Yes ☒ No ☐									
Time	Total depth(m)	Sample depth(m)	Temp °C	Conductivity µmhos	Salinity ppt	pH s.u.	D.O. mg/L	D.O. % sat	Secchi m	Tide Ebb/ Flood Slack				
<i>1055</i>	<i>0.7</i>	<i>0.3</i>	<i>12.9</i>	<i>486.3</i>	<i>0.24</i>	<i>7.32</i>	<i>7.31</i>	<i>69.5</i>	<i>0.7</i>	<i>JOB</i>				
Comments and Weather:														



CHAIN OF CUSTODY / LABORATORY ANALYSIS REQUEST FORM

9143 Philips Highway, Suite 200 Jacksonville, FL 32256 / Ph(904) 739-2277 / FAX (904) 739-2011

SR #

ALS Contact: Mandy Sullivan

Page 1 of 1

[illegible]