

Chemical Analysis Report

CEN-DIST-2021-04-15-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Sweetwater/Salt Creek/Long Branch**
Request ID: **RQ-2021-03-01-11**
Customer: **CEN-DIST**
Project ID: **LKJES-BMAP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Kalina Warren

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Colin Wright, Program Administrator

Date Certified: 05-MAY-2021 15:01



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited by the Florida Department of Health Environmental Laboratory Certification Program or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Salt Creek @ Packard Ave.

Collection Date/Time: 04/14/2021 11:40

Field ID: G2CE0088

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2239917	EPA 350.1 Rev. 2.0	Ammonia-N	0.16		mg N/L	P397094	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.97		mg N/L	P396978	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.15		mg N/L	P396843	
	EPA 365.1 Rev. 2.0	Total-P	0.25		mg P/L	P396701	
2239919	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.20		mg P/L	P396623	
2239921	SM 2540 D-2011	TSS	5	I	mg/L	P396815	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: Sweetwater Creek at Florida Avenue

Collection Date/Time: 04/14/2021 10:42

Field ID: G2CE0123

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2239918	EPA 350.1 Rev. 2.0	Ammonia-N	0.99		mg N/L	P397094	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.9		mg N/L	P396978	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.75		mg N/L	P396843	
	EPA 365.1 Rev. 2.0	Total-P	0.46		mg P/L	P396701	
2239920	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.36		mg P/L	P396623	
2239922	SM 2540 D-2011	TSS	4	IA	mg/L	P396714	

Ref. Method and Comment:

SM 2540 D-2011: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P397094

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P396978

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P396843

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P396701

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P396623

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2540 D-2011
Batch ID: P396714

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 2540 D-2011
Batch ID: P396815

Component	Result	Code	Units
TSS	2	U	mg/L

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P397094

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	97.6		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P396978

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	98.6		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P396843

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	97.0		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P396701

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	104		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P396623

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	96.7		P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P397094

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2239876	Ammonia-N	94.6	96.0	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P396978

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2239889	Kjeldahl Nitrogen	103	108	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P396843

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2239876	NO2NO3-N	98.0	99.0	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P396701

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2239840	Total-P	103	104	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P396623

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2239758	O-Phosphate-P	95.8	98.2	P/P	90 - 110

Quality Assurance Report Precision

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P397094

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2239876	Ammonia-N	1.31	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P396978

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2239889	Kjeldahl Nitrogen	3.12	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P396843

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2239876	NO2NO3-N	1.02	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P396701

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2239840	Total-P	0.338	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P396623

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2239758	O-Phosphate-P	2.38	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A104762

Included Lab Sample IDs: 2239919, 2239920

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	99.1	98.8	P/P	90 - 110
O-Phosphate-P	99.3	99.9	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A104810

Included Lab Sample IDs: 2239917, 2239918

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	102	99.9	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A104844

Included Lab Sample IDs: 2239917, 2239918

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	97.5	97.5	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A104950

Included Lab Sample IDs: 2239917, 2239918

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	99.5	99.9	P/P	90 - 110
Ammonia-N	99.9	99.3	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A104990

Included Lab Sample IDs: 2239917, 2239918

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	98.0	97.4	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		Precision LCS	Precision SMP	MS
EPA 350.1 Rev. 2.0	Ammonia-N	97.6	94.6	96.0			1.31
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	98.6	103	108			3.12
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	97.0	98.0	99.0			1.02
EPA 365.1 Rev. 2.0	Total-P	104	103	104			0.338
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	96.7	95.8	98.2			2.38

Reference Method Descriptions

Method	Description	Associated Samples
EPA 350.1 Rev. 2.0	Ammonia in aqueous matrices as mg N/L	2239917, 2239918
EPA 351.2 Rev. 2.0	Total Kjeldahl Nitrogen in aqueous matrices	2239917, 2239918
EPA 353.2 Rev. 2.0	Nitrite/Nitrate in aqueous matrices as mg N/L	2239917, 2239918
EPA 365.1 Rev. 2.0	Total Phosphorus in aqueous matrices as mg P/L	2239917, 2239918
EPA 365.1 Rev. 2.0 dissolved	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2239919, 2239920
SM 2540 D-2011	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2239921, 2239922

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 350.1 Rev. 2.0	04/15/2021			04/23/2021 18:01	Ping Hua	2239917
	04/15/2021			04/23/2021 18:24	Ping Hua	2239918
EPA 351.2 Rev. 2.0	04/15/2021	04/22/2021 14:21	Benjamin T. Nordmann	04/26/2021 15:31	Virginia P. Brown	2239917
	04/15/2021	04/22/2021 14:21	Benjamin T. Nordmann	04/26/2021 15:33	Virginia P. Brown	2239918
EPA 353.2 Rev. 2.0	04/15/2021			04/19/2021 15:30	Touraj Touran	2239917
	04/15/2021			04/19/2021 15:31	Touraj Touran	2239918
EPA 365.1 Rev. 2.0	04/15/2021	04/16/2021 13:47	Yen Ta	04/16/2021 18:24	Shengyu Ding	2239917
	04/15/2021	04/16/2021 13:47	Yen Ta	04/16/2021 18:25	Shengyu Ding	2239918
EPA 365.1 Rev. 2.0 dissolved	04/15/2021			04/15/2021 13:44	Lipi Saha	2239919
	04/15/2021			04/15/2021 14:25	Lipi Saha	2239920
SM 2540 D-2011	04/15/2021			04/15/2021 15:21	Yijie Li	2239922
	04/15/2021			04/16/2021 15:36	Yijie Li	2239921