Meter Calibra[illegible]/Verification Logsheet
(FDEP SOP FT 1000-FT 1600, FD 1000-FD 4000)

Meter ID: ~~71500~~ 738999
Project: CHICO BMAP

Method/Parameter	Technician	MMDDYY	24 HH:MM								Acceptance Critera
DEP SOP FT 1400 *Temperature*	Initials	Date	Time	Previous CV		Bath Type	NIST Thermometer S/N	NIST Thermometer Temp °C	Meter Temp °C		± 0.5 °C
CAL ICV CCV MAINT				7.16.20							Pass Fail
DEP SOP FT 1500 *Dissolved Oxygen*	Initials	Date	Time	Baro in Hg	Water Temp °C	Saturation mg/L	Meter Baro in Hg	Meter DO mg/L	Meter % DO	Meter Temp °C	± 0.3 mg/l
CAL ICV CCV MAINT	tm	04.13.21	08:53	29.96	19.2	9.25	29.99	9.32	101.22	19.37	Pass Fail
CAL ICV CCV MAINT	KP	04.13.21	18:30	30.01	19.8	~~9.27 KP~~ 9.15	30.01	9.28	101.45	19.91	Pass Fail
CAL ICV CCV MAINT											Pass Fail
CAL ICV CCV MAINT											Pass Fail
CAL ICV CCV MAINT											Pass Fail
DEP SOP FT 1200 *Specific Conductance*	Initials	Date	Time	Standard µmhos/cm	Standard Expiration	Standard ID	Meter Temp °C	Meter µmhos/cm			± 5%
CAL ICV CCV MAINT	tm	04.13.21	08:59	146.9	10.21	INV210431	20.97	139.93			Pass Fail
CAL ICV CCV MAINT	KP	04.13.21	18:34	10,000	11.21	INV200219	21.03	9,905.3			Pass Fail
CAL ICV CCV MAINT											Pass Fail
CAL ICV CCV MAINT											Pass Fail
CAL ICV CCV MAINT											Pass Fail
DEP SOP FT 1100 *pH*	Initials	Date	Time	Standard SU	Standard Expiration	Standard ID	Meter pH mV	Meter SU			± 0.2 SU
CAL ICV CCV MAINT	tm	04.13.21	09:06	4	07.22	INV200833	146.1	4.02			Pass Fail
CAL ICV CCV MAINT	KP	04.13.21	18:42	10	12.22	INV210368	-215.4	10.15			Pass Fail
CAL ICV CCV MAINT											Pass Fail
CAL ICV CCV MAINT											Pass Fail
CAL ICV CCV MAINT											Pass Fail
CAL ICV CCV MAINT											Pass Fail
CAL ICV CCV MAINT											Pass Fail
DEP SOP FT 1000 *ORP*	Initials	Date	Time	Standard mV	Standard Expiration	Standard ID	Meter Temp °C	Meter mV			± 20 mV
CAL ICV CCV MAINT											Pass Fail
CAL ICV CCV MAINT											Pass Fail
CAL ICV CCV MAINT											Pass Fail
CAL ICV CCV MAINT											Pass Fail
CAL ICV CCV MAINT											Pass Fail

Notes:

- NEW BATCH OF 146.9 COND. KCl STD