



Meter ID: AT5000
Project: DMAP CH40

Method/Parameter	Technician	MMDDYY	24 HH:MM	Acceptance Criteria			
DEP SOP FT 1400 Temperature	Initials	Date	Time	Previous CV	Bath Type	NIST Thermometer	
						S/N	Meter
						Temp °C	Temp °C
DEP SOP FT 1500 Dissolved Oxygen	Initials	Date	Time	Baro in Hg	Water Temp °C	Saturation mg/L	Meter
						Baro in Hg	DO mg/L
						% DO	Temp °C
							Temp °C
DEP SOP FT 1200 Specific Conductance	Initials	Date	Time	µmhos/cm	Expiration	ID	Meter
							Temp °C
							µmhos/cm
DEP SOP FT 1100 pH	Initials	Date	Time	SU	Expiration	ID	Meter
							pH mV
							SU
DEP SOP FT 1000 ORP	Initials	Date	Time	mV	Expiration	ID	Meter
							Temp °C
							mV

Notes:

• pH VALUE 8.9, 21 POST SAMPLE CV FAILED - pH CLIMBING OVER COURSE OF SEVERAL MINUTES - STABILIZED AT 10.46 - C.A. CALIBRATION