



Method/Parameter		Technician	MMDYY	24 HH:MM	Previous CV				Bath Type	NIST Thermometer		Meter		Acceptance Criteria	
DEP SOP FT 1400 Temperature		Initials	Date	Time						S/N		Temp °C	Temp °C	± 0.5 °C	
CAL	ICV	CCV	MAINT											Pass	Fail
DEP SOP FT 1500 Dissolved Oxygen															
		Initials	Date	Time	Baro in Hg	Water Temp °C	Saturation mg/L	Baro in Hg		DO mg/L	% DO	Temp °C			
CAL	ICV	CCV	MAINT	11/8/21	0750	30.19	19.8	9.90	30.29	9.37	99.65	18.92	Pass	Fail	
CAL	ICV	CCV	MAINT	11/8/21	15:54	30.14	19.0	9.35	30.19	9.28	99.68	19.24	Pass	Fail	
CAL	ICV	CCV	MAINT										Pass	Fail	
CAL	ICV	CCV	MAINT										Pass	Fail	
DEP SOP FT 1200 Specific Conductance															
		Initials	Date	Time	µmhos/cm	Expiration		ID	Temp °C		µmhos/cm				
CAL	ICV	CCV	MAINT	11/8/21	0755	1467	3/2022	INV210952	20.56	14.71		Pass	Fail		
CAL	ICV	CCV	MAINT	11/8/21	16:02	50,000	3/2023	INV210663	20.71	48.92		Pass	Fail		
CAL	ICV	CCV	MAINT										Pass	Fail	
CAL	ICV	CCV	MAINT										Pass	Fail	
DEP SOP FT 1100 pH															
		Initials	Date	Time	SU	Expiration		ID	pH mV		SU				
CAL	ICV	CCV	MAINT	11/8/21	0800	4	4/23/21	INV210362	45.5	9.04		Pass	Fail		
CAL	ICV	CCV	MAINT	11/8/21	16:10	10.0	3/23	INV210504	30.56	9.91		Pass	Fail		
CAL	ICV	CCV	MAINT										Pass	Fail	
CAL	ICV	CCV	MAINT										Pass	Fail	
CAL	ICV	CCV	MAINT										Pass	Fail	
CAL	ICV	CCV	MAINT										Pass	Fail	
DEP SOP FT 1000 ORP															
		Initials	Date	Time	mV	Expiration		ID	Temp °C		mV				
CAL	ICV	CCV	MAINT										Pass	Fail	
CAL	ICV	CCV	MAINT										Pass	Fail	
CAL	ICV	CCV	MAINT										Pass	Fail	
CAL	ICV	CCV	MAINT										Pass	Fail	
CAL	ICV	CCV	MAINT										Pass	Fail	

Notes: