

FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
CHAIN OF CUSTODY FORM

Date: 9/13/2021
Customer: WAS
RQ: RQ-2021-09-13-06

Collected By: Evelyn Becerra, Blake Spencer
Sampling Agency: FDEP
Project ID: SMP

PLEASE RETURN THE ORIGINAL OF THIS FORM WITH SAMPLE SUBMITTAL
FORMS AND THE COOLER PACKING WORKSHEET TO THE LAB


Number of Coolers Shipped: 2

Delivery Method: HandDelivered

FIELD ID (Labels) submitted under this RQ for this collection date.

Site Location	Field ID/WIN ID
Merritt's Mill Pond 500 meters upstream of dam	 G2WA0009
SHANGRI LA SPRING Vent	 G2WA0022
Merritt's Mill Pond 600 meters downstream of Jackson Blue Sp	 G2WA0007
Merritt's Mill Pond mid under powerlines	 G2WA0008

RELINQUISHED BY(SIGNATURE):



DATE/TIME: 9/13/2021 EST

THIS SECTION TO BE COMPLETED BY LABORATORY

Rec'd/Inspected By:

Rec'd/Inspected Date:



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PLAN FIELD SHEET

January 2020

Meter Passed End
of Day CCV
Yes / No

WIN Station Name: Merritt's Mill Pond 600 meters downstream of Jackson Blue Sp **Date:** 9/13/2021
WIN/Field ID: G2WA0007 **WBID:** 180A **ENR:** **Latitude:** 30.78886917
County: JACKSON **Org ID:** 21FLWQA **Longitude:** -85.14250969
Receiving Body of Water: - **RQ:** 2021-09-13-06

Sampling Team Members		WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
☒	Evelyn Becerra	x		x		
☒	Blake Spencer		x		x	x
☐						
☐						

Sampling Equipment	
☒ Sample Container	☐ Carboy
☐ Dipper	☐ Van Dorn
☒ Syringe	☐ Pole
Depth measured with: Sonar	
Secchi Equipment ID:	
Collection Method	
☐ Wading	☐ Canoe/Kayak
☐ From Shore	☐ Electric Motor
☐ Other	☒ Gasoline Motor

Water Level: ☒ Normal ☐ Low ☐ High ☐ Dry ☐ Pooled ☐ Flooded						Meter ID# EXO #1					
Photos: NO		Flow: NA		Tide:		Tide Times: High: Low:					
Biological Samples Collected: ☒ None ☐ HA ☐ SCI ☐ RPS ☐ Algae ID ☐ LVS ☐ LVI Other:											
SBIO ID Numbers:											
Notes:											

Duplicate Sample

Time Zone: EST	Time:	Field ID: N/A	(WIN ID_DUP)
WIN DMT Activity ID:			

Blank

Time Zone:	Time:	Field ID:	(Blank Type_DDDMMYYY)
DI Source:		Location:	
DI Type:		Equipment ID:	
DI Container (Name/ID):		DI Container Type:	

LIMS EVENT ID: WAS-SECT-2021-09-13-01 **WIN Import ID:** 21FLWQA_WAS-SECT-2021-09-13-01_10-13-2021.TXT
Enter Field Parameters, Verify Station ID In LIMS DMT: ESB 9/16/21 **QC Station ID, Field Parameters In LIMS DMT:** CC 10/06/2021
(Initials/Date) (Initials/Date)
Scan/Save Field Sheets ESB 9/16/21 **Migrate Data to WIN:** DT 10/13/21 **Verify Data In WIN:**
(Initials/Date) (Initials/Date) (Initials/Date)



RQ-2021-09-13-06

Customer: WQAP

Project ID: SMP

Collected By: Evelyn Becerra, Blake Spencer

Field Report Prepared By: Evelyn Becerra

Sampling Agency: FDEP

Place Label Here

WIN ID/Field ID:



G2WA0007

Location: Merritt's Mill Pond 600 meters downstream of Jackson Blue Sp

Comments:

Matrix: Fresh

(Check one)

☒ GRAB☐ COMPOSITE

Date	Time	DO (mg/L)	DO (%SAT)	Temp. (C°)	pH	Sample Depth (m)	Secchi Disk (m)	Total Depth (m)	Sp COND (µmhos/cm)	Salinity (PPT)		
	1359 EST	9.99	113.2	21.475	9.96 J	0.3	1.8 ☒ VOB (S)	1.8	278.2	0.13		
9/13/2021												
	1359 EST	10.08	115	21.478	10.9 J	1.3			277.5	0.13		
Parameter Suite		Parameter Methods				Preservation		pH Check	# Bottles sent to Lab	Bottle Group		
Preserved Nutrients (P-500ML)		☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				☒ Ice ☒ H2SO4*		☒ <2	1	A		
		Lot #	SA0293040		Exp:	10/19/21						
Metals (P-500ML)		☒ W-ICPMS ☒ W-ICP ☒ W-HARD				☒ Ice ☒ HNO3*		☒ <2	1	A		
		Lot #	NA0344060		Exp:	01/04/22						
Filtered Nutrients (P125ML)		☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter				☒ Ice			1	A		
		Filter Lot #	17347262									
		Syringe Lot #	1091364									
Chlorophyll (BP-1L)		☒ CHLSUITE-W				☒ Ice			1	A		
Physical/Aggregate (Fresh) (P-1L)		☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-Cl-IC / W-COLOR / W-SO4-IC / W-TDS				☒ Ice			1	A		
Physical/Aggregate (Marine) (P-1L)		☐ Alkalinity / W-BR-IC / W-F / W-TSS / TURBIDITY				☐ Ice						
Macroinvertebrates (PJ-2L)		☐ MI-FW-QLDC	Formalin Lot #			☐ Formalin						
Periphyton (PT-50ML)		☐ ALGAE-ID				☐ Ice						
Microbiology (Contract Lab Bottle)		☐ E Coli ☐ F Coli ☐ Enterococci ☐ Contract Lab				☐ Ice						
Molecular (QPCR-P-500ML)		☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-DG3				☐ Ice						
		☐ PCR-GULL2 ☐ PCR-GFD ☐				☐ Ice						
Tracers (BG-500ML)		☐ W-E8321-DI ☐ W-E8321-MS				☐ Ice						
Pesticides (BG-500ML) (BG-1L)		W-E8321-DI	W-E8321-MS	W-PSNP-TQ	W-CT-CL-R	☐ Ice						
Phytoplankton/Periphyton		W-PCL-TQ-R	W-CARB-AA (5 ml of MCAA** Buffer Solution)			☐ Ice						
				DTY-QN-C		☐ Ice						
				PKD-QN-BV								
Name: Evelyn Becerra Blake Spencer					Signature: <i>Blake Spencer</i>				Date: 9/13/21			

*2.0 ml of Preservative used unless noted otherwise.

**MCAA: Monochloroacetic acid buffer solution

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Meter ID:

Exo #1

RQ:

2021-09-13-06

Project:

MMP BMAP

boldly "X" this box if there are quality data on this page

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Blake Spencer	9/13/21	0739	20.7	761.5	9.0	9.01	100.2	—	1.1492	P/F	L/F
ICV	Blake Spencer	9/13/21	0740	20.9	761.4	8.9	9.09	101.7	—	—	P/F	L/F
CCV	Blake Spencer	9/13/21	1717	24.06	759.0	8.4	8.41	99.9%	—	—	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard µmhos/cm	Meter Reading µmhos/cm	Pass / Fail	Lab / Field
Calibr.	Blake Spencer	9/13/21	0746	210817J	09/22	1000	1000	P/F	L/F
ICV	Blake Spencer	9/13/21	0750	4103146	07/2023	100	1099.8	P/F	L/F
CCV	Blake Spencer	9/13/21	1651	210817J	09/22	1000	1001	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria ± 5%

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Blake Spencer	09/13/21	0755	210513A	11/22	7.02	21.5	7.02	-29.3	P/F	L/F
Calibr.	Blake Spencer	09/13/21	0758	210513B	11/22	4.00	21.6	4.00	146.4	P/F	L/F
Calibr.	Blake Spencer	9/13/21	0801	201209G	06/22	10.04	21.6	10.63	-190.2	P/F	L/F
ICV	Blake Spencer	9/13/21	0904	210513B	11/22	4.00	21.5	4.00	145.5	P/F	L/F
CCV	Blake Spencer	9/13/21	1523	201209G	06/22	10.04	21.5	10.63	145.5	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50; mV pH 4 Range +180 ± 50; mV pH 10 Range -180 ± 50;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: _____

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air						P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria ± 5 % or ± 0.05m, whichever is greater.

COMMENTS:

Cooler Packing Worksheet for Request: RQ-2021-09-13-06

Merrit's Mill Pond BMAP

Ship Cooler On: 9/9/2021

Customer/Project: WAS-SECT/BMAP

Assigned To: KLUG_K

Requester: Curtis Musson

Priority: 3

850-245-8453

2600 Blair Stone Rd

Tallahassee, FL 32399

Attn: Curtis Musson

Please return
packing list with
sample submittal
forms.

Comments:

Group A: Nutrients, Metals x3

Group B: Algal_ID x1

Group C: Nutrients, Metals, Pesticides x 1

Include carbamates preservation at 1 tube/container

Packing Notes:

No Fedex Labels

No Trash Bags

Sampling Note:

Group A: 3 sties

Group B/C: 1 site - Powerlines

Requested Analyses:

Bottle Group: A

of Sites: 3

Container ID: P-1L

Qty: 3 Preservation: ICE

Lot # 00680826

Description: Plastic Bottle 1L

Analysis

ALKALINITY

TURBIDITY

W-CL-IC

W-COLOR

W-F

W-SO4-IC

W-TDS

W-TSS

Description

Alkalinity in aqueous matrices as mg CaCO3/L

Turbidity in aqueous matrices

Chloride in aqueous matrices

True Color measured at 450 nm (not valid for NPDES monitoring)

Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)

Sulfate in aqueous matrices

Total Dissolved Solids in aqueous matrices

Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample

Container ID: BP-1L

Qty: 3 Preservation: ICE

Lot # 1326477

Description: Brown Plastic Bottle 1L

Analysis

CHLSUITE-W

Description

Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry

Container ID: P-500ML

Qty: 3 Preservation: HNO3

Lot # 00080369

Description: Plastic Bottle 500 mL

Affix RED nitric acid dot to container.

Analysis

W-HARD

W-ICP

W-ICPMS

Description

Hardness by calculation, in aqueous matrices, reported as mg/L CaCO3

Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects

Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects

Container ID: P-500ML

Qty: 3 Preservation: ICE And H2SO4

Lot # 00080369

Description: Plastic Bottle 500 mL

Affix YELLOW sulfuric acid dot to container.

Analysis

W-NH3

W-NO2NO3

W-S-A-TP

W-TKN

W-TOC

Description

Ammonia in aqueous matrices as mg N/L

Nitrite/Nitrate in aqueous matrices as mg N/L

Total Phosphorus in aqueous matrices as mg P/L

Total Kjeldahl Nitrogen in aqueous matrices

Nonpurgeable Organic Carbon in aqueous matrices

Container ID: P-125ML

Qty: 3 Preservation: FILTER-ICE

Lot # 00072954

Description: Plastic Bottle 125 mL

Analysis

W-PO4-F

Description

Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L

Bottle Group: B

of Sites: 1

Container ID: PT-50ML

Qty: 1 Preservation: ICE

Lot # 11621053

Description: Plastic Tube 50 mL

Analysis

ALGAL_ID

Description

Qualitative assessment of algal taxa present without counts

Bottle Group: C

of Sites: 1

Container ID: P-1L

Qty: 1 Preservation: ICE

Lot # 00080826

Description: Plastic Bottle 1L

Analysis

ALKALINITY

TURBIDITY

W-CL-IC

Description

Alkalinity in aqueous matrices as mg CaCO3/L

Turbidity in aqueous matrices

Chloride in aqueous matrices

W-COLOR	True Color measured at 450 nm (not valid for NPDES monitoring)
W-F	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
W-SO4-IC	Sulfate in aqueous matrices
W-TDS	Total Dissolved Solids in aqueous matrices
W-TSS	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample

Container ID: BP-1L

Qty: 1 Preservation: ICE

Lot # 1326677

Description: Brown Plastic Bottle 1L

Analysis

CHLSUITE-W

Description

Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry

Container ID: BG-500ML

Qty: 1 Preservation: CH2CICOOH and Ice

Lot # 00080737

Description: Brown Glass Bottle 500ML *FILL Completely*

Affix BLUE dot to container.

Analysis

W-CARB-AA

Description

Carbamates in water matrices by HPLC/MS/MS

Container ID: BG-500ML FILL TO TOP Qty: 4 Preservation: ICE

Lot # 00080737

Description: Brown Glass Bottle 500ML *Fill Completely*

Affix BLUE dot to container.

Analysis

W-CT-CI-R

W-PCL-TQ-R

Description

Organochlorine pesticides in water matrices by GC/MS with Negative Chemical Ionization

Organochlorine pesticides, pyrethroids and tracers in water matrices by GC/MS/MS

Container ID: BG-500ML

Qty: 1 Preservation: ICE

Lot # 00080737

Description: Brown Glass Bottle 500ML *FILL Completely*

Affix BLUE dot to container.

Analysis

W-E8321-DI

W-E8321-MS

Description

Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS

Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS

Container ID: P-500ML

Qty: 1 Preservation: HNO3

Lot # 00080369

Description: Plastic Bottle 500 mL

Affix RED nitric acid dot to container.

Analysis

W-HARD

W-ICP

W-ICPMS

Description

Hardness by calculation, in aqueous matrices, reported as mg/L CaCO3

Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects

Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects

Container ID: P-500ML

Qty: 1 Preservation: ICE And H2SO4

Lot # 00080369

Description: Plastic Bottle 500 mL

Affix YELLOW sulfuric acid dot to container.

Analysis

W-NH3

W-NO2NO3

W-S-A-TP

W-TKN

W-TOC

Description

Ammonia in aqueous matrices as mg N/L

Nitrite/Nitrate in aqueous matrices as mg N/L

Total Phosphorus in aqueous matrices as mg P/L

Total Kjeldahl Nitrogen in aqueous matrices

Nonpurgeable Organic Carbon in aqueous matrices

Container ID: P-125ML

Qty: 1 Preservation: FILTER-ICE

Lot # 00072954

Description: Plastic Bottle 125 mL

Analysis

W-PO4-F

Description

Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L

Container ID: BG-1L

Qty: 4 Preservation: ICE

Lot # 00080147

Description: Brown Glass Bottle 1L

Affix BLUE dot to container.

Analysis

W-PSNP-TQ

Description

Organonitrogen and phosphorus pesticides in water matrices by GC/MS/MS

Cooler Packed By: 14K

Number of Coolers: 2

Date: 9.3.21

Kit must also include:

- ☒ Field Sheets
- ☒ Temperature Control Bottle
- ☐ FedEx Bills, if applicable
- ☐ Plastic Bags/Zip Ties
- ☐ Acid Labels (sets of 24)

If Preservation Included:

ID <u>MCA Buffer soln</u>	Exp <u>2.9.22</u>	Lot # <u>118437</u>
ID _____	Exp _____	Lot # _____
ID _____	Exp _____	Lot # _____
ID _____	Exp _____	Lot # _____
ID _____	Exp _____	Lot # _____

PLEASE RETURN ALL COOLERS FOR RQ-2021-09-13-06