

From: [Jeremy Bosso](#)
To: [DEP-Overflowmicro](#); [Butler, Myranda K](#); [King, Michael](#); [Dana Morton](#)
Subject: microbiology report for RQ-2022-01-17-46 (Bayou Chico BMAP) samples collected 01/18/2022
Date: Wednesday, January 19, 2022 5:43:23 PM
Attachments: [FF 2022 01 18A.txt](#)
[STF FDEP 2022 01 18A.pdf](#)
[DEP Reports 2022 Bayou Chico BMAP.xlsx](#)

EXTERNAL MESSAGE

This email originated outside of DEP. Please use caution when opening attachments, clicking links, or responding to this email.


Good afternoon. I've attached the microbiology report for RQ-2022-01-17-46 (Bayou Chico BMAP) samples collected 01/18/2022.

Samples analyzed for Enterococcus by **Enterolert (QT)** were incubated from 01/18/2022 (15:05) to 01/19/2022 (15:43).

Samples analyzed for *E. coli* by **SM 9223B (Colilert-18 QT)** were incubated from 01/18/2022 (15:15) to 01/19/2022 (11:30).

Note: The "J2" qualifier on sample 22-48933 (33020JD4 SW Arm Bayou Chico N of Mandalay Dr)) is due to our lab duplicate for that sample being outside of our internal precision acceptance criteria. I believe this is most likely due to random error and because the duplicate results were both such low concentrations (1.0 versus 3.1 MPN/100 mL before dilution factor). We always report the higher number of duplicate analyses.

Thank you,
Jeremy Bosso, MPH
Technical Director, [Wetlands Research Laboratory](#)
An Analytical Arm of the [Center for Environmental Diagnostics & Bioremediation](#)
University of West Florida
11000 University Pkwy, Pensacola, FL 32514
(850) 857-6096



Confidentiality Notice: This e-mail communication and any attachments may contain confidential and privileged information for the use of the designated recipient(s) named above. If you are not the intended recipient, you are hereby notified that you have received this communication in error and that any review, disclosure, dissemination, distribution or copying of it or its contents is prohibited.