

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Boldly "X" this box if there is qualified data on this page.

Project/Site:

Reference Laker / Moore Lake

Date:

8/24/2020

Meter #

154185

Temperature (Quarterly)		For Date of Last Temperature Verification see _____ in log book									
Dissolved Oxygen	DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
CAL	ICV	CCV	NKW	8/24/20	0835		8.85	21.5	100.2	8.85	P F
CAL	ICV	CCV	NKW	8/24/20	1254		8.68	22.3	100.0	8.693	P F
CAL	ICV	CCV									P F
CAL	ICV	CCV									P F
CAL	ICV	CCV									P F
CAL	ICV	CCV									P F
CAL	ICV	CCV									P F
CAL	ICV	CCV									P F
CAL	ICV	CCV									P F

Specific Conductance		DEP SOP FT 1200									
Initials	Date	Time	Standard µmhos/cm	Exp. Date	Lot #	Bottle #	Cell Constant	Reading µmhos/cm	Pass or Fail		
CAL	ICV	CCV	NKW	8/24/20	0830	100	08/2021	2908C81	106	P F	
CAL	ICV	CCV	NKW	8/24/20	0832	100	08/2021	2908C81	100	P F	
CAL	ICV	CCV	NKW	8/24/20	0833	100	08/2021	2908C81	100	P F	
CAL	ICV	CCV	NKW	8/24/20	1250	100	08/2021	2908C81	95	P F	
CAL	ICV	CCV								P F	
CAL	ICV	CCV								P F	
CAL	ICV	CCV								P F	
CAL	ICV	CCV								P F	

pH		DEP SOP FT 1100									
Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	Slope	Reading SU	Pass or Fail		
CAL	ICV	CCV	NKW	8/24/20	0823	7.0	10/2021	200407B	7.13	P F	
CAL	ICV	CCV	NKW	8/24/20	0825	7.0	10/2021	200407B		P F	
CAL	ICV	CCV	NKW	8/24/20	0826					P F	
CAL	ICV	CCV	NKW	8/24/20	0827	4.0	06/2021	191120A	4.12	P F	
CAL	ICV	CCV	NKW	8/24/20	1252	7.0	10/2021	200407B	7.12	P F	
CAL	ICV	CCV								P F	
CAL	ICV	CCV								P F	
CAL	ICV	CCV								P F	
CAL	ICV	CCV								P F	

Maintenance: Weekly pH Slope: _____

Specific Conductance Probe Cleaned? Yes No

Dissolved Oxygen Membrane Changed: Yes No

Notes:

Perform only in Calibrate Mode:
Perform only in Run Mode:
Perform only in Run Mode:

CAL - Calibrate -
ICV - Initial Calibration Verification
CCV - Continuing Calibration Verification