

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Boldly "X" this box if there is qualified data on this page.

Project/Site: Ref Lakes Porter Pond + ANF

Date: 9/25/2020

Meter # 154185

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

Dissolved Oxygen	DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
CAL ICV <u>CCV</u>		JJ	9/25/2020	3:23p			8.53	22.6	98.7	8.59	(P) F
CAL ICV <u>CCV</u>		AO	9/28/2020	4:33p			7.86	27.0	98.7	7.91	(P) F
CAL ICV CCV											P F
CAL ICV CCV											P F
CAL ICV CCV											P F
CAL ICV CCV											P F
CAL ICV CCV											P F

Specific Conductance	DEP SOP FT 1200	Initials	Date	Time	Standard μ hos/cm	Exp. Date	Lot #	Bottle #	Cell Constant	Reading μ hos/cm	Pass or Fail
CAL ICV <u>CCV</u>		JJ	9/25/2020	3:25p	1,000	12/2020	191120D	N/A		987	(P) F
CAL ICV <u>CCV</u>		AO	9/28/2020	4:36p	100	11/2021	2911C93	N/A		102.2	(P) F
CAL ICV CCV											P F
CAL ICV CCV											P F
CAL ICV CCV											P F
CAL ICV CCV											P F

pH	DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	Slope	Reading SU	Pass or Fail
CAL ICV <u>CCV</u>		JJ	9/25/2020	3:31p	7.0	10/2021	200407B	N/A		6.98	(P) F
CAL ICV <u>CCV</u>		JJ	9/25/2020	3:34p	4.01	10/2021	191120A	N/A		3.98	(P) F
CAL ICV <u>CCV</u>		AO	9/28/2020	4:43p	7.0	10/2021	200407B	N/A		6.99	(P) F
CAL ICV CCV											P F
CAL ICV CCV											P F
CAL ICV CCV											P F
CAL ICV CCV											P F
CAL ICV CCV											P F

Maintenance: Weekly pH Slope: _____

Specific Conductance Probe Cleaned? Yes No

Dissolved Oxygen Membrane Changed: Yes No

Notes:

transcribed by AO 10/2/2020

Perform only in Calibrate Mode:

Perform only in Run Mode:

Perform only in Run Mode:

CAL - Calibrate -

ICV - Initial Calibration Verification

CCV - Continuing Calibration Verification