

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Boldly "Y" this box
if there is qualified
data on this page.

Project/Site: Reference Streams

Date: 8/21/17 P2-2017-06-05-47

Meter # DEP 113529

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

Dissolved Oxygen	DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
------------------	--------------------	----------	------	------	-----------------	---------------	------	---------	------	------------------------------------	-----------------

Acceptance Criteria: $\pm 0.3 \text{ mg/L}$

CAL	ICV	CCV	<u>W</u>	<u>8/21</u>	<u>0715</u>		<u>8.0</u>	<u>21.3</u>	<u>101</u>	<u>7.926</u>	P
CAL	ICV	CCV	<u>W</u>	<u>8/22</u>	<u>0700</u>		<u>8.3</u>	<u>25.0</u>	<u>101</u>	<u>8.243</u>	P
CAL	ICV	CCV									F
CAL	ICV	CCV									F
CAL	ICV	CCV									F
CAL	ICV	CCV									F
CAL	ICV	CCV									F
CAL	ICV	CCV									F

Specific Conductance	DEP SOP FT 1200	Initials	Date	Time	Standard $\mu\text{mhos/cm}$	Exp. Date	Lot #	Bottle #	Cell Constant	Reading $\mu\text{mhos/cm}$	Pass or Fail
----------------------	--------------------	----------	------	------	---------------------------------	-----------	-------	----------	------------------	--------------------------------	-----------------

Acceptance Criteria: $\pm 5\%$

CAL	ICV	CCV	<u>W</u>	<u>8/21</u>	<u>0710</u>	<u>12900</u>	<u>66118</u>	<u>KEC 17-2</u>		<u>12900</u>	P
CAL	ICV	CCV	<u>W</u>	<u>8/21</u>	<u>0727</u>	<u>1413</u>	<u>104117</u>	<u>KEC 17-1</u>		<u>1414</u>	P
CAL	ICV	CCV	<u>W</u>	<u>8/22</u>	<u>0705</u>	<u>100</u>	<u>02019</u>	<u>2703994</u>		<u>103</u>	P
CAL	ICV	CCV									F
CAL	ICV	CCV									F
CAL	ICV	CCV									F
CAL	ICV	CCV									F
CAL	ICV	CCV									F

pH	DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
----	--------------------	----------	------	------	----------------	-----------	-------	----------	------------	---------------	-----------------

Acceptance Criteria: $\pm 0.2 \text{ SU}$

CAL	ICV	CCV	<u>W</u>	<u>8/21</u>	<u>0725</u>	<u>7</u>	<u>200804</u>		<u>-47.0</u>	<u>7.0</u>	P
CAL	ICV	CCV			<u>0714</u>	<u>4</u>	<u>2011083</u>		<u>130.4</u>	<u>4.0</u>	P
CAL	ICV	CCV			<u>0727</u>	<u>10</u>	<u>1005846</u>		<u>-208.1</u>	<u>10.0</u>	P
CAL	ICV	CCV	<u>W</u>	<u>8/22</u>	<u>0730</u>	<u>4</u>	<u>2011083</u>		<u>4.01</u>	<u>4.01</u>	P
CAL	ICV	CCV			<u>0710</u>	<u>10</u>	<u>1005846</u>			<u>10.01</u>	P
CAL	ICV	CCV									F
CAL	ICV	CCV									F
CAL	ICV	CCV									F
CAL	ICV	CCV									F

Maintenance: Weekly pH Slope: Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No

Notes: Team 2 Mery, N. Muley. Partner ck 33030002. Fine log ck-23040015