

Boldly "X" this box if there is qualified data on this page.

CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 02-05-2018

Project/Site: Ref. Stream/ Aucilla + Bethel Date: 4/29/2019 Meter # 154185
Rocky + Steinhatchee

Maintenance performed? (circle one) YES NO Date of maintenance: _____ Maintenance Log Page: # _____

Temperature (Quarterly)		For Date of Last Temperature Verification see _____ in log book										
Dissolved Oxygen	DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	Saturation mg/L (from chart)	Temp °C	% DO	Reading mg/L	Pass or Fail	Qualifier
Acceptance Criteria: +/- 0.3mg/l												
CAL ICV CCV	SEN	4/29/19	1345				8.514	23.4	100.0		P F	
CAL ICV CCV	SEN	4/29/19	1347				8.514	23.4	100.0	8.55	P F	
CAL ICV CCV	AO	4/11/19	0845				8.418	24.0	100.1	8.44	P F	
CAL ICV CCV	NKw	5/3/19	1138				8.985	20.6	99.3	8.92	P F	
CAL ICV CCV											P F	
CAL ICV CCV											P F	
CAL ICV CCV											P F	
Acceptance Criteria: +/- 5%												
CAL ICV CCV	SEN	4/29/19	1322				1000	7/2019	180621C	1	P F	
CAL ICV CCV	SEN	4/29/19	1325				100	7/2020	2808884	1	P F	
CAL ICV CCV	AO	5/1/19	0850				1000	7/2019	180621C	1	P F	
CAL ICV CCV	NKw	5/3/19	1133				100	7/2020	2808884	1	P F	
CAL ICV CCV											P F	
CAL ICV CCV											P F	
CAL ICV CCV											P F	
Acceptance Criteria: +/- 0.2 SU												
CAL ICV CCV	SEN	4/29/19	1330				7.0	10/2019	180403C	1	P F	
CAL ICV CCV	SEN	4/29/19	1333				4.0	7/2019	180621D	1	P F	
CAL ICV CCV	SEN	4/29/19	1335				4.0	7/2019	180621D	1	P F	
CAL ICV CCV	AO	5/1/19	0853				10.0	5/1/2020	180621F	1	P F	
CAL ICV CCV											P F	
CAL ICV CCV											P F	
CAL ICV CCV											P F	
CAL ICV CCV											P F	
CAL ICV CCV											P F	
CAL ICV CCV											P F	

Weekly pH Slope: _____ Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No

Notes:

Barometric pressure 761.7 AO
 solubility @ 24.0°C = 8.4 so no adjustment needed
 pH probe replacement needed; not fixed after maintenance by Sarah AO
 pH data not recorded for Rocky + Steinhatchee. New probe ordered. NKw

Perform only in Calibrate Mode: CAL - Calibrate -
 Perform only in Run Mode: ICV - Initial Calibration Verification
 Perform only in Run Mode: CCV - Continuing Calibration Verification