

Boldly "X" this box if there is qualified data on this page.

CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 02-05-2018

Project/Site: Little Crooked Creek

Date: 11/30/21

Meter # 154185

Maintenance performed? (circle one) YES ☒ NO ☐ Date of maintenance: _____ Maintenance Log Page: # _____

Temperature (Quarterly)		For Date of Last Temperature Verification see _____ in log book										
Dissolved Oxygen	DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	Saturation mg/L (from chart)	Temp °C	% DO	Reading mg/L	Pass or Fail	Qualifier
Acceptance Criteria: +/- 0.3mg/l												
CAL ICV CCV		JP	11/30/21	7:58			8.932	20.9	100.3	8.95	P F	
CAL ICV CCV		JP	11/30/21	8:00			8.95	20.8	100.7	8.96	P F	
CAL ICV CCV		CH	11/30/21	3:56			8.86	21.3	94.8	8.86	P F	
CAL ICV CCV											P F	
CAL ICV CCV											P F	
CAL ICV CCV											P F	
CAL ICV CCV											P F	
CAL ICV CCV											P F	
Acceptance Criteria: +/- 5%												
Specific Conductance	DEP SOP FT 1200	Initials	Date	Time	Standard umhos/cm	Exp. Date	Lot #	Bottle #	Cell Constant	Reading umhos/cm	Pass or Fail	Qualifier
CAL ICV CCV		JP	11/30/21	7:41	1000	08/2022	210817J			1000	P F	
CAL ICV CCV		JP	11/30/21	7:45	1000	03/2023	4103146			99.7	P F	
CAL ICV CCV		CH	11/30/21	3:54	1000	03/2023	4103146			103.6	P F	
CAL ICV CCV											P F	
CAL ICV CCV											P F	
CAL ICV CCV											P F	
CAL ICV CCV											P F	
Acceptance Criteria: +/- 0.2 SU												
pH	DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	Slope	Reading SU	Pass or Fail	Qualifier
CAL ICV CCV		JP	11/30/21	7:48	7.0	02/2023	210817D			7.02	P F	
CAL ICV CCV		JP	11/30/21	7:53	4.0	02/2023	210817B			4.0	P F	
CAL ICV CCV		JP	11/30/21	7:55	4.0	04/2023	210817B			4.0	P F	
CAL ICV CCV		CH	11/30/21	3:51	5.0	02/2023	210817D			7.01	P F	
CAL ICV CCV											P F	
CAL ICV CCV											P F	
CAL ICV CCV											P F	
CAL ICV CCV											P F	
CAL ICV CCV											P F	
CAL ICV CCV											P F	

Weekly pH Slope: _____ Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No

Notes:

Perform only in Calibrate Mode: CAL - Calibrate -
 Perform only in Run Mode: ICV - Initial Calibration Verification
 Perform only in Run Mode: CCV - Continuing Calibration Verification