

Chemical Analysis Report

SO-DIST-2018-10-02-02

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Caloosa Creeks**
Request ID: **RQ-2018-10-01-69**
Customer: **SO-DIST**
Project ID: **BMAP**

Send Reports to:
FL Dept. of Environmental Protection
FLDEP South District Office
2295 Victoria Ave. Suite 364
Fort Myers, FL 33901-3381
Attn: Gary Snorek

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Liang Lin, Environmental Administrator

Date Certified: 24-OCT-2018 14:30

A handwritten signature in black ink, appearing to read "Liang Lin", is shown on a light-colored background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Hancock Creek

Collection Date/Time: 10/01/2018 08:30

Field ID: G3SD0130

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2030230	SM 2120 B	Color (true)	93		PCU	P352637	
	SM 2320 B-97	Total Alkalinity	180		mg CaCO3/L	P352884	
	SM 2540 D-97	TSS	5	I	mg/L	P353126	
	SM 4500 F-C-97	Fluoride	0.24		mg F/L	P352887	
2030233	SM 5310 B-00 dissolved	Organic Carbon	14		mg C/L	P352973	
2030236	EPA 350.1 Rev. 2.0	Ammonia-N	0.036		mg N/L	P353033	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.0		mg N/L	P352822	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.088		mg N/L	P352776	
	EPA 365.1 Rev. 2.0	Total-P	0.18		mg P/L	P352745	
	SM 5310 B-00	Organic Carbon	14		mg C/L	P352973	
2030239	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.13		mg P/L	P352617	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: Orange River

Collection Date/Time: 10/01/2018 09:40

Field ID: G3SD0128

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2030231	SM 2120 B	Color (true)	57		PCU	P352637	
	SM 2320 B-97	Total Alkalinity	216	A	mg CaCO3/L	P352884	
	SM 2540 D-97	TSS	4	I	mg/L	P353127	
	SM 4500 F-C-97	Fluoride	0.15		mg F/L	P352887	
2030234	SM 5310 B-00 dissolved	Organic Carbon	10		mg C/L	P352973	
2030237	EPA 350.1 Rev. 2.0	Ammonia-N	0.020		mg N/L	P353034	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.69		mg N/L	P352822	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.14		mg N/L	P352776	
	EPA 365.1 Rev. 2.0	Total-P	0.044		mg P/L	P352745	
	SM 5310 B-00	Organic Carbon	11		mg C/L	P352973	
2030240	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.016		mg P/L	P352617	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: Telegraph Creek

Collection Date/Time: 10/01/2018 10:35

Field ID: G3SD0129

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2030232	SM 2120 B	Color (true)	220		PCU	P352674	
	SM 2320 B-97	Total Alkalinity	52		mg CaCO3/L	P352884	
	SM 2540 D-97	TSS	4	IA	mg/L	P353127	
	SM 4500 F-C-97	Fluoride	0.075	I	mg F/L	P352887	
2030235	SM 5310 B-00 dissolved	Organic Carbon	20		mg C/L	P352973	
2030238	EPA 350.1 Rev. 2.0	Ammonia-N	0.038		mg N/L	P353034	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.0		mg N/L	P353082	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.056		mg N/L	P352776	
	EPA 365.1 Rev. 2.0	Total-P	0.095		mg P/L	P352745	
	SM 5310 B-00	Organic Carbon	20		mg C/L	P352973	
2030241	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.064		mg P/L	P352618	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P353033

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P353034

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P352822

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P353082

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P352776

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P352745

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P352617

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P352618

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P352637

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Quality Assurance Report Method Blank Results

Reference Method: SM 2120 B
Batch ID: P352674

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P352884

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 D-97
Batch ID: P353126

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P353127

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P352887

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P352973

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Reference Method: SM 5310 B-00 dissolved
Batch ID: P352973

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P353033

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	99.4		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P353034

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	97.2		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P352822

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	103		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P353082

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	109		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P352776

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	101		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P352745

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	101		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P352617

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	104		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P352618

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	100		P	90 - 110

Reference Method: SM 2120 B
Batch ID: P352637

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	102		P	85 - 115

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 2120 B
Batch ID: P352674

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	101		P	85 - 115

Reference Method: SM 2320 B-97
Batch ID: P352884

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	103		P	98 - 107

Reference Method: SM 4500 F-C-97
Batch ID: P352887

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	99.2		P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P352973

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	99.0		P	85 - 115

Reference Method: SM 5310 B-00 dissolved
Batch ID: P352973

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	99.0		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P353033

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2030229	Ammonia-N	102	101	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P353034

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2030237	Ammonia-N	98.5	99.8	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P352822

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2030237	Kjeldahl Nitrogen	103	96.8	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P353082

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2030238	Kjeldahl Nitrogen	107	106	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P352776

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2030114	NO2NO3-N	102	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P352745

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2030114	Total-P	96.5	101	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P352617

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2030240	O-Phosphate-P	97.5	98.7	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P352618

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2030293	O-Phosphate-P	98.6	100	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P352887

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2030231	Fluoride	99.0	99.0	P/P	91 - 105

Quality Assurance Report Matrix Spike Accuracy

Reference Method: SM 5310 B-00
Batch ID: P352973

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2030114	Organic Carbon	90.8	91.8	P/P	85 - 115

Reference Method: SM 5310 B-00 dissolved
Batch ID: P352973

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2030114	Organic Carbon	90.8	91.8	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P353033

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2030229	Ammonia-N	0.778	Spike	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P353034

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2030237	Ammonia-N	0.959	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P352822

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2030237	Kjeldahl Nitrogen	3.73	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P353082

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2030238	Kjeldahl Nitrogen	0.221	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P352776

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2030114	NO2NO3-N	0.980	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P352745

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2030114	Total-P	4.96	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P352617

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2030240	O-Phosphate-P	1.05	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P352618

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2030293	O-Phosphate-P	1.37	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2120 B
 Batch ID: P352637

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2030179	Color (true)	2.23	Sample	P	0 - 20

Reference Method: SM 2120 B
 Batch ID: P352674

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2030158	Color (true)	2.71	Sample	P	0 - 20

Reference Method: SM 2320 B-97
 Batch ID: P352884

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2030231	Total Alkalinity	0.642	Sample	P	0 - 1.6

Reference Method: SM 4500 F-C-97
 Batch ID: P352887

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2030231	Fluoride	0.0	Spike	P	0 - 3.5

Reference Method: SM 5310 B-00
 Batch ID: P352973

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2030114	Organic Carbon	0.630	Spike	P	0 - 20

Reference Method: SM 5310 B-00 dissolved
 Batch ID: P352973

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2030114	Organic Carbon	0.630	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
 Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A86850

Included Lab Sample IDs: 2030239, 2030240, 2030241

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	99.0	99.7	P/P	90 - 110
O-Phosphate-P	99.9	101	P/P	90 - 110

Reference Method: SM 2120 B

Run ID: A86859

Included Lab Sample IDs: 2030230, 2030231

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	99.7	101	P/P	90 - 115

Reference Method: SM 2120 B

Run ID: A86870

Included Lab Sample IDs: 2030232

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	100	99.9	P/P	90 - 115

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A86915

Included Lab Sample IDs: 2030236, 2030237, 2030238

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	100	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A86932

Included Lab Sample IDs: 2030236, 2030238

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	101	103	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A86959

Included Lab Sample IDs: 2030230, 2030231, 2030232

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	100	103	P/P	90 - 110
Total Alkalinity	103	100	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A86959

Included Lab Sample IDs: 2030230, 2030231, 2030232

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	101	99.5	P/P	90 - 110
Fluoride	99.5	101	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A86974

Included Lab Sample IDs: 2030237

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	103	103	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A86981

Included Lab Sample IDs: 2030236, 2030237

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	94.8	95.6	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A86990

Included Lab Sample IDs: 2030236, 2030237, 2030238

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	93.2	91.8	P/P	90 - 110

Reference Method: SM 5310 B-00 dissolved

Run ID: A86990

Included Lab Sample IDs: 2030233, 2030234, 2030235

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	93.2	91.8	P/P	90 - 110
Organic Carbon	94.5	93.2	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A87014

Included Lab Sample IDs: 2030236, 2030237, 2030238

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	102	101	P/P	90 - 110
Ammonia-N	102	103	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A87086

Included Lab Sample IDs: 2030238

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	99.7	100	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 350.1 Rev. 2.0	Ammonia-N	99.4	102	101			0.778
	Ammonia-N	97.2	98.5	99.8			0.959
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	103	103	96.8			3.73
	Kjeldahl Nitrogen	109	107	106			0.221
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	101	102	102			0.980
EPA 365.1 Rev. 2.0	Total-P	101	96.5	101			4.96
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	104	97.5	98.7			1.05
	O-Phosphate-P	100	98.6	100			1.37
SM 2120 B	Color (true)	102				2.23	
	Color (true)	101				2.71	
SM 2320 B-97	Total Alkalinity	103				0.642	
SM 4500 F-C-97	Fluoride	99.2	99.0	99.0			0.0
SM 5310 B-00	Organic Carbon	99.0	90.8	91.8			0.630
SM 5310 B-00 dissolved	Organic Carbon	99.0	90.8	91.8			0.630

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	2030236, 2030237, 2030238
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	2030236, 2030237, 2030238
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	2030236, 2030237, 2030238
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	2030236, 2030237, 2030238
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2030239, 2030240, 2030241
SM 2120 B / E31780	True Color measured at 450 nm	2030230, 2030231, 2030232
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2030230, 2030231, 2030232
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2030230, 2030231, 2030232
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2030230, 2030231, 2030232
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	2030236, 2030237, 2030238
SM 5310 B-00 dissolved / E31780	Dissolved Nonpurgeable Organic Carbon in filtered, aqueous matrices	2030233, 2030234, 2030235

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 350.1 Rev. 2.0	10/02/2018			10/04/2018 13:26	Ping Hua	2030236
	10/02/2018			10/04/2018 13:35	Ping Hua	2030237
	10/02/2018			10/04/2018 14:02	Ping Hua	2030238
EPA 351.2 Rev. 2.0	10/02/2018	10/05/2018 10:15	Adrian Shelby	10/08/2018 13:25	Joseph Suarez	2030236
	10/02/2018	10/05/2018 10:15	Adrian Shelby	10/08/2018 13:26	Joseph Suarez	2030237
	10/02/2018	10/16/2018 11:00	Adrian Shelby	10/17/2018 11:03	Joseph Suarez	2030238
EPA 353.2 Rev. 2.0	10/02/2018			10/04/2018 13:58	Lauren Bottorf	2030236
	10/02/2018			10/04/2018 13:59	Lauren Bottorf	2030237
	10/02/2018			10/04/2018 14:00	Lauren Bottorf	2030238
EPA 365.1 Rev. 2.0	10/02/2018	10/04/2018 12:35	Sima Feizi	10/05/2018 08:57	Beverly Y. Sanders	2030236
	10/02/2018	10/04/2018 12:35	Sima Feizi	10/05/2018 08:59	Beverly Y. Sanders	2030238
	10/02/2018	10/04/2018 12:35	Sima Feizi	10/05/2018 14:03	Beverly Y. Sanders	2030237
EPA 365.1 Rev. 2.0 dissolved	10/02/2018			10/02/2018 15:44	Julia Storbeck	2030240
	10/02/2018			10/02/2018 16:19	Julia Storbeck	2030239
	10/02/2018			10/02/2018 16:20	Julia Storbeck	2030241
SM 2120 B	10/02/2018			10/02/2018 15:10	Chelsea Hernandez	2030230, 2030231
	10/02/2018			10/03/2018 09:46	DeAsia Armster	2030232
SM 2320 B-97	10/02/2018			10/06/2018 02:16	Dale L. Simmons	2030231
	10/02/2018			10/06/2018 05:05	Dale L. Simmons	2030230
	10/02/2018			10/06/2018 05:15	Dale L. Simmons	2030232
SM 2540 D-97	10/02/2018			10/05/2018 16:36	Yijie Li	2030230, 2030231, 2030232
SM 4500 F-C-97	10/02/2018			10/06/2018 02:08	Dale L. Simmons	2030231
	10/02/2018			10/06/2018 05:05	Dale L. Simmons	2030230
	10/02/2018			10/06/2018 05:15	Dale L. Simmons	2030232
SM 5310 B-00	10/02/2018			10/05/2018 01:53	Megan Treadwell	2030236
	10/02/2018			10/05/2018 02:05	Megan Treadwell	2030237
	10/02/2018			10/05/2018 02:17	Megan Treadwell	2030238
SM 5310 B-00 dissolved	10/02/2018			10/05/2018 01:04	Megan Treadwell	2030234
	10/02/2018			10/05/2018 01:41	Megan Treadwell	2030235
	10/02/2018			10/05/2018 12:52	Megan Treadwell	2030233