

Quarterly Temperature Verification Regional Operation Centers

DEP SOP FT 1400; Acceptance Criteria for Temp. $\pm 0.5^{\circ}\text{C}$.

Correction factor needed if acceptance criteria are exceeded.

Record all digits displayed for temperature readings. Do not round before reporting measurements.

Meter ID#:	<u>7</u>	Date of Last Verification:	<u>7/27/17</u>
Date:	<u>9/28/17</u>	Time:	<u>1430</u> <u>ETZ</u> / CTZ NIST Reference Device ID#:
<u>4167760</u>			
Cold Bath (between 0 – 10°C): Sonde: <u>7.23</u> °C; NIST: <u>7.10</u> °C; Result: <u>Pass</u> / Fail			
Warm Bath (between 30 – 40°C): Sonde: <u>38.45</u> °C; NIST: <u>38.3</u> °C; Result: <u>Pass</u> / Fail			
Is Correction Factor Needed? Y / <u>(N)</u> If yes, describe: _____			
Person Performing Verification: <u>K. DeArth</u>			

Comments: _____

Meter ID#:	<u>7</u>	Date of Last Verification:	<u>9/28/17</u>
Date:	<u>1/8/18</u>	Time:	<u>1500</u> <u>ETZ</u> / CTZ NIST Reference Device ID#:
<u>4167760</u>			
Cold Bath (between 0 – 10°C): Sonde: <u>1.7</u> °C; NIST: <u>1.7</u> °C; Result: <u>Pass</u> / Fail			
Warm Bath (between 30 – 40°C): Sonde: <u>38.4</u> °C; NIST: <u>38.3</u> °C; Result: <u>Pass</u> / Fail			
Is Correction Factor Needed? Y / <u>(N)</u> If yes, describe: _____			
Person Performing Verification: <u>Gary Shork</u>			

Comments: _____

Meter ID#:	<u>7</u>	Date of Last Verification:	<u>9/18/18</u>
Date:	<u>3/22/18</u>	Time:	<u>1410</u> <u>ETZ</u> / CTZ NIST Reference Device ID#:
<u>4167275</u>			
Cold Bath (between 0 – 10°C): Sonde: <u>5.83</u> °C; NIST: <u>5.50</u> °C; Result: <u>Pass</u> / Fail			
Warm Bath (between 30 – 40°C): Sonde: <u>32.93</u> °C; NIST: <u>32.80</u> °C; Result: <u>Pass</u> / Fail			
Is Correction Factor Needed? Y / <u>(N)</u> If yes, describe: _____			
Person Performing Verification: <u>JOHN BORRINGTON</u>			

Comments: _____