

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID:

President Camacho

RQ:

Project:

CCHMN

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

4/14/22

DO	Name	Date	Time	Temp	Baro	DO	DO	%DO	Probe	Probe	Pass	Lab
DEP SOP			GMT	°C	meter	Gain	mg/L		Offset	Gain		
FT 1500					range	mg/L						
Calibr.	Monica Jaeger	4/20/22	652	25.5	765.4	8.2	8.24	100.8		1.023	P/F	L/F
ICV	"	"	653	25.6	765.6	8.2	8.24	100.8			P/F	L/F
CCV	"	"	1505	26.9	765.4	8.0	8.01	100.6			P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec	Name	Date	Time	Temp	Standard	Water	Pass	Lab
Gain			GMT	°C	DO	Reading		
FT 4000					mg/L	mg/L		
Calibr.	Monica Jaeger	4/20/22	655	22.0309	3/23	50.000	50.000	P/F L/F
ICV	"	"	701	21.1228	12/23	100	103.9	P/F L/F
CCV	"	"	1507	22.0309	3/23	50000	50173	P/F L/F
CCV								P/F L/F

Conductivity Acceptance criteria $\pm 5\%$

pH	Name	Date	Time	Temp	Expt	pH	Temp	Water	mV	Pass	Lab
DEP SOP			GMT	°C	Date	Buffer	°C	Reading			
FT 1000						Buffer		Buffer			
Calibr.	Monica Jaeger	4/20/22	705	11.12468	11/23	7.0	25.7	7.0	✓	P/F	L/F
Calibr.	"	"	708	22.03024	3/24	4.01	25.6	4.01	✓	P/F	L/F
Calibr.	"	"	710	42.02048	6/23	10.0	25.5	10.0	✓	P/F	L/F
ICV	"	"	712	22.03024	3/24	4	25.5	4.0		P/F	L/F
CCV	"	"	1508	42.02048	6/23	10.0	23.8	10.03		P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 $\frac{165}{10}$, slope from 4 to 7 $\frac{180}{3}$ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / **NO** / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification:

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: 3312A ; Date of last verification: 2/24/22

Depth	Name	Date	Time	Gain	DO	Pass	Lab
Depth			GMT	mg/L	mg/L		
Depth							
Pressure mode in air						P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: