

Login Checklist

RQ- 2022-02-01-59

Login Station ID: #4

Shipping Method: FedEx | UPS | HD | Greyhound

Date/Time of Receipt: 02/ 11 /22 @ 9:15

*Cooler Temperature	Samples Frozen?	* All Samples in Cooler Preserved with		Number of Sample Containers in Cooler	Evidence Tape				Waybill Tracking #
					Present?		*Intact?		
	YES	HNO ₃	Formalin		Yes	No	Yes	No	
0.9				10		X			5225 5542 5275

COOLER CHECK

*All samples received on ice unless otherwise noted. HNO3 and Formalin preserved samples do NOT have a temperature requirement.

- If the temperature of a cooler exceeds 6.0 oC (above 10.0 oC for W-1-4-DIOX and microbiology) or received without ice, identify affected samples in an NCR. NCR# _____
- DOH coolers can exceed 2.0 °C above the standard temperature (6.0 or 10.0 °C for W-1-4-DIOX) without login confirmation from customer for samples requiring ice preservation; identify affected samples in an NCR. NCR# _____
- If cooler or samples are received with a damaged Evidence Seal; identify the affected samples in an NCR. NCR# _____
- If coolers or samples are received late; identify the affected samples in an NCR. NCR# _____

Chain of Custody/ Field Sheet(s) Included?

Yes ✓ No _____

Micro-Biology Overflow Chain of Custody/ Field Sheet(s) Included?

Yes _____ No X

CONTAINER CHECK

Evidence Tape on Bottles?

Yes _____ No _____ NA X

Evidence Tape intact? Yes _____ No _____

If Criminal, photographs taken?

Yes _____ No _____

Containers intact? Yes ✓ No _____

Caps on tight?

Yes ✓ No _____

Sufficient Sample Volume:

Yes ✓ No _____

If NO is checked above, generate an NCR listing affected sample IDs in its appropriate category. NCR# _____

CHLORINE CHECK REQUIRED? (Blue dot on container)

Yes _____ No X Init: ZR

Chlorine detected? (One container checked per Field ID)

Yes _____ No _____ Init: _____

If chlorine is detected, generate an NCR listing affected sample IDs . NCR# _____

PRESERVATION CHECK (EXCLUDING SAMPLES SENT TO OVERFLOW LABS, "OV-")

**Acid Preserved Samples: pH < 2.0?

Yes ✓ No _____ NA _____ Init: ZR

**W-1-4-DIOX (Green dot on container) preserved to pH < 4.0?

Yes _____ No _____ NA X Init: ZR

If samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

Coolers Unpacked/Checked by: ZR

Date: 02/ 11 /22

Event contains NCRs (Y/N)? N

Caloosa Creeks

Event ID SO-DIST-2022-02-11-02 (BMAP)

Date Received: 11-FEB-2022 09:15

Login Checklist

Login Preservation Equipment

Lot # & Exp: Date

Additional Comments:

Infrared Thermometer ID:	#4 Temp Gun EXP:10/19/2022
pH Test Strips 0 – 6	N/A
pH Paper 0 – 3	N/A
pH Paper 0 – 13	210620 EXP:04/15/2023
Chlorine Test Strips	0261 EXP:05/2023

ENCORE SAMPLES

S-VOC-MS samples in Encores must be frozen within 48 hours of collection.

Were Encore samples received? Yes _____ Init: _____

Date and time placed in freezer _____

Were all samples placed in freezer within 48 hours of collection? Yes _____ No _____

If no, were samples shipped on dry ice? Yes _____ No _____

If S-VOC-MS samples were not preserved correctly, generate an NCR listing affected sample IDs . NCR# _____

FOR CLEAN LAB USE ONLY – Total Mercury Preservation Check(W-HG-AF;W-HG-AF-F):

Samples preserved within 48 hours? Yes _____ No _____

Date and time samples preserved: _____

Caloosa Creeks

Central Laboratory Submittal Form

last revised Jan 25, 2018

Customer: SO-DIST

Requester: Gary Snorek

Field Report Prepared By: GARY SNOREK

Project ID: BMAP

Collected By: JOHN BARRON

Sampling Agency: FDEP SDRUC 8052

Location		Field ID/WIN ID		G3SD0128		☐ Comp ☒ Grab		Collection (comp begin or grab)		Eastern Central		Composite end		Bottle Group(s)	
Field ID		Location		Orange River @ BoatRamp		Matrix (type, e.g., Salt, Fresh, etc)		Diss. Oxygen		☒ mg/L ☐ % sat		Total Res. Chlorine (mg/L)		A	
WIN/STI		Temp (C)		19.7		pH		7.72		Sample Depth		0.3		☐ ft ☒ m	
Latitude		Longitude		Salinity (ppt)		0.33		Comments							
Location		Field ID/WIN ID		G3SD0146		☐ Comp ☒ Grab		Collection (comp begin or grab)		Eastern Central		Composite end		Bottle Group(s)	
Field ID		Location		Roberts Canal @ SR 80		Matrix (type, e.g., Salt, Fresh, etc)		Diss. Oxygen		☒ mg/L ☐ % sat		Total Res. Chlorine (mg/L)		A	
WIN/STOI		Temp (C)		21.2		pH		8.03		Sample Depth		0.3		☐ ft ☒ m	
Latitude		Longitude		Salinity (ppt)		0.30		Comments							
Location		Field ID/WIN ID		G3SD0088		☐ Comp ☒ Grab		Collection (comp begin or grab)		Eastern Central		Composite end		Bottle Group(s)	
Field ID		Location		Nine Mile Canal @ SR 80		Matrix (type, e.g., Salt, Fresh, etc)		Diss. Oxygen		☒ mg/L ☐ % sat		Total Res. Chlorine (mg/L)		A	
WIN/STORI		Temp (C)		19.2		pH		7.20		Sample Depth		0.3		☐ ft ☒ m	
Latitude		Longitude		Salinity (ppt)		0.23		Comments							
Location		Field ID/WIN ID		G3SD0085		☐ Comp ☒ Grab		Collection (comp begin or grab)		Eastern Central		Composite end		Bottle Group(s)	
Field ID		Location		Bee Branch @ Loblolly Bay		Matrix (type, e.g., Salt, Fresh, etc)		Diss. Oxygen		☒ mg/L ☐ % sat		Total Res. Chlorine (mg/L)		A	
WIN/STOREI		Temp (C)		15.9		pH		6.68		Sample Depth		0.3		☐ ft ☒ m	
Latitude		Longitude		Salinity (ppt)		0.15		Comments							

Relinquished By:	Date/Time	Shipping Method	Number of Coolers Shipped:	Received By:	Date/Time
JOHN BARRON	2/10/22, 1600	FOEX	2	JR	2-11-2022 9:15

* Only received two sites - mr

Group: A	Bottle Type:	P-1L	# of Bottles: 5	Preservative:	ICE	Enter Number of Bottles Sent to Lab	5
	ALKALINITY						
	W-COLOR						
	W-F						
	W-TSS						
Group: A	Bottle Type:	BP-1L	# of Bottles: 5	Preservative:	ICE	Enter Number of Bottles Sent to Lab	5
	CHLSUITE-W						
Group: A	Bottle Type:	P-125ML	# of Bottles: 5	Preservative:	FILTER-H2SO4-I CE	Enter Number of Bottles Sent to Lab	5
	W-DOC						
Group: A	Bottle Type:	P-500ML	# of Bottles: 5	Preservative:	ICE And H2SO4	Enter Number of Bottles Sent to Lab	5
	W-NH3						
	W-NO2NO3						
	W-S-A-TP						
	W-TKN						
	W-TOC						
Group: A	Bottle Type:	P-125ML	# of Bottles: 5	Preservative:	FILTER-ICE	Enter Number of Bottles Sent to Lab	5
	W-PO4-F						

SA1105060

Florida Department of Environmental Protection

Central Laboratory Submittal Form

Customer: SO-DIST

Requester: Gary Snorek

Field Report Prepared By:

GARY SNOREK

Project ID: BMAP

Collected By:

JOHN BARRINGTON

Sampling Agency:

FDEP SURGE 8252

Loc Field ID/WIN ID	G3SD0129		☐ Comp ☒ Grab	Collection (comp begin or grab) Date 2/10/22 Time 1511	☒ Eastern ☐ Central	Composite end Date Time	Bottle Group(s) (See pg 2)	
Field Location	Telegraph Creek @ USGS		Matrix (type, e.g., Salt, Fresh, etc) Fresh		Diss. Oxygen 5.44	☒ mg/L ☐ % sat	Total Res. Chlorine (mg/L)	
WIN			Temp (C) 18.5	pH 7.46	Sample Depth 0.3	☐ ft ☒ m	Sp. Conductance (umho/cm) 562	
Latitude	☐ dd ☐ dms	Longitude	☐ dd ☐ dms	Salinity (ppt) 0.27	Comments			
Location			☐ Comp ☐ Grab	Collection (comp begin or grab) Date Time	☐ Eastern ☐ Central	Composite end Date Time	Bottle Group(s)	
Field ID			Matrix (type, e.g., Salt, Fresh, etc)		Diss. Oxygen	☐ mg/L ☐ % sat	Total Res. Chlorine (mg/L)	
WIN/STORET #			Temp (C)	pH	Sample Depth	☐ ft ☐ m	Sp. Conductance (umho/cm)	
Latitude	☐ dd ☐ dms	Longitude	☐ dd ☐ dms	Salinity (ppt)	Comments			
Location			☐ Comp ☐ Grab	Collection (comp begin or grab) Date Time	☐ Eastern ☐ Central	Composite end Date Time	Bottle Group(s)	
Field ID			Matrix (type, e.g., Salt, Fresh, etc)		Diss. Oxygen	☐ mg/L ☐ % sat	Total Res. Chlorine (mg/L)	
WIN/STORET #			Temp (C)	pH	Sample Depth	☐ ft ☐ m	Sp. Conductance (umho/cm)	
Latitude	☐ dd ☐ dms	Longitude	☐ dd ☐ dms	Salinity (ppt)	Comments			
Location			☐ Comp ☐ Grab	Collection (comp begin or grab) Date Time	☐ Eastern ☐ Central	Composite end Date Time	Bottle Group(s)	
Field ID			Matrix (type, e.g., Salt, Fresh, etc)		Diss. Oxygen	☐ mg/L ☐ % sat	Total Res. Chlorine (mg/L)	
WIN/STORET #			Temp (C)	pH	Sample Depth	☐ ft ☐ m	Sp. Conductance (umho/cm)	
Latitude	☐ dd ☐ dms	Longitude	☐ dd ☐ dms	Salinity (ppt)	Comments			
Location			☐ Comp ☐ Grab	Collection (comp begin or grab) Date Time	☐ Eastern ☐ Central	Composite end Date Time	Bottle Group(s)	
Field ID			Matrix (type, e.g., Salt, Fresh, etc)		Diss. Oxygen	☐ mg/L ☐ % sat	Total Res. Chlorine (mg/L)	
WIN/STORET #			Temp (C)	pH	Sample Depth	☐ ft ☐ m	Sp. Conductance (umho/cm)	
Latitude	☐ dd ☐ dms	Longitude	☐ dd ☐ dms	Salinity (ppt)	Comments			

Relinquished By:	Date/Time	Shipping Method	Number of Coolers Shipped:	Received By:	Date/Time
JOHN BARRINGTON	2/10/22 1600	FedEx	2	JK	2-11-2022 9:15

Group: A Bottle Type: P-1L # of Bottles: 5 Preservative: ICE Enter Number of Bottles Sent to Lab _____

ALKALINITY
W-COLOR
W-F
W-TSS

Group: A Bottle Type: BP-1L # of Bottles: 5 Preservative: ICE Enter Number of Bottles Sent to Lab _____

CHLSUITE-W

Group: A Bottle Type: P-125ML # of Bottles: 5 Preservative: FILTER-H2SO4-I
CE Enter Number of Bottles Sent to Lab _____

W-DOC

Group: A Bottle Type: P-500ML # of Bottles: 5 Preservative: ICE And H2SO4 Enter Number of Bottles Sent to Lab _____

W-NH3
W-NO2NO3
W-S-A-TP
W-TKN
W-TOC

Group: A Bottle Type: P-125ML # of Bottles: 5 Preservative: FILTER-ICE Enter Number of Bottles Sent to Lab _____

W-PO4-F

Date of Request: 10-FEB-2022
 Created By: SNOREK_G On: 10-FEB-2022
 Modified By: AYRES_J On: 10-FEB-2022
 Customer: SO-DIST
 Project: BMAP
 Division: Division of Environmental Assessment and Restoration
 District: South District
 Event Description: Caloosa Creeks

Send Coolers To:

Phone: 239-344-5683
 FLDEP South District Office
 2295 Victoria Ave. Suite 364
 Fort Myers, FL 33901-3381
 Attn: Gary Snorek
 Cooler Ship EMail: Gary.Snorek@dep.state.fl.us

Priority: 3
 Event Status: A
 Criminal Investigation: NO
 Chemistry Request Reviewed By: AYRES_J On: 10-FEB-2022
 Biology Request Reviewed By: AYRES_J On: 10-FEB-2022
 Sampling Kit Required: NO Ship on:
 Sampling Kit Shipped:
 Sampling Kit Packed By:
 Preservatives Needed: NO
 Date to Receive Samples: 10-FEB-2022
 Logged In By:

Send Final Report To:

FLDEP South District Office
 2295 Victoria Ave. Suite 364
 Fort Myers, FL 33901-3381
 Attn: Gary Snorek
 Report Notification EMail:
 Gary.Snorek@Dep.State.fl.us;kirby.Wolfe@Dep.state.fl.us;Nicole
 .reistad@floridaDEP.gov

Comments: caloosa creeks sampling

Bottle Group A (Water) With 5 Samples:

Bottle Type: P-1L	Number of Bottles: 5	Preserved With: ICE
ALKALINITY	Template: DEFAULT	(SM 2320 B-2011) Alkalinity in aqueous matrices as mg CaCO ₃ /L
W-COLOR	Template: DEFAULT	(DEP SOP: NU-094-1) True Color measured at 450 nm (not valid for NPDES monitoring)
*Color (true)		
W-F	Template: DEFAULT	(SM 4500 F-C-2011) Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)
W-TSS	Template: DEFAULT	(SM 2540 D-2011) Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample
Bottle Type: BP-1L	Number of Bottles: 5	Preserved With: ICE
CHLSUITE-W	Template: DEFAULT	(SM 10200 H (mod.)) Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry
Bottle Type: P-125ML	Number of Bottles: 5	Preserved With: FILTER-H ₂ SO ₄ -ICE
W-DOC	Template: DEFAULT	(SM 5310 B-2011) Dissolved Nonpurgeable Organic Carbon in filtered, aqueous matrices
Bottle Type: P-500ML	Number of Bottles: 5	Preserved With: ICE And H ₂ SO ₄
W-NH ₃	Template: DEFAULT	(EPA 350.1 Rev. 2.0) Ammonia in aqueous matrices as mg N/L
W-NO ₂ NO ₃	Template: DEFAULT	(EPA 353.2 Rev. 2.0) Nitrite/Nitrate in aqueous matrices as mg N/L
W-S-A-TP	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Total Phosphorus in aqueous matrices as mg P/L
W-TKN	Template: DEFAULT	(EPA 351.2 Rev. 2.0) Total Kjeldahl Nitrogen in aqueous matrices
W-TOC	Template: DEFAULT	(SM 5310 B-2011) Nonpurgeable Organic Carbon in aqueous matrices
Bottle Type: P-125ML	Number of Bottles: 5	Preserved With: FILTER-ICE
W-PO ₄ -F	Template: DEFAULT	(EPA 365.1 Rev. 2.0) Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L

* - The laboratory is not NELAP certified for this analyte/method, or certification is not applicable.