

YSI Datasonde Datasheet v1.1 – Deployment

SECN Coastal Water Quality Monitoring

Station Name

Clapboard Creek

Calibration QC Pre-Deployment

Date 11-13-2017

QC'd By

Battery Voltage 11.6

Turbidity Wiper Replaced Y / (N)

Wiper parking 180° (Y) N

Replace Batteries Y / N

ODO Wiper Replaced Y / (N)

Wiper parking 180° (Y) N

If Yes, New Voltage

Free memory (days) 185

Battery Life (days) 115

Barometric Pressure 766

pH7 (mv) -22.6

pH10/4 (mv) -192.4 / 148.0

Conductivity Cell Constant 5.1116

ODO Gain 1.0862

pH Slope 169.8

Parameter	Temp (C)	Sp. Cond.	pH7	pH10	pH4	ODO %	Depth (m)	Turbidity (L)	Turbidity (H)
Standard	23.2	56.00	7.00	10.00	4.00	100	0.0	0.0	126
Measured	23.1	56.01	7.00			100.8	0.0	1.3	
Calibrate		56.00	7.01	10.01	4.00	100.9	0.0	0.0	125.8

Start Time (Sonde) 10:00

File Name CC11132017 Start Date 11-13-2017 Start Time (Actual) 10:00

Deployment Information

Deployed By BD/AK

Instrument # JEA 1 Date 11-13-2017

Time 10:25

Notes

Site Observations: Fish Kill ☐ Algal Bloom ☐ Debris ☐ Hurricane / TS ☐ Other ☐

Observation Description

Windy, choppy conditions
~ 70°

Recorded By AKalmbacher Computer Entry By AKalmbacher

Certified By

YSI Datasonde Datasheet v1.1 – Retrieval

Station Name

SECN Coastal Water Quality Monitoring

Clapboard Creek

Retrieval Information:

Retrieved By:

PO/AK

Instrument # JEA 1

Date 11-30-2017 Time 10:23

Barometric Pressure 305.9 mmHg pH7 (mv) -25.4 pH 10 / 4 (mv) 150.7 -195.3

Battery Voltage 12.3 Turbidity wiper parked 180° (Y) / (N) ODO (Y) / (N)

Post Calibration Date 11-30-2017 Cond. Cell Constant 5.1116 pH Slope 169.9

Parameter	Temp (C)	Sp. Cond.	pH 7	pH 10	pH 4	ODO %	Depth (m)	Turbidity (L)	Turbidity (H)
Standard	22.8	50.00	7.00	10.0	4.0	100	0.0	0.0	
Measured	22.65	50.09	7.05	10.06	3.91	100.1	0.03	0.6	

Fouling presence: Please record type of fouling AND amount (eg A, B, E, / H)

Types A= Algae, B= Barnacles, C= Crabs, E= Eggs, F= Fish, O= Other, N= none / Amount H=Heavy, M= Moderate, L= Light, N= None.

Sonde/Guard: N External Screen: N Temp/Cond: N

ODO Probe: N pH Probe: N Turbidity: N

Comments:

Outgoing tide, ~ 70°. Calm
partly cloudy

File Name CC11132017

Any Probe Malfunctions

(Y) / (N)

If Y, please describe above

Download File (Y) / (N)

Date 12/1/2017 Initials AK

Sent Datafile (Y) / (N)

Date 12/1/2017 Initials AK

Sent Datasheet (Y) / (N)

Date 12/1/2017 Initials AK

Recorded By: AKalmbacher

Computer Entry By: AKalmbacher

Certified By:

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page.

Meter SEA 1 RQ- N/A Project Clapboard Creek

- Notes: (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).
 (2) Always wait for meter to stabilize before recording any readings.
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site _____

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 1 / 1

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	<u>AKalmbacher</u>	<u>11/13/17</u>	<u>0845</u>	<u>23.3</u>	<u>8.53</u>	<u>8.60</u>	<u>100.9</u>	<u>—</u>	<u>1.0862</u>		
ICV		<u>↓</u>	<u>0850</u>	<u>23.3</u>	<u>8.53</u>	<u>8.60</u>	<u>100.8</u>		<u>1.0862</u>	<u>P</u>	<u>L</u>
CCV	<u>AKalmbacher</u>	<u>11/30/17</u>	<u>1145</u>	<u>22.6</u>	<u>8.64</u>	<u>8.65</u>	<u>100.1</u>		<u>1.0862</u>	<u>P</u>	<u>L</u>
CCV											

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	<u>AKalmbacher</u>	<u>11/13/17</u>	<u>0852</u>	<u>170720X</u>	<u>8/2018</u>	<u>50.00</u>	<u>50.00</u>		
ICV		<u>↓</u>	<u>0855</u>	<u>" "</u>		<u>50.00</u>	<u>50.01</u>	<u>P</u>	<u>L</u>
CCV		<u>↓</u>	<u>0857</u>	<u>1124820</u>	<u>6/2018</u>	<u>24.82</u>	<u>25.41</u>	<u>P</u>	<u>L</u>
CCV	<u>AKalmbacher</u>	<u>11-30-17</u>	<u>1149</u>	<u>170720X</u>	<u>8/2018</u>	<u>50.00</u>	<u>50.09</u>	<u>P</u>	<u>L</u>

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading mV	Pass / Fail	Lab / Field
Calibr.	<u>AKalmbacher</u>	<u>11/13/17</u>	<u>0900</u>	<u>2708860</u>	<u>7/2019</u>	<u>7.0</u>	<u>7.01 -22.6</u>		
Calibr.		<u>↓</u>	<u>0902</u>	<u>2710816</u>	<u>9/2019</u>	<u>4.0</u>	<u>4.00 -148.0</u>		
Calibr.		<u>↓</u>	<u>0905</u>	<u>2704798</u>	<u>9/2018</u>	<u>10.0</u>	<u>10.01 -192.4</u>		
ICV		<u>↓</u>	<u>0910</u>	<u>2708860</u>	<u>7/2019</u>	<u>7.0</u>	<u>7.00 -22.7</u>	<u>P</u>	<u>L</u>
CCV	<u>AKalmbacher</u>	<u>11-30-17</u>	<u>1155</u>	<u>2708860</u>	<u>7/2019</u>	<u>7.0</u>	<u>7.05 -25.4</u>	<u>P</u>	<u>L</u>
CCV									

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU; mV pH7 Range 0 $\pm$ 50;

mV pH 4 Range +180 $\pm$ 50;

mV pH 10 Range -180 $\pm$ 50;

Slope from 7 to 10 and 4 to 7 must be between 165 and 180mV

Depth (Quarterly)

Date of Last Depth Verification _____

Depth Sensor (Daily)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air	<u>AKalmbacher</u>	<u>11-13-2017</u>	<u>0912</u>	<u>0.000</u>	<u>0.00</u>	<u>P</u>	<u>L</u>

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater