

YSI Datasonde Datasheet v1.1 – Deployment

SECN Coastal Water Quality Monitoring

Station Name

Clapboard Creek

Calibration QC Pre-Deployment

Date 2-4-2019

QC'd By _____

Battery Voltage _____ Turbidity Wiper Replaced Y / (N) Wiper parking 180° (Y) / N

Replace Batteries (Y) / N ODO Wiper Replaced Y / (N) Wiper parking 180° (Y) / N

If Yes, New Voltage 12.8 Free memory (days) 104 Battery Life (days) 117

Barometric Pressure 762.0 pH7 (mv) -324 pH10/4 (mv) -195.6 / 132.2

Conductivity Cell Constant 4.99598 ODO Gain 1.02519 pH Slope 163.2 / 164.6

Parameter	Temp (C)	Sp. Cond.	pH7	pH10	pH4	ODO %	Depth (m)	Turbidity (L)	Turbidity (H)
Standard	21.55	50.00	7.0	10.0	4.0	100	0.0	0.0	126.0
Measured	21.82	49.99	7.05	9.98	4.01	100.2	0.0	0.0	125.2
Calibrate		50.00	7.00	10.00	4.00	100.2	0.0	0.0	127.5

Start Time (Sonde) 10:30

File Name CC02042019 Start Date 02-04-2019 Start Time (Actual) 10:30

Deployment Information

Deployed By BD/AR

Instrument # SEA 1 Date 02-04-2019

Time 11:13

Notes

Very high tide, overcast, misty
~60's°

Site Observations: Fish Kill ☐ Algal Bloom ☐ Debris ☐ Hurricane / TS ☐ Other ☐

Observation Description

Recorded By A. Kalmbacher Computer Entry By A. Kalmbacher

Certified By _____

YSI Datasonde Datasheet v1.1 – Retrieval

SECN Coastal Water Quality Monitoring

Station Name

Clapboard Creek

Retrieval Information:

Retrieved By:

PO/AK

Instrument #

JEA1

Date

2-25-2019

Time

11:35

Barometric Pressure

763.5

mmHg

pH7 (mv)

-34.1

pH 10 / 4 (mv)

196.5 / 125.7

Battery Voltage

11.8

Turbidity wiper parked 180°

☒ Y ☐ N

ODO

☒ Y ☐ N

Post Calibration Date

2-25-2019

Cond. Cell Constant

1.9881

pH Slope

162.4 / 159.8 *

Parameter	Temp (C)	Sp. Cond.	pH 7	pH 10	pH 4	ODO %	Depth (m)	Turbidity (L)	Turbidity (H)
Standard	22.5	50.00	7.00	10.0	4.0	100	0.0	0.0	126.0
Measured	22.0	49.87	7.04	10.0	4.11	102	0.08	0.6	—

Fouling presence: Please record type of fouling AND amount (eg A, B, E, / H)

Types A= Algae, B= Barnacles, C= Crabs, E= Eggs, F= Fish, O= Other, N= none / Amount H=Heavy, M= Moderate, L= Light, N= None.

Sonde/Guard:

F, C

External Screen:

N

Temp/Cond:

N

ODO Probe:

N

pH Probe:

N

Turbidity:

N

Comments:

High Tide, windy ~ 65°

* pH slope out of range

File Name CC02042019

Any Probe Malfunctions

☐ Y ☒ N

If Y, please describe above

Download File

☒ Y ☐ N

Date 2-25-2019

Initials

AK

Sent Datafile

☐ Y ☐ N

Date 2-25-2019

Initials

AK

Sent Datasheet

☐ Y ☐ N

Date 2-25-2019

Initials

AK

Recorded By:

A. Kalmbacher

Computer Entry By:

A. Kalmbacher

Certified By

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: JEAI

RQ-

Project: Clapboard Creek

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass/Fail
Calibr.	<u>AKumbacher</u>	<u>2-4-19</u>	<u>0835</u>	<u>21.6</u>	<u>8.81</u>	<u>760.2</u>	<u>8.81</u>	<u>100.2</u>		<u>1.025</u>	<u>P</u> / <u>F</u> / <u>L</u> / <u>F</u>
ICV			<u>0840</u>	<u>21.7</u>	<u>760.2</u>	<u>8.79</u>	<u>8.79</u>	<u>100.2</u>		<u>1.025</u>	<u>P</u> / <u>F</u> / <u>L</u> / <u>F</u>
CCV		<u>2-25-19</u>	<u>1400</u>	<u>17.2</u>	<u>763.5</u>	<u>9.62</u>	<u>9.82</u>	<u>102.0</u>		<u>1.025</u>	<u>P</u> / <u>F</u> / <u>L</u> / <u>F</u>
CCV											<u>P</u> / <u>F</u> / <u>L</u> / <u>F</u>

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass/Fail
Calibr.	<u>AKumbacher</u>	<u>2-4-19</u>	<u>0815</u>	<u>180919</u>	<u>9/2019</u>	<u>50,000</u>	<u>49,990</u>	<u>P</u> / <u>F</u> / <u>L</u> / <u>F</u>
ICV			<u>0848</u>	<u>1024820</u>	<u>9/2019</u>	<u>24.820</u>	<u>24.580</u>	<u>P</u> / <u>F</u> / <u>L</u> / <u>F</u>
CCV		<u>2-25-19</u>	<u>1405</u>	<u>180919</u>	<u>9/2019</u>	<u>50,000</u>	<u>49,870</u>	<u>P</u> / <u>F</u> / <u>L</u> / <u>F</u>
CCV								<u>P</u> / <u>F</u> / <u>L</u> / <u>F</u>

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass/Fail
Calibr.	<u>AKumbacher</u>	<u>2-4-19</u>	<u>0850</u>	<u>2809841</u>	<u>8/2020</u>	<u>7.</u>	<u>21.9</u>	<u>7.00</u>	<u>-32.4</u>	<u>P</u> / <u>F</u> / <u>L</u> / <u>F</u>
Calibr.			<u>0855</u>	<u>2810832</u>	<u>9/2020</u>	<u>4.</u>	<u>21.8</u>	<u>4.00</u>	<u>132.2</u>	<u>P</u> / <u>F</u> / <u>L</u> / <u>F</u>
Calibr.			<u>0900</u>	<u>1086551</u>	<u>12/2019</u>	<u>10.</u>	<u>21.9</u>	<u>10.00</u>	<u>-195.6</u>	<u>P</u> / <u>F</u> / <u>L</u> / <u>F</u>
ICV			<u>0903</u>	<u>2809841</u>	<u>8/2020</u>	<u>7.0</u>	<u>21.9</u>	<u>7.05</u>	<u>-34.4</u>	<u>P</u> / <u>F</u> / <u>L</u> / <u>F</u>
CCV		<u>2-25-19</u>	<u>1410</u>	<u>2809841</u>	<u>8/2020</u>	<u>7.0</u>	<u>22.3</u>	<u>7.04</u>	<u>-34.1</u>	<u>P</u> / <u>F</u> / <u>L</u> / <u>F</u>
CCV							<u>22.5</u>	<u>10.0</u>	<u>-196.5</u>	<u>P</u> / <u>F</u> / <u>L</u> / <u>F</u>

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 163.2, slope from 4 to 7 164.6 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? (YES) NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification:

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: ; Date of last verification:

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass/Fail
Pressure mode in air	<u>AKumbacher</u>	<u>2-4-19</u>	<u>0905</u>	<u>0.00</u>	<u>0.00</u>	<u>P</u> / <u>F</u> / <u>L</u> / <u>F</u>

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

pH slope out of range