

YSI Datasonde Datasheet v1.1 – Deployment

SECN Coastal Water Quality Monitoring

Station Name

Clayboard Creek

Calibration QC Pre-Deployment

Date 1/2/2019

QC'd By

Battery Voltage 11.4

Turbidity Wiper Replaced Y/N

Wiper parking 180° Y/N

Replace Batteries Y/N

ODO Wiper Replaced Y/N

Wiper parking 180° Y/N

If Yes, New Voltage

Free memory (days) 165

Battery Life (days) 111

Barometric Pressure 762.8

pH7 (mv) -12.6

pH10/4 (mv) -163.4/12.0

Conductivity Cell Constant 4.90729

ODO Gain 0.99301

pH Slope 154.6/150.8

Parameter	Temp (C)	Sp. Cond.	pH7	pH10	pH4	ODO %	Depth (m)	Turbidity (L)	Turbidity (H)
Standard	22.7	50.00	7.0	10.0	4.0	100	0.0	0.0	126
Measured	22.9	50.01	7.10	10.20	3.99	100.2	0.004	0.3	122.5
Calibrate		50.00	7.00	10.04	3.99	100.2	0.00	0.0	125.2

Start Time (Sonde) 10:30

File Name CC01022019

Start Date 01-02-2019

Start Time (Actual) 10:30

Deployment Information

Deployed By BD/AK

Instrument # SEA 2

Date 01-02-2019

Time 11:10

Notes

pH probe barely calibrating in range on 10.0, pH slope out of range

Site Observations: Fish Kill ☐ Algal Bloom ☐ Debris ☐ Hurricane / TS ☐ Other ☐

Observation Description

Low tide, calm surface pt. cloudy
high 70°

Recorded By AKalmbacher

Computer Entry By AKalmbacher

Certified By

YSI Datasonde Datasheet v1.1 – Retrieval

Station Name

SECN Coastal Water Quality Monitoring

Clapboard Crk

Retrieval Information:

Retrieved By:

BD/AK

Instrument # SEA 2 Date 2-4-2019 Time 11:00Barometric Pressure 762.0 mmHg pH7 (mv) -20.5 pH 10 / 4 (mv) -178.6 / 138.2Battery Voltage 10.5 Turbidity wiper parked 180° ☒ Y / ☐ N ODO ☒ Y / ☐ NPost Calibration Date 02-04-2019 Cond. Cell Constant 4.9209 pH Slope 158.7 / 158.7

Parameter	Temp (C)	Sp. Cond.	pH 7	pH 10	pH 4	ODO %	Depth (m)	Turbidity (L)	Turbidity (H)
Standard	<u>20.4</u>	<u>50.00</u>	<u>7.0</u>	<u>10.0</u>	<u>4.0</u>	<u>100</u>	<u>0.0</u>	<u>0.0</u>	<u>126.0</u>
Measured	<u>21.3</u>	<u>49.68</u>	<u>7.19</u>	<u>10.34</u>	<u>4.04</u>	<u>99.4</u>	<u>-0.17</u>	<u>0.8</u>	<u>—</u>

Fouling presence: Please record type of fouling AND amount (eg A, B, E, / H)

Types A= Algae, B= Barnacles, C= Crabs, E= Eggs, F= Fish, O= Other, N= none / Amount H=Heavy, M= Moderate, L= Light, N= None.

Sonde/Guard: C External Screen: C Temp/Cond: NODO Probe: N pH Probe: N Turbidity: N

Comments:

Very High Tide
Overcast, Misty
~ 60's°

File Name CC01022019

Any Probe Malfunctions

☐ Y / ☐ N

If Y, please describe above

Download File

☒ Y / ☐ NDate 2-4-2019

Initials

AK

Sent Datafile

☐ Y / ☐ NDate 2-5-2019

Initials

AK

Sent Datasheet

☐ Y / ☐ NDate 2-5-2019

Initials

AK

Recorded By:

AKalmbacher

Computer Entry By:

AKalmbacher

Certified By

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: SEA2

RQ:

Project: Claphorn Creek

- Notes:** (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	<u>AKalmbacher</u>	<u>1-2-2019</u>	<u>0835</u>	<u>22.9</u>	<u>762.8</u>	<u>8.59</u>	<u>8.60</u>	<u>100.2</u>	<u>—</u>	<u>0.9930</u>	<u>P / F</u>	<u>L / F</u>
ICV			<u>0837</u>	<u>23.0</u>	<u>762.8</u>	<u>8.58</u>	<u>8.58</u>	<u>100.2</u>	<u>—</u>	<u>0.9930</u>	<u>P / F</u>	<u>L / F</u>
CCV	<u>AKalmbacher</u>	<u>2-4-2019</u>	<u>1315</u>	<u>21.2</u>	<u>762.0</u>	<u>8.88</u>	<u>8.80</u>	<u>99.4</u>	<u>—</u>	<u>0.9930</u>	<u>P / F</u>	<u>L / F</u>
CCV											<u>P / F</u>	<u>L / F</u>

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	<u>AKalmbacher</u>	<u>1-2-2019</u>	<u>0840</u>	<u>180919</u>	<u>09/2019</u>	<u>50,000</u>	<u>50,000</u>	<u>P / F</u>	<u>L / F</u>
ICV			<u>0842</u>	<u>1024820</u>	<u>09/2019</u>	<u>24,820</u>	<u>24,440</u>	<u>P / F</u>	<u>L / F</u>
CCV	<u>AKalmbacher</u>	<u>2-4-2019</u>	<u>1317</u>	<u>180919</u>	<u>9/2019</u>	<u>50,000</u>	<u>49,680</u>	<u>P / F</u>	<u>L / F</u>
CCV								<u>P / F</u>	<u>L / F</u>

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	<u>AKalmbacher</u>	<u>1-2-2019</u>	<u>0845</u>	<u>2809841</u>	<u>8/2020</u>	<u>7.</u>	<u>23.0</u>	<u>7.00</u>	<u>-26.1</u>	<u>P / F</u>	<u>L / F</u>
Calibr.			<u>0848</u>	<u>2810832</u>	<u>9/2020</u>	<u>4.</u>	<u>23.0</u>	<u>3.99</u>	<u>142.0</u>	<u>P / F</u>	<u>L / F</u>
Calibr.			<u>0852</u>	<u>1806F51</u>	<u>12/2019</u>	<u>10.</u>	<u>23.0</u>	<u>10.04</u>	<u>-163.4</u>	<u>P / F</u>	<u>L / F</u>
ICV			<u>0900</u>	<u>2809841</u>	<u>8/2020</u>	<u>7.0</u>	<u>23.0</u>	<u>7.10</u>	<u>-15.9</u>	<u>P / F</u>	<u>L / F</u>
CCV	<u>AKalmbacher</u>	<u>2-4-2019</u>	<u>1320</u>	<u>2809841</u>	<u>8/2020</u>	<u>7.0</u>	<u>22.2</u>	<u>7.19</u>	<u>-2.5</u>	<u>P / F</u>	<u>L / F</u>
CCV										<u>P / F</u>	<u>L / F</u>

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 156.8, slope from 4 to 7 154.6 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification:

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: ; Date of last verification:

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	<u>AKalmbacher</u>	<u>1-02-2019</u>	<u>0905</u>	<u>0.00</u>	<u>0.004</u>	<u>P / F</u>	<u>L / F</u>

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: