

SECN Coastal Water Quality Monitoring
Datasonde Datasheet v1.42 (11/2019)—DEPLOYMENT

Station Name

TIMUclap02

Calibration / Pre-Deployment

Instrument (Sonde) # (SN) 9AE
 Calibration Date 9/24/20 Calibrated By Kim Appleby
 Battery Voltage Pre-Calibration 10.2 Battery Voltage >5.3 ☒ Y ☐ N Battery Voltage Post-Calibration 10.2
 Central Wiper Replaced ☐ Y ☒ N Central Wiper Parking 180° ☒ Y ☐ N
 Barometric Pressure (mmHg) 760.3 Conductivity Cell Constant 4.89515 ODO Gain 1.00494
 pH7 (mv) / pH10 (mv) / pH4 (mv)(check) / pH Slope /

Parameter	Temp (°C)	Sp. Cond. (mS/cm)	Depth (m)	ODO (% Sat.)	ODO (mg/l)	pH7	pH10
Standard	<u>23.72</u>	<u>50.00</u>	<u>0.000</u>	<u>100.0</u>	<u>8.5</u>	<u>/</u>	<u>/</u>
Measured	<u>23.67</u>	<u>50.14</u>	<u>0.001</u>	<u>100.1</u>	<u>8.46</u>	<u>/</u>	<u>/</u>
Calibrated	<u>/</u>	<u>50.00</u>	<u>0.000</u>	<u>100.1</u>	<u>8.46</u>	<u>/</u>	<u>/</u>

Parameter	pH4 (check)	Turbidity Low (FNU)	Turbidity High (FNU)	Chl Low (µg/l)	Chl High (µg/l)	BGA-PC Low (µg/l)	BGA-PC High (µg/l)
Standard	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>
Measured	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>
Calibrated	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>

Deployment

Deployed By: Kim Appleby File Name: TIMUclap02-20200924
 (Station name _YYYYMMDD)
 Sonde Start Date: 9/24/20 Sonde Start Time: 1200 Eastern Standard Time (EST)
 Actual Deployment Date: 9/24/20 Actual Time Placed in Water: 1148 (EST)

Comments:

Site Observations: Fish Kill ☐ Algal Bloom ☐ Debris ☐ Hurricane / Tropical Storm ☐ Other ☐

Observation Description:

Recorded By: Kim Appleby Computer Entry By / Date: _____

Certified By / Date: _____

SECN Coastal Water Quality Monitoring
YSI Datasonde Datasheet v1.42 (11/2019)— RETRIEVAL

Station Name

TIMUclap02

Retrieval

In situ Sonde # (SN)

9AE

Retrieved By

Kim Appleby + Jeremy Parrish

Retrieval Date

10/15/20

Retrieval Time

1003

(EST)

Field Meter # (SN)

JEA2

Field Meter Calibration Date

10/15/20

On Site Readings

	Temp (°C)	Baro. Pres. (mmHg)	ODO (% Sat.)	ODO (mg/l)	Sp. Cond. (mS/cm)	Salinity (psu)	pH	Turb. (FNU)	Depth (m)	Chla (µg/l)	BGA (µg/l)
Field Meter	26.32	760.3	79.8 82.6	5.72 6.00	33.07	20.81			3.7		
In situ Sonde	26.31	760.3	75.1	5.39	33.11	20.68			4.066		

Fouling presence: Please record type of fouling **AND** amount (e.g. A, B, E, / H). Types A= Algae, B= Barnacles, C= Crabs, E= Eggs, F= Fish, O= Other, N= none, S= Sediment / Amount H=Heavy, M= Moderate, L= Light, N= None

Sonde/Guard: N External Screen: N Temp/Cond: N

ODO Probe: N pH Probe: / Turbidity: /

Post Deployment

Post Deployment Reading Date

10/15/20

Battery Voltage

10.2

Central Wiper Parking 180°

☒ Y / N

Barometric Pressure (mmHg)

760.3

Conductivity Cell Constant

4.89515

pH7 (mv)

/

pH10 (mv)

/

pH Slope

/

pH4 (mv)

/

Parameter	Temp (°C)	Sp. Cond. (mS/cm)	Depth (m)	ODO (% Sat.)	ODO (mg/l)	pH7	pH10
Standard	23.37	50.00	0.00	100.0	8.50		
Measured (clean sensors)	23.29	50.08	-0.018	99.9	8.49		

Parameter	pH4 (check)	Turbidity Low (FNU)	Turbidity High (FNU)	Chl Low (µg/l)	Chl High (µg/l)	BGA-PE Low (µg/l)	BGA-PE High (µg/l)
Standard							
Measured (clean sensors)							

Any Probe Malfunctions?

Y ☒ N

If Y, describe in the comments.

File Name:

TIMUclap02-20201015

(Station name _YYYYMMDD)

Comments:

Recorded By

Kim Appleby

Computer Entry By / Date

Certified By / Date

Used as "field meter" on 9/24/20

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: **9AE**

RQ-

Project: **Clapboard Creek**

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	K. Appleby	9/24/20	7:17	23.62	760.3	8.5	8.46	100.1	-	1.00494	(P) F	(L) F
ICV	↓	↓	7:20	23.67	760.3	8.5	8.46	100.1	-	1.00494	(P) F	(L) F
CCV	↓	10/15/20	14:50	23.29	760.3	8.5	8.49	99.9	-	1.00494	(P) F	(L) F
CCV											P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Kim Appleby	9/24/20	7:22	200407A	4/2021	50,000	50,000	(P) F	(L) F
ICV	↓	↓	7:24	200407B	4/2021	20,000	19,520	(P) F	(L) F
CCV	↓	10/15/20	14:52	200407A	4/2021	50,000	50,080	(P) F	(L) F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.						7.				P / F	L / F
Calibr.						4.				P / F	L / F
Calibr.						10.				P / F	L / F
ICV										P / F	L / F
CCV										P / F	L / F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: _____

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Kim Appleby	9/24/20	7:25	0.00	0.00	(P) F	(L) F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

used as "field meter" on 10/15/20

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: SEA2

RQ-

Project: Clapboard Creek

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Kim Appleby	10/15/20	713	23.26	761.5	8.5	8.52	100.0	-	1.03983	P/F	L/F
ICV	↓	↓	716	23.41	761.5	8.5	8.51	100.2	-	1.03983	P/F	L/F
CCV	↓	11/3/20	1450	24.12	768.0	8.5	8.76	104.3	-	1.03983	P/F	L/F
CCV			1415								P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Kim Appleby	10/15/20	718	200407A	4/2021	50,000	50,000	P/F	L/F
ICV	↓	↓	720	200407B	4/2021	20,000	19,260	P/F	L/F
CCV	↓	11/3/20	1454	200407A	4/2021	50,000	49,220	P/F	L/F
CCV			1417					P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.						7.				P/F	L/F
Calibr.						4.				P/F	L/F
Calibr.						10.				P/F	L/F
ICV										P/F	L/F
CCV										P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 $\pm$ 50; mV pH 4 Range +180 $\pm$ 50; mV pH 10 Range -180 $\pm$ 50;

If mV are recorded: slope from 7 to 10 , slope from 4 to 7 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification:

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: ; Date of last verification:

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Kim Appleby	10/15/20	725	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≥ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: