

SECN Coastal Water Quality Monitoring
Datasonde Datasheet v1.42 (11/2019)—DEPLOYMENT

Station Name

TIMUclap02

Calibration / Pre-Deployment

Instrument (Sonde) # (SN) 3AD

Calibration Date

1/18/20

Calibrated By

Kim Appleby

Battery Voltage Pre-Calibration

11.8

Battery Voltage >5.3

Y/N

Battery Voltage Post-Calibration

11.8

Central Wiper Replaced

Y/N

Central Wiper Parking 180°

Y/N

Barometric Pressure (mmHg)

763.5

Conductivity Cell Constant

4.86767

ODO Gain

1.02262

pH7 (mv)

/

pH10 (mv)

/

pH4 (mv)(check)

/

pH Slope

/

Parameter	Temp (°C)	Sp. Cond. (mS/cm)	Depth (m)	ODO (% Sat.)	ODO (mg/l)	pH7	pH10
Standard	<u>21.48</u>	<u>50.00</u>	<u>0.00</u>	<u>100.0</u>	<u>8.9</u>	<u>/</u>	<u>/</u>
Measured	<u>21.30</u>	<u>50.02</u>	<u>0.00</u>	<u>100.7</u>	<u>8.91</u>	<u>/</u>	<u>/</u>
Calibrated	<u>/</u>	<u>50.00</u>	<u>0.00</u>	<u>100.2</u>	<u>8.86</u>	<u>/</u>	<u>/</u>

Parameter	pH4 (check)	Turbidity Low (FNU)	Turbidity High (FNU)	Chl Low (µg/l)	Chl High (µg/l)	BGA-PC Low (µg/l)	BGA-PC High (µg/l)
Standard	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>
Measured	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>
Calibrated	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>

Deployment

Deployed By: Kim Appleby + Kelly McDonald

File Name: TIMUclap02-20210118

(Station name_YYYYMMDD)

Sonde Start Date: 1/18/21

Sonde Start Time: 1700

Eastern Standard Time (EST)

Actual Deployment Date: 1/18/21

Actual Time Placed in Water: 1656

(EST)

Comments:

Site Observations: Fish Kill ☐ Algal Bloom ☐ Debris ☐ Hurricane / Tropical Storm ☐ Other ☐

Observation Description:

Recorded By: Kim Appleby

Computer Entry By / Date:

Certified By / Date:

SECN Coastal Water Quality Monitoring
YSI Datasonde Datasheet v1.42 (11/2019)— RETRIEVAL

Station Name

TIMUclap02

Retrieval

In situ Sonde # (SN)

JEAZ

Retrieved By

Kim Appleby + Brian Davis

Retrieval Date

2/4/21

Retrieval Time

1525 (EST)

Field Meter # (SN)

9AE

Field Meter Calibration Date

2/4/21

On Site Readings

	Temp (°C)	Baro. Pres. (mmHg)	ODO (% Sat.)	ODO (mg/l)	Sp. Cond. (mS/cm)	Salinity (psu)	pH	Turb. (FNU)	Depth (m)	Chla (µg/l)	BGA (µg/l)
Field Meter	13.45	/	97.2	8.57	41.76	26.79	/	/	3.56	/	/
In situ Sonde	13.39	/	88.2	7.83	40.92	26.19	/	/	3.78	/	/

Fouling presence: Please record type of fouling **AND** amount (e.g. A, B, E, / H). Types A= Algae, B= Barnacles, C= Crabs, E= Eggs, F= Fish, O= Other, N= none, S= Sediment / Amount H=Heavy, M= Moderate, L= Light, N= None

Sonde/Guard: S External Screen: — Temp/Cond: N

ODO Probe: N pH Probe: — Turbidity: —

Post Deployment

Post Deployment Reading Date

2/4/21

Battery Voltage

11.2

Central Wiper Parking 180°

☒ Y / ☐ N

Barometric Pressure (mmHg)

762.4

Conductivity Cell Constant

4.86767

pH7 (mv)

/

pH10 (mv)

/

pH Slope

/

pH4 (mv)

/

Parameter	Temp (°C)	Sp. Cond. (mS/cm)	Depth (m)	ODO (% Sat.)	ODO (mg/l)	pH7	pH10
Standard	19.02	50.00	0.00	100.0	9.3	/	/
Measured (clean sensors)	19.04	49.34	-0.11	101.3	9.34	/	/

Parameter	pH4 (check)	Turbidity Low (FNU)	Turbidity High (FNU)	Chl Low (µg/l)	Chl High (µg/l)	BGA-PE Low (µg/l)	BGA-PE High (µg/l)
Standard	/	/	/	/	/	/	/
Measured (clean sensors)	/	/	/	/	/	/	/

Any Probe Malfunctions?

☐ Y / ☒ N

If Y, describe in the comments.

File Name:

TIMUclap02-20210118

(Station name _YYYYMMDD)

Comments:

Recorded By

Kim Appleby

Computer Entry By / Date

Certified By / Date

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: TEA7

RQ:

Project:

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 10/14/20

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Kim Appleby	11/18/21	1522	21.26	763.5	8.9	8.86	100.2	-	1.0267	(P) F	(L) F
ICV	↓	↓	1524	21.30	763.5	8.9	8.91	100.7	-	1.0262	(P) F	(L) F
CCV	↓	2/4/21	1735	19.04	762.4	9.3	9.34	101.3	-	1.0262	(P) F	(L) F
CCV											P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	Kim Appleby	11/18/21	1526	201222L	1/2022	50.000	50.000	(P) F	(L) F
ICV	↓	↓	1528	200107B	4/2021	20,000	19,600	(P) F	(L) F
CCV	↓	2/4/21	1737	201222L	1/2022	50.000	49,340	(P) F	(L) F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.						7.				P / F	L / F
Calibr.						4.				P / F	L / F
Calibr.						10.				P / F	L / F
ICV										P / F	L / F
CCV										P / F	L / F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: _____

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Kim Appleby	11/18/21	1533	0.00	0.00	(P) F	(L) F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: 9AERQ- Project: Clapboard

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Kim Appleby	2/4/21	908	20.23	764.0	9.1	9.10	100.4	-	1.0029	P/F	L/F
ICV	↓	↓	910	20.23	764.0	9.1	9.10	100.9	-	1.0029	P/F	L/F
CCV		2/15/21	1357	22.01	759.5	8.7	8.57	98.0	-	1.0029	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Kim Appleby	2/4/21	912	201222L	1/2022	50,000	50,000	P/F	L/F
ICV	↓	↓	914	200407B	4/2021	20,000	19,590	P/F	L/F
CCV	↓	2/15/21	1359	201222L	1/2022	50,000	49,850	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.						7.				P/F	L/F
Calibr.						4.				P/F	L/F
Calibr.						10.				P/F	L/F
ICV										P/F	L/F
CCV										P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: _____

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Kim Appleby	2/4/21	915	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: