

SECN Coastal Water Quality Monitoring
Datasonde Datasheet v1.42 (11/2019)—DEPLOYMENT

Station Name

TIMUclap02

Calibration / Pre-Deployment

Instrument (Sonde) # (SN) 18E101511 (LIL0)

Calibration Date 4/5/21

Calibrated By Kim Appleby

Battery Voltage Pre-Calibration 2.40 Battery Voltage >5.3 Y / ~~N~~ Battery Voltage Post-Calibration 2.40

Central Wiper Replaced Y / ~~N~~ Central Wiper Parking 180° Y / N

Barometric Pressure (mmHg) 764.8 Conductivity Cell Constant 0.470655 ODO Gain 1.084118

pH7 (mv) — pH10 (mv) — pH4 (mv)(check) — pH Slope —

Parameter	Temp (°C)	Sp. Cond. (mS/cm)	Depth (m)	ODO (% Sat.)	ODO (mg/l)	pH7	pH10
Standard	<u>24.07</u>	<u>50.00</u>	<u>0.00</u>	<u>100.0</u>	<u>9.0</u>		
Measured	<u>24.05</u>	<u>50.22</u>	<u>0.00</u>	<u>100.7</u>	<u>8.95</u>		
Calibrated		<u>50.00</u>	<u>0.00</u>	<u>100.6</u>	<u>8.95</u> N/A		

Parameter	pH4 (check)	Turbidity Low (FNU)	Turbidity High (FNU)	Chl Low (µg/l)	Chl High (µg/l)	BGA-PC Low (µg/l)	BGA-PC High (µg/l)
Standard							
Measured							
Calibrated							

Deployment

Deployed By: Kim Appleby + Brian Davis File Name: TIMUclap02-20210405
(Station name_YYYYMMDD)

Sonde Start Date: 4/5/21 Sonde Start Time: 1045 Eastern Standard Time (EST)

Actual Deployment Date: 4/5/21 Actual Time Placed in Water: 1048 (EST)

Comments:

Site Observations: Fish Kill ☐ Algal Bloom ☐ Debris ☐ Hurricane / Tropical Storm ☐ Other ☐

Observation Description:

Recorded By: Kim Appleby Computer Entry By / Date: _____

Certified By / Date: _____

SECN Coastal Water Quality Monitoring
YSI Datasonde Datasheet v1.42 (11/2019)— RETRIEVAL

Station Name

TIMUclap02

Retrieval

In situ Sonde # (SN)

L1LO-18E101511

Retrieved By

Kim Appleby + Brian Davis

Retrieval Date

4/22/21

Retrieval Time

1011

(EST)

Field Meter # (SN)

18J103752 (Mow)

Field Meter Calibration Date

4/22/21

On Site Readings

	Temp (°C)	Baro. Pres. (mmHg)	ODO (% Sat.)	ODO (mg/l)	Sp. Cond. (mS/cm)	Salinity (psu)	pH	Turb. (FNU)	Depth (m)	Chla (µg/l)	BGA (µg/l)
Field Meter	21.26	764.7	87.8	6.79	37.08	23.51	/	/	2.85	/	/
In situ Sonde	20.054	764.7	94.6 66.6	8.5 5.14	35.43	22.21	/	/	12.882**	/	/

Fouling presence: Please record type of fouling **AND** amount (e.g. A, B, E, / H). Types A= Algae, B= Barnacles, C= Crabs, E= Eggs, F= Fish, O= Other, N= none, S= Sediment / Amount H=Heavy, M= Moderate, L= Light, N= None

Sonde/Guard: N External Screen: - Temp/Cond: N

ODO Probe: N pH Probe: - Turbidity: -

Post Deployment

Post Deployment Reading Date

4/22/21

Battery Voltage

2.36

Central Wiper Parking 180°

☒ Y / ☐ N

Barometric Pressure (mmHg)

763.6

Conductivity Cell Constant

0.470655

pH7 (mv)

/

pH10 (mv)

/

pH Slope

/

pH4 (mv)

/

Parameter	Temp (°C)	Sp. Cond. (mS/cm)	Depth (m)	ODO (% Sat.)	ODO (mg/l)	pH7	pH10
Standard	22.71	50.00	0.00	100.0	8.9	/	/
Measured (clean sensors)	22.52	49.99	0.00*	100.8	8.90	/	/

Parameter	pH4 (check)	Turbidity Low (FNU)	Turbidity High (FNU)	Chl Low (µg/l)	Chl High (µg/l)	BGA-PE Low (µg/l)	BGA-PE High (µg/l)
Standard	/	/	/	/	/	/	/
Measured (clean sensors)	/	/	/	/	/	/	/

Any Probe Malfunctions?

☒ Y / ☐ N

File Name:

TIMUclap02-20210422

If Y, describe in the comments.

(Station name_YYYYMMDD)

Comments: ** Depth reading at 10.xx post deployment. Recorded data for depth incorrect.
* Read 0.00 after re-calibrating post deployment.

Recorded By Kim Appleby

Computer Entry By / Date

Certified By / Date

used as "field meter" on 4/5/21

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: L.T.L.O

RQ-

Project: RMAP
Lapboard Creek

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 2/15/21

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Kim Appleby	4/5/21	947	21.1	764.8	9.0	8.95	100.6	-	1.084118	P/F	L/F
ICV	↓	↓	949	21.2	764.8	9.0	8.95	100.7	-	1.084118	P/F	L/F
CCV	↓	4/22/21	1424	21.45	763.6	8.9	8.90	100.8	-	1.084118	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	Kim Appleby	4/5/21	955	201222L	1/2022	50,000	50,000	P/F	L/F
ICV	↓	4/5/21	957	200407B	4/2021	20,000	19,613	P/F	L/F
CCV	↓	4/22/21	1426	201222L	1/2022	50,000	49.99	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$.

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.						7.				P/F	L/F
Calibr.						4.				P/F	L/F
Calibr.						10.				P/F	L/F
ICV										P/F	L/F
CCV										P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 $\pm$ 50; mV pH 4 Range +180 $\pm$ 50; mV pH 10 Range -180 $\pm$ 50;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 2/15/21

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Kim Appleby	4/5/21	1010	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

used as "field meter" on ~~5/22/21~~ ^{4/22/21} ~~5/22/21~~

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Moe

RQ:

Project: Clayboard Creek

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 2/15/21

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Kim Appleby	4/22/21	843	22.3	763.5	8.7	8.76	100.5	-	1.07724	P/F	L/F
ICV	↓	4/22/21	846	22.4	763.6	8.7	8.76	101.0	-	1.07724	P/F	L/F
CCV	↓	5/10/21	1656				8.27	101.2	-	1.07724	P/F	L/F
CCV		5/10/21	1723	25.38	759.4	8.2	8.26	100.7	-	1.07724	P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	Kim Appleby	4/22/21	850	201222L	1/2022	50,000	50,000	P/F	L/F
ICV	↓	↓	851	200407B	4/2021	20,000	19,751	P/F	L/F
CCV	↓	5/10/21	1741	201222L	1/2022	50,000	52,090	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.						7.				P/F	L/F
Calibr.						4.				P/F	L/F
Calibr.						10.				P/F	L/F
ICV										P/F	L/F
CCV										P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 2/15/21

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Kim Appleby	4/22/21	855	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: