

Used as "field meter" on 8/2/21

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: E.L.S.A.

RQ-

Project: Clipboard

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 7/8/21

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Kim Appleby	8/2/21	913	20.3	761.5	9.1	—	100.2	—	1.0126	☒ P / ☒ F	☒ L / ☒ F
ICV	↓	↓	918	20.5	761.5	8.7	8.65	100.3	—	1.0126	☒ P / ☒ F	☒ L / ☒ F
CCV	↓	↓	932	22.3	761.4			99.5			P / F	L / F
CCV	↓	8/19/21	944	21.04	762.4	8.9	9.97	100.8	—	1.0126	☒ P / ☒ F	☒ L / ☒ F

DO Acceptance criteria from Table ± 0.3 mg/L. 1445 Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	Kim Appleby	8/2/21	918	210426E	5/22	50,000	50,000	☒ P / ☒ F	☒ L / ☒ F
ICV	↓	↓	920	210426D	5/22	20,000	19,930	☒ P / ☒ F	☒ L / ☒ F
CCV	↓	8/19/21	922	210426E	5/22	50,000	50,134	☒ P / ☒ F	☒ L / ☒ F
CCV			1448					P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.						7.				P / F	L / F
Calibr.						4.				P / F	L / F
Calibr.						10.				P / F	L / F
ICV										P / F	L / F
CCV										P / F	L / F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 $\pm$ 50; mV pH 4 Range +180 $\pm$ 50; mV pH 10 Range -180 $\pm$ 50;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 5/27/21

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Kim Appleby	8/2/21	930	0.00	0.00	☒ P / ☒ F	☒ L / ☒ F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

Used as "Field meter" on 8/18/21

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Moc

RQ-

Project: BMAP (lapboard)

Creek

- Notes: (1) Always wait for meter to stabilize before recording any readings.
(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 8/16/21

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Kim Appleby	8/18/21	844	22.6	764.4	8.7	8.7 NA	100.6	-	1.083	P/F	P/F
ICV		↓	846	22.8	764.5	8.7	8.69	100.6	-	1.083	P/F	P/F
CCV		8/30/21	1340	22.8	760.8		8.64	98.5	-	1.083	P/F	P/F
CCV		↓	1355	23.51	760.4	8.5	8.64	101.7	-	1.083	P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Kim Appleby	8/18/21	856	210426E	5/22	50,000	50,000	P/F	P/F
ICV		↓	858	210426D	5/22	20,000	19,904	P/F	P/F
CCV		8/30/21	1415	210426E	5/22	50,000	49,807	P/F	P/F
CCV		8/30/21	1418	210426L	5/22	1413	1,449	P/F	P/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.						7.				P/F	L/F
Calibr.						4.				P/F	L/F
Calibr.						10.				P/F	L/F
ICV										P/F	L/F
CCV										P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 $\pm$ 50; mV pH 4 Range $\pm 180 \pm 50$; mV pH 10 Range $\pm 180 \pm 50$;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 8/19/21

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Kim Appleby	8/18/21	850	0.00	0.00	P/F	P/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

SECN Coastal Water Quality Monitoring
Datasonde Datasheet v1.42 (11/2019)—DEPLOYMENT

Station Name

TIMUclap02

Calibration / Pre-Deployment

Instrument (Sonde) # (SN) E.L.S.A - 20D101688
 Calibration Date 8/2/21 Calibrated By Kim Appleby
 Battery Voltage Pre-Calibration 2.61 Battery Voltage >5.3 (Ex03) Y ☒ N ☐ Battery Voltage Post-Calibration 2.61
 Central Wiper Replaced Y ☒ N ☐ Central Wiper Parking 180° (X) N ☒
 Barometric Pressure (mmHg) 761.5 Conductivity Cell Constant 0.472899 ODO Gain 1.0126
 pH7 (mv) pH10 (mv) pH4 (mv)(check) pH Slope

Parameter	Temp (°C)	Sp. Cond. (mS/cm)	Depth (m)	ODO (% Sat.)	ODO (mg/l)	pH7	pH10
Standard	<u>21.92</u>	<u>50.00</u>	<u>0.00</u>	<u>100.0</u>	<u>8.7</u>		
Measured	<u>19.821.86</u>	<u>49.99</u>	<u>0.00</u>	<u>99.5</u>	<u>8.65</u>		
Calibrated		<u>50.00</u>	<u>0.00</u>	<u>100.2</u>	<u>N/A</u>		

Parameter	pH4 (check)	Turbidity Low (FNU)	Turbidity High (FNU)	Chl Low (µg/l)	Chl High (µg/l)	BGA-PC Low (µg/l)	BGA-PC High (µg/l)
Standard							
Measured							
Calibrated							

Deployment

Deployed By: K. Appleby File Name: TIMUclap02 - 20210302
 (Station name _YYYYMMDD)
 Sonde Start Date: 8/2/21 Sonde Start Time: 1119 Eastern Standard Time (EST)
 Actual Deployment Date: 8/2/21 Actual Time Placed in Water: 1111 (EST)

Comments:

Site Observations: Fish Kill ☐ Algal Bloom ☒ Debris ☐ Hurricane / Tropical Storm ☐ Other ☐

Observation Description:

Algae bloom at site, samples to surface water samples taken

Recorded By: Kim Appleby Computer Entry By / Date:

Certified By / Date:

SECN Coastal Water Quality Monitoring
YSI Datasonde Datasheet v1.42 (11/2019)— RETRIEVAL

Station Name

TIMUclap02

Retrieval

In situ Sonde # (SN)

^{BU}
~~MOE~~ ELSA

Retrieved By

BZvis K Appleby

Retrieval Date

8/18/21

Retrieval Time

1050

(EST)

Field Meter # (SN)

MOE

Field Meter Calibration Date

~~MOE~~ 8/18/21

On Site Readings

	Temp (°C)	Baro. Pres. (mmHg)	ODO (% Sat.)	ODO (mg/l)	Sp. Cond. (mS/cm)	Salinity (psu)	pH	Turb. (FNU)	Depth (m)	Chla (µg/l)	BGA (µg/l)
Field Meter	29.30	765.9	67.0	4.72	24.967	15.13	/	/	2.72	/	/
In situ Sonde	29.40		47.9	3.77	24.074	14.52	/	/		/	/

Fouling presence: Please record type of fouling AND amount (e.g. A, B, E, / H). Types A= Algae, B= Barnacles, C= Crabs, E= Eggs, F= Fish, O= Other, N= none, S= Sediment / Amount H=Heavy, M= Moderate, L= Light, N= None

Sonde/Guard: S light External Screen: IV Temp/Cond: IV

ODO Probe: IV A (light film) pH Probe: N Turbidity: IV

Post Deployment

Post Deployment Reading Date

8/19/21

Battery Voltage

2.58

Central Wiper Parking 180°

☒ N

Barometric Pressure (mmHg)

762.3

Conductivity Cell Constant

0.472899

pH7 (mv)

/

pH10 (mv)

/

pH Slope

/

pH4 (mv)

/

Parameter	Temp (°C)	Sp. Cond. (mS/cm)	Depth (m)	ODO (% Sat.)	ODO (mg/l)	pH7	pH10
Standard	22.38	50,000	0.000	8.9	100.0	/	/
Measured (clean sensors)	22.50	50,134	0.006	8.97	100.8	/	/

Parameter	pH4 (check)	Turbidity Low (FNU)	Turbidity High (FNU)	Chl Low (µg/l)	Chl High (µg/l)	BGA-PE Low (µg/l)	BGA-PE High (µg/l)
Standard							
Measured (clean sensors)							

Any Probe Malfunctions?

☒ N

If Y, describe in the comments.

File Name:

TIMUclap02-20210802

(Station name_YYYYMMDD)

Comments:

Recorded By

Kim Appleby

Computer Entry By / Date

Certified By / Date