

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID:

ELSA

RQ:

—

Project:

BMAP clipboard

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 9/8/23

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	KV Hinges	11/6/23	834	22.2	762.8	8.7	—	100.4	—	1.012	(P) F	(L) F
ICV			837	22.3	762.8	8.7	8.75	100.5	—	1.012	(P) F	(L) F
CCV		12/11/23	1014	20.53	765	8.9	8.91	99.6	—	1.012	(P) F	(L) F
CCV						9.1	9.01	100.1	(2)		P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	KV Hinges	11/6/23	839	230131C	2/24	50000	50,000	(P) F	(L) F
ICV			845	230131D	2/24	15000	15033	(P) F	(L) F
CCV		12/11/23	1618	230612C	6/24	50000	50340	(P) F	(L) F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	KV					7.				P / F	L / F
Calibr.						4.				P / F	L / F
Calibr.						10.				P / F	L / F
ICV										P / F	L / F
CCV										P / F	L / F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 9/8/23

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	KV Hinges	11/6/23	846	0.000	0.000	(P) F	(L) F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

SECN Coastal Water Quality Monitoring
 Datasheet v1.42 (11/2019)—DEPLOYMENT

Station Name

TIMUclap02

Calibration / Pre-Deployment

Instrument (Sonde) # (SN) ELSA

Calibration Date 11/6/23

Calibrated By KVillinger

Battery Voltage Pre-Calibration 2.46

Battery Voltage >5.3 ☒ Y ☐ N

Battery Voltage Post-Calibration 2.41

Central Wiper Replaced ☒ Y ☐ N

Central Wiper Parking 180° ☒ Y ☐ N

Barometric Pressure (mmHg) 762.8

Conductivity Cell Constant 0.471

ODO Gain 1.012

pH7 (mv) /

pH10 (mv) /

pH4 (mv)(check) /

pH Slope /

Parameter	Temp (°C)	Sp. Cond. (mS/cm)	Depth (m)	ODO (% Sat.)	ODO (mg/l)	pH7	pH10
Standard	<u>22.60</u>	<u>50.000</u>	<u>0.000</u>	<u>100.0</u>	<u>8.7</u>	<u>/</u>	<u>/</u>
Measured	<u>22.598</u>	<u>50.003</u>	<u>0.000</u>	<u>100.5</u>	<u>8.75</u>	<u>/</u>	<u>/</u>
Calibrated	<u>/</u>	<u>50.000</u>	<u>0.000</u>	<u>100.4</u>	<u>-</u>	<u>/</u>	<u>/</u>

Parameter	pH4 (check)	Turbidity Low (FNU)	Turbidity High (FNU)	Chl Low (µg/l)	Chl High (µg/l)	BGA-PC Low (µg/l)	BGA-PC High (µg/l)
Standard	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>
Measured	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>
Calibrated	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>	<u>/</u>

Deployment

Deployed By: KVillinger, Haylee Brown File Name: CC11062023

(Station name _YYYYMMDD)

Sonde Start Date: 11/6/23 Sonde Start Time: 1145 Eastern Standard Time (EST)

Actual Deployment Date: 11/6/23 Actual Time Placed in Water: 1146 (EST)

Comments:

Site Observations: Fish Kill ☐ Algal Bloom ☐ Debris ☐ Hurricane / Tropical Storm ☐ Other ☐

Observation Description:

Recorded By: KVillinger Computer Entry By / Date: _____

Certified By / Date: _____

SECN Coastal Water Quality Monitoring
YSI Datasonde Datasheet v1.42 (11/2019)— **RETRIEVAL**

Station Name

TIMUclap02

Retrieval

In situ Sonde # (SN)

ELSA

Retrieved By

K. Villings, J. Parrish

Retrieval Date

12/11/23

Retrieval Time

1107

(EST)

Field Meter # (SN)

L110

Field Meter Calibration Date

12/11/23

On Site Readings

	Temp (°C)	Baro. Pres. (mmHg)	ODO (% Sat.)	ODO (mg/l)	Sp. Cond. (mS/cm)	Salinity (psu)	pH	Turb. (FNU)	Depth (m)	Chla (µg/l)	BGA (µg/l)
Field Meter	17.22	760.7	88.3	7.79	24.000	14.66 3.285	X	X	3.274	X	X
In situ Sonde							X	X		X	X

Fouling presence: Please record type of fouling **AND** amount (e.g. A, B, E, / H). Types A= Algae, B= Barnacles, C= Crabs, E= Eggs, F= Fish, O= Other; N= none, S = Sediment / Amount H=Heavy, M= Moderate, L= Light, N= None

Sonde/Guard: F, B, C, A, D External Screen: _____ Temp/Cond: _____

ODO Probe: _____ pH Probe: _____ Turbidity: _____

Post Deployment

Post Deployment Reading Date

12/11/23

Battery Voltage

1.1A

Central Wiper Parking 180°

Y Y N

Barometric Pressure (mmHg)

765

Conductivity Cell Constant

5.471

pH7 (mv)

/

pH10 (mv)

/

pH Slope

/

pH4 (mv)

/

Parameter	Temp (°C)	Sp. Cond. (mS/cm)	Depth (m)	ODO (% Sat.)	ODO (mg/l)	pH7	pH10
Standard	21.62	50.000	0.000	100.0	9.1	X	X
Measured (clean sensors)	21.59	50.340	0.025	100.1	9.01	X	X

Parameter	pH4 (check)	Turbidity Low (FNU)	Turbidity High (FNU)	Chl Low (µg/l)	Chl High (µg/l)	BGA-PE Low (µg/l)	BGA-PE High (µg/l)
Standard							
Measured (clean sensors)							

Any Probe Malfunctions?

Y / N

If Y, describe in the comments.

File Name:

CC11062023

(Station name _YYYYMMDD)

Comments:

Sonde batteries depleted during deployment.
Had to change batteries pre-decat.
Will be a gap in data.

Recorded By

K. Villings

Computer Entry By / Date

Certified By / Date