

SECN Coastal Water Quality Monitoring
Datasonde Datasheet v1.42 (11/2019)—DEPLOYMENT

Station Name

TIMUclap02

Calibration / Pre-Deployment

Instrument (Sonde) # (SN) M02
Calibration Date 7/22/24 Calibrated By K. Villings
Battery Voltage Pre-Calibration 2.46 Battery Voltage >5.3 ☒ Y / ☐ N Battery Voltage Post-Calibration 2.43
Central Wiper Replaced ☒ Y / ☐ N Central Wiper Parking 180° ☒ Y / ☐ N
Barometric Pressure (mmHg) 766.6 Conductivity Cell Constant 0.47 ODO Gain 1.07
pH7 (mv) / pH10 (mv) / pH4 (mv)(check) / pH Slope /

Parameter	Temp (°C)	Sp. Cond. (mS/cm)	Depth (m)	ODO (% Sat.)	ODO (mg/l)	pH7	pH10
Standard	<u>21.00</u>	<u>50.00</u>	<u>0.135</u>	<u>100.0</u>	<u>9.1</u>		
Measured	<u>20.969</u>	<u>49.99</u>	<u>0.135</u>	<u>100.4</u>	<u>9.12</u>		
Calibrated		<u>50.00</u>	<u>0.135</u>	<u>100.4</u>			

Parameter	pH4 (check)	Turbidity Low (FNU)	Turbidity High (FNU)	Chl Low (µg/l)	Chl High (µg/l)	BGA-PC Low (µg/l)	BGA-PC High (µg/l)
Standard							
Measured							
Calibrated							

Deployment

Deployed By: K. Villings Hayler Bam File Name: CC07222024
Alc. Rosen (Station name YYYYMMDD)
Sonde Start Date: 7/22/24 Sonde Start Time: 1315 Eastern Standard Time (EST)
Actual Deployment Date: 7/22/24 Actual Time Placed In Water: 1305 (EST)

Comments:

Site Observations: Fish Kill ☐ Algal Bloom ☐ Debris ☐ Hurricane / Tropical Storm ☐ Other ☐

Observation Description:

Recorded By: K. Villings Computer Entry By / Date: _____

Certified By / Date: _____

SECN Coastal Water Quality Monitoring
YSI Datasonde Datasheet v1.42 (11/2019) — RETRIEVAL

Station Name

TIMUclap02

Retrieval

In situ Sonde # (SN)

1053
Moe

Retrieved By

Jeremy Parrish, Kyle Coker

Retrieval Date

8/14/24

Retrieval Time

1050 (EST)

Field Meter # (SN)

DEST

Field Meter Calibration Date

8/14/24

On Site Readings

	Temp (°C)	Baro. Pres. (mmHg)	ODO (% Sat.)	ODO (mg/l)	Sp. Cond. (mS/cm)	Salinity (psu)	pH	Turb. (FNU)	Depth (m)	Chla (µg/l)	BGA (µg/l)
Field Meter	29.5	761.5	84.1	5.8	29875	18.39			3.24		
In situ Sonde	29.7	761.5	76.7	5.3	27913	17.07			3.58		

Fouling presence: Please record type of fouling AND amount (e.g. A, B, E, / H). Types A= Algae, B= Barnacles, C= Crabs, E= Eggs, F= Fish, O= Other, N= none, S= Sediment / Amount H=Heavy, M= Moderate, L= Light, N= None

Sonde/Guard: Light External Screen: Temp/Cond:

ODO Probe: pH Probe: Turbidity:

Post Deployment

Post Deployment Reading Date

8/15/24

Battery Voltage

2.26

Central Wiper Parking 180°

Y (N)

Barometric Pressure (mmHg)

761.1

Conductivity Cell Constant

0.47

pH7 (mv)

pH10 (mv)

pH Slope

pH4 (mv)

Parameter	Temp (°C)	Sp. Cond. (mS/cm)	Depth (m)	ODO (% Sat.)	ODO (mg/l)	pH7	pH10
Standard	22.13	50.00	0.135	100.0	8.6		
Measured (clean sensors)	22.22	50.01	0.113	99.7	8.59		

Parameter	pH4 (check)	Turbidity Low (FNU)	Turbidity High (FNU)	Chl Low (µg/l)	Chl High (µg/l)	BGA-PE Low (µg/l)	BGA-PE High (µg/l)
Standard							
Measured (clean sensors)							

Any Probe Malfunctions?

Y (N)

File Name: C07222024

If Y, describe in the comments.

(Station name _YYYYMMDD)

Comments:

5-quality. Brush mooring upon retrieval

Recorded By

K. Villages

Computer Entry By / Date

Certified By / Date

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: Moe

RQ: ---

Project: BMAPCM/Clarkson Creek

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 7/3/24

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	KVillings	7/22/24	1012	20.0	762.6	9.1	9.12	100.4	—	1.07	(P) F	(L) F
ICV	↓	↓	1015	20.2	762.6	9.1	9.10	100.4	—	↓	(P) F	(L) F
CCV	↓	6/15/24	1555	22.8	761.1	8.6	8.59	100.1	—	↓	(P) F	(L) F
CCV											P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	KVillings	7/22/24	1020	231103C	11/24	50000	50000	(P) F	(L) F
ICV	↓	↓	1025	231110	11/24	15000	14879	(P) F	(L) F
CCV	↓	6/15/24	1605	231103C	11/24	50000	50014	(P) F	(L) F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.						7.				P / F	L / F
Calibr.						4.				P / F	L / F
Calibr.						10.				P / F	L / F
ICV										P / F	L / F
CCV										P / F	L / F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 $\pm$ 50; mV pH 4 Range +180 $\pm$ 50; mV pH 10 Range -180 $\pm$ 50;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 7/3/24

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	KVillings	7/22/24	1030	0.135	0.135	(P) F	(L) F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: