



Certificate of Analysis

14325 60th St N · Clearwater, FL 33760

(727) 524-7732

www.exaxol.com · info@exaxol.com

Product: Conductivity Standard 10,000 μ -mhos

Issue Date: 10/26/2020

Item Number: C1065-10000

Lot #: 201026B

Expires: (1 yr) 11/2021

Product Description and Intended Use: For calibration of analytical instruments and/or validation of analytical methods

Actual Sample Value and Associated Uncertainty:

10,000 μ mhos $\pm$ 0.5% (9,950 μ mhos - 10,050 μ mhos)*. Conductivity values change according to temperature as follows:

$^{\circ}$ C	20	21	22	23	24	25	26	27	28	29	30
μ mhos	9075	9260	9445	9630	9815	10000	10185	10370	10555	10740	10925

*Testing performed on-site at Exaxol's QC laboratory using in-house procedures for all standardized solutions. These results were obtained in an environment of $25^{\circ}\text{C} \pm 1^{\circ}\text{C}$ with an RH of 40%-60% and relate only to the item specified on this certificate.

Homogeneity: Solutions are mixed according to in-house procedures and are guaranteed to be homogeneous.

Storage and Handling: Store in an ambient environment (15°C - 30°C). Keep product capped to avoid evaporation, which could alter the certified value.

NIST Traceability and Measurement Procedure for Defined Measurands:

Traceable to NIST SRM 999b and NIST Class S weights, calibrated via Test #FL-14057, according to NIST Handbook 145: Standard Operating Procedures Number 5: 3-1 Weight Design. Environmental conditions (temperature and humidity) is monitored with a NIST traceable thermometer and hygrometer. Exaxol Reference Materials are also Traceable to NIST Certified SRM's through an unbroken chain of comparisons.

Per ASTM E288 requirements, Exaxol uses strictly Class A pipettes and volumetric glassware, maintained in accordance with ASTM E542-01. Exaxol calibrates their balances using certified Class S NIST weights certified by the State of Florida Metrology Laboratory. Calibration records are maintained on site. All glassware and pipettes are also pre-calibrated and operated using the guidelines set forth by ASTM E542-01.

Exaxol is compliant with ISO 17025 and ISO 17034 standards. Exaxol solutions and standards are made in accordance with Exaxol's documented ISO 9001:2015 Quality Management System. Production records and test results are maintained and secured in the laboratory's Manufacturing Log, according to lot number.

Joseph E Papa, Sr. Chemist