

Certificate of Analysis List

For request number 1334742

Catalog Number Entered	Lot Number Entered	Related Catalog Number	Related Lot Code	Description	
2660501 NTU	0346	2971205	0357A	STABLCAL [®] FORMAZIN STANDARD	800
2684801 NTU	0336	2971205	0357A	STABLCAL [®] FORMAZIN STANDARD	20
2684801 NTU	0338	2971205	0357A	STABLCAL [®] FORMAZIN STANDARD	20
2684901 NTU	0335	2971205	0357A	STABLCAL [®] FORMAZIN STANDARD	100
2684901 NTU	0336	2971205	0357A	STABLCAL [®] FORMAZIN STANDARD	100
2684901 NTU	0342	2971205	0357A	STABLCAL [®] FORMAZIN STANDARD	100
2961801	0339	2971205	0357A	10 NTU Verification Standard	

Total Enclosures: 7

***Certificate of Analysis******This is a Component of 2971205 lot A0357A***

Page 1

COMMODITY: **STABLCAL® FORMAZIN STANDARD 800 NTU**COMMODITY NUMBER: **2660501**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A0346****12/22/2020****1/7/2021**

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	760 to 840 NTU	821.0 NTU

The expiration date is Mar 2022

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

Certified by _____

Scott Als
Analytical Services Chemist

***Certificate of Analysis******This is a Component of 2971205 lot A0357A***

Page 1

COMMODITY: **STABLCAL® FORMAZIN STANDARD 20 NTU**COMMODITY NUMBER: **2684801**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A0336****12/21/2020****12/29/2020**

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	19 to 21 NTU	20.4 NTU

The expiration date is Mar 2022

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

Certified by _____

Scott Als
Analytical Services Chemist

***Certificate of Analysis******This is a Component of 2971205 lot A0357A***

Page 1

COMMODITY: **STABLCAL® FORMAZIN STANDARD 20 NTU**COMMODITY NUMBER: **2684801**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A0338****12/28/2020****12/31/2020**

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	19 to 21 NTU	20.4 NTU

The expiration date is Mar 2022

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

Certified by _____

Scott Als
Analytical Services Chemist

***Certificate of Analysis******This is a Component of 2971205 lot A0357A***

Page 1

COMMODITY: **STABLCAL® FORMAZIN STANDARD 100 NTU**COMMODITY NUMBER: **2684901**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A0335****12/28/2020****1/4/2021**

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	95 to 105 NTU	103.0 NTU

The expiration date is Mar 2022

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

Certified by _____

Scott Als
Analytical Services Chemist

***Certificate of Analysis******This is a Component of 2971205 lot A0357A***

Page 1

COMMODITY: **STABLCAL® FORMAZIN STANDARD 100 NTU**COMMODITY NUMBER: **2684901**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A0336****12/21/2020****12/28/2020**

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	95 to 105 NTU	102.0 NTU

The expiration date is Mar 2022

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

Certified by _____

Scott Als
Analytical Services Chemist

***Certificate of Analysis******This is a Component of 2971205 lot A0357A***

Page 1

COMMODITY: **STABLCAL® FORMAZIN STANDARD 100 NTU**COMMODITY NUMBER: **2684901**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A0342****12/30/2020****1/4/2021**

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	95 to 105 NTU	102.0 NTU

The expiration date is Mar 2022

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

Certified by _____

Scott Als
Analytical Services Chemist

***Certificate of Analysis******This is a Component of 2971205 lot A0357A***

Page 1

COMMODITY: **10 NTU Verification Standard**COMMODITY NUMBER: **2961801**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A0339****12/30/2020****1/5/2021**

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	9.5 to 10.5 NTU	10.20 NTU

The expiration date is Mar 2022

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

Certified by _____

Scott Als
Analytical Services Chemist