

Certificate of Analysis List

For request number 1937785

Catalog Number Entered	Lot Number Entered	Related Catalog Number	Related Lot Code	Description
2684801 STANDARD	4011 20 NTU	N/A	N/A	STABLCAL sup TS sup0 FORMAZIN

Total Enclosures: 1

***Certificate of Analysis***

Page 1

COMMODITY: **STABLCAL|sup|TS|sup0 FORMAZIN STANDARD 20 NTU**
COMMODITY NUMBER: **2684801** MANUFACTURE DATE: DATE OF ANALYSIS:
LOT NUMBER: **A4011** **1/26/2024** **2/2/2024**

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	19 to 21 NTU	20.1 NTU

The expiration date is Apr 2025

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

Certified by _____

Scott Als
Analytical Services Chemist