

# Certificate of Analysis

**Inner Solution Concentrate**  
**Private labeled for ECI****Lot Number:** 2207G05**Product Number:** TAU325**Manufacture Date:** JUL 26, 2022**Expiration Date:** JUL 2024

This product is formulated according to ECI specifications and tested to verify conformance to ECI specifications.

| Name         | CAS#        | Grade            |
|--------------|-------------|------------------|
| Trade Secret | Proprietary | Commercial Grade |

| Test                | Specification    | Result |
|---------------------|------------------|--------|
| Appearance          | Colorless liquid | Passed |
| Suitability for Use | Passes Test      | Passed |

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

| Part Number | Size / Package Type       | Shelf Life (Unopened Container) |
|-------------|---------------------------|---------------------------------|
| TAU325      | 50 mL natural LDPE disper | 24 months                       |

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)

Myrlande Gilles (07/26/2022)

Quality Control

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.