

EXAXOL

manufacturers of laboratory solutions/standards since 1988

CERTIFICATE OF ANALYSIS

Product: Conductivity Standard, 1,000 μ -mhos (C1065-1000)

Lot: 180522A

Expires: (1 yr) 06/2019

Assay (ppm, M, N, %, etc)

Actual: 1,000

Contains: Potassium chloride, water: both ACS

Other results: 1000 μ -mhos @ 25° C

TDS= 525.0 ppm KCl

Made in accordance with written standard Good Laboratory Practice Procedures. Volumetric glassware used is strictly Class A per ASTM E288 requirements, maintained in accordance with ASTM E 542. Analytical balance(s) calibrated periodically using certified Class S NIST weights, and documented. Thermometers used are NIST calibrated. Standardized products prepared according to Master Lab Record book, with lot #, batch size, customer, and testing results.

STORAGE: ambient (15 degC-30 degC)

Standards used:

NIST: Potassium Chloride KCl 3193

Exaxol:

Joseph E Papa, Sr. Chemist

EXAXOL CHEMICAL CORP, 14325 60th St N, Clearwater, FL 33760

Ph: (727) 524-7732 Fax: (727) 532-8221