



Certificate of Analysis

1 Reagent Lane
Fair Lawn, NJ 07410
201.796.7100 tel
201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System
Standard ISO9001:2015 by SAI Global Certificate Number CERT – 0120632

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	SB101	Quality Test / Release Date	05/31/2019
Lot Number	193247	Expiration Date	May/2021
Description	BUFFER SOLUTION, PH 4.00, CERTIFIED		
Country of Origin	United States		
Chemical Origin	Organic - non animal		
BSE/TSE Comment	No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product.		
Chemical Comment	This material does not contain any potential allergens, including peanuts, milk, wheat, eggs, gluten, or soybean and does not contain any genetically modified organisms (GMO's)		

N/A			
Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	Clear, red liquid
IDENTIFICATION	PASS/FAIL	= PASS TEST	PASS TEST
NIST STD USED: pH 7a	POT DIHYDRO PHOS	= LOT 186IG	LOT 186IG
NIST STD USED: pH 7b	DISOD HYDRO PHOS	= LOT 186IIG	LOT 186IIG
PH AT 25 DEGREES C		Inclusive Between 3.99 - 4.01	4.00
NIST STD USED: pH 4	POT HYDROGEN PHTHAL	= LOT 185I	LOT 185I
OPTICAL ABSORBANCE	PASS/FAIL	= PASS TEST	PASS TEST
TRACEABLE TO NIST	PASS/FAIL	= PASS TEST	PASS TEST

Jerusa Bailey-Wyche

Quality Assurance Specialist - Certificate of Analysis Fair Lawn

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above. If there are any questions with this certificate, please call at (800) 227-6701.

*Based on suggested storage condition.