

Chemical Analysis Report

BSSP-2018-07-17-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Status Lake LVI**
Request ID: **RQ-2018-07-16-48**
Customer: **BSSP**
Project ID: **ADDLAKES**

Send Reports to:
FL Dept. of Environmental Protection
2600 Blair Stone Rd.
Tallahassee, FL 32399-2400
Attn: Sarah Panek, MS 6511

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Liang Lin, Environmental Administrator

Date Certified: 02-AUG-2018 16:32

A handwritten signature in black ink, appearing to read "Liang Lin", is shown on a light-colored background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Lake Fontaine

Collection Date/Time: 07/17/2018 12:05

Field ID: Lake Fontaine

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2008168	SM 2120 B	Color (true)	30		PCU	P348609	
	SM 2320 B-97	Total Alkalinity	20		mg CaCO3/L	P348736	
2008169	EPA 350.1 Rev. 2.0	Ammonia-N	0.002	I	mg N/L	P348997	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.42		mg N/L	P348840	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P348759	
	EPA 365.1 Rev. 2.0	Total-P	0.022		mg P/L	P348924	
2008170	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P348600	

Quality Assurance Report Method Blank Results

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P348997

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P348840

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P348759

Component	Result	Code	Units
NO ₂ NO ₃ -N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P348924

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P348600

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P348609

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P348736

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P348997

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	104		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P348840

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	101		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P348759

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	99.5		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P348924

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	97.1		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P348600

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	97.1		P	90 - 110

Reference Method: SM 2120 B
Batch ID: P348609

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	99.7		P	85 - 115

Reference Method: SM 2320 B-97
Batch ID: P348736

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	102		P	98 - 107

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P348997

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2008095	Ammonia-N	101	103	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P348840

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2008169	Kjeldahl Nitrogen	95.8	97.5	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P348759

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2007832	NO2NO3-N	102	104	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P348924

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2008300	Total-P	98.3	96.0	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P348600

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2007893	O-Phosphate-P	95.2	93.8	P/P	90 - 110

Quality Assurance Report Precision

Reference Method: EPA 350.1 Rev. 2.0
 Batch ID: P348997

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2008095	Ammonia-N	1.60	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
 Batch ID: P348840

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2008169	Kjeldahl Nitrogen	1.29	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P348759

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2007832	NO2NO3-N	1.54	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
 Batch ID: P348924

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2008300	Total-P	2.17	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P348600

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2007893	O-Phosphate-P	1.48	Spike	P	0 - 20

Reference Method: SM 2120 B
 Batch ID: P348609

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2008050	Color (true)	0.0964	Sample	P	0 - 20

Reference Method: SM 2320 B-97
 Batch ID: P348736

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2007885	Total Alkalinity	0.368	Sample	P	0 - 1.6

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
 Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A85281

Included Lab Sample IDs: 2008170

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	97.3	100	P/P	90 - 110

Reference Method: SM 2120 B

Run ID: A85283

Included Lab Sample IDs: 2008168

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	102	103	P/P	90 - 115

Reference Method: SM 2320 B-97

Run ID: A85328

Included Lab Sample IDs: 2008168

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	99.8	104	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A85343

Included Lab Sample IDs: 2008169

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	99.5	99.5	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A85405

Included Lab Sample IDs: 2008169

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	92.6	98.5	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A85415

Included Lab Sample IDs: 2008169

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	101	101	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A85425

Included Lab Sample IDs: 2008169

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	100	100	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements. Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 350.1 Rev. 2.0	Ammonia-N	104	101	103			1.60
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	101	95.8	97.5			1.29
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	99.5	102	104			1.54
EPA 365.1 Rev. 2.0	Total-P	97.1	98.3	96.0			2.17
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	97.1	95.2	93.8			1.48
SM 2120 B	Color (true)	99.7				0.0964	
SM 2320 B-97	Total Alkalinity	102				0.368	

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	2008169
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	2008169
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	2008169
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	2008169
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2008170
SM 2120 B / E31780	True Color measured at 450 nm	2008168
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2008168

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 350.1 Rev. 2.0	07/17/2018			07/23/2018 12:24	Ping Hua	2008169
EPA 351.2 Rev. 2.0	07/17/2018	07/20/2018 11:00	Adrian Shelby	07/23/2018 10:41	Joseph Suarez	2008169
EPA 353.2 Rev. 2.0	07/17/2018			07/19/2018 11:19	Lauren Bottorf	2008169
EPA 365.1 Rev. 2.0	07/17/2018	07/23/2018 12:25	Sima Feizi	07/24/2018 10:53	Beverly Y. Sanders	2008169
EPA 365.1 Rev. 2.0 dissolved	07/17/2018			07/17/2018 16:05	Julia Storbeck	2008170
SM 2120 B	07/17/2018			07/17/2018 16:39	Tanya Welborn	2008168
SM 2320 B-97	07/17/2018			07/19/2018 04:45	Dale L. Simmons	2008168