

Holds 297 mL liquid

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD4000)

Boldly "X" this box if there is qualified data on this page.

Meter YSI 556#1

RQ- 2016-05-02-29

Project Nassau West

- Notes: (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).
 (2) Always wait for meter to stabilize before recording any readings.
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site L.C.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 4/16

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	B. Davis	5/3/16	741	22.85	8.59	8.59	100	46.1		P	L
ICV			742	22.89	8.59	8.58	97.8	46.1		P	L
CCV			1346	22.8							
CCV			1359	22.8	8.74	8.74	103.4	49.2		P	L

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mho./cm	Pass / Fail	Lab / Field
Calibr.	B. Davis	5/1/16	744	160216A	2/17	1000	1000	P	L
ICV			745			1000	1000	P	L
CCV			1350			100	104		
CCV									

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	B. Davis	5/3/16	747	2509677	8/17	7.0	7.0		P	L
Calibr.			749	2509774	7/17	4.0	4.0		P	L
Calibr.			750	2509774	4/17	10.0	10.0		P	L
ICV			750	2509774	2/17	7.0	7.05	-59.4	P	L
CCV		5/3/16		2509677	8/17					
CCV			1352	2509774	7/17	4.0	4.05	117.3	P	L

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU; mV pH7 Range 0 $\pm$ 50;

mV pH 4 Range +180 $\pm$ 50;

mV pH 10 Range -180 $\pm$ 50;

Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly)

Date of Last Depth Verification 4/16

Depth Sensor (Daily)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air	B. Davis	5/3/16	800	0.000	0.00	P	L

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater