

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page.

Meter #3 RQ- 2016-10-24-17 Project SWW3/Tumblah SCI

- Notes: (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).
 (2) Always wait for meter to stabilize before recording any readings.
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 03/04/2016 7/2016

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	J. Parrish	10/31/16	718	23.3	8.53	8.53	103	59.4	-	P	L
ICV	2	2	719	23.4	8.53	8.44	99.7	58.4	-	P	L
CCV											
CCV	N. Frederik	10/31/16	1530	23.58	8.49	8.46	99.7	59.4	-	P	L

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	Stann	10/31/16	724	150615A	5/17	1000	1000	P	L
ICV	2	2	724	05151413	10/16	1413	1403	P	L
CCV									
CCV	N. Frederik	10/31/16	1535	160616B	12/16	100	102	P	L

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Stann	10/31/16	0730	250604	6/18	7.0	7.04	-	P	L
Calibr.				2602553	1/18	4.0				
Calibr.				2510A88	4/17	10.0				
ICV										
CCV	N. Frederik	10/31/16	1540	2602553	1/18	4.0	4.02	-	P	L
CCV										

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU; mV pH7 Range 0 $\pm$ 50;

mV pH 4 Range +180 $\pm$ 50;

mV pH 10 Range -180 $\pm$ 50;

Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly)

Date of Last Depth Verification

Depth Sensor (Daily)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air	Stann	10/31/16	730	0.000	0.00	P	

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater