

Chemical Analysis Report

NE-DIST-2016-08-10-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Nassau Mid West**
Request ID: **RQ-2016-08-08-16**
Customer: **NE-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
8800 Baymeadows Way West
Jacksonville, FL 32256-7577
Attn: Jeremy Parrish

For additional information please contact
Timothy W. Fitzpatrick
Liang-Tsair Lin, Ph.D.
Colin Wright, Ph.D.
Thekkekalathil Chandrasekhar, PhD, QA Officer
Phone (850) 245-8085

Certified by: Colin Wright, Environmental Administrator

Date Certified: 26-AUG-2016 14:00

A handwritten signature in black ink, appearing to read "Colin Wright", is written over a light blue rectangular background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: STMARY SPRONG 23C

Collection Date/Time: 08/09/2016 09:55

Field ID: 19010023

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1822928	EPA 180.1 Rev. 2.0	Turbidity	3.4	A	NTU	P309512	
	EPA 300.0 Rev. 2.1	Chloride	17	A	mg Cl/L	P309940	
		Sulfate	6.5	A	mg SO4/L	P309943	
	SM 2120 B	Color (true)	73		PCU	P309523	
	SM 2320 B-97	Total Alkalinity	45		mg CaCO3/L	P309623	
	SM 2540 C-97	TDS	112		mg/L	P310094	
	SM 2540 D-97	TSS	3	I	mg/L	P309998	
	SM 4500 F-C-97	Fluoride	0.12		mg F/L	P309626	
1822931	EPA 350.1 Rev. 2.0	Ammonia-N	0.015		mg N/L	P309589	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.32		mg N/L	P309785	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.39		mg N/L	P309614	
	EPA 365.1 Rev. 2.0	Total-P	0.22		mg P/L	P309792	
	SM 5310 B-00	Organic Carbon	6.5		mg C/L	P310047	
1822934	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.16		mg P/L	P309518	

Sample Location: CALKIN TRIB HAMP REG.

Collection Date/Time: 08/09/2016 10:30

Field ID: 19010042

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1822929	EPA 180.1 Rev. 2.0	Turbidity	3.8		NTU	P309512	
	EPA 300.0 Rev. 2.1	Chloride	8.0		mg Cl/L	P309940	
		Sulfate	1.4		mg SO4/L	P309943	
	SM 2120 B	Color (true)	570		PCU	P309524	
	SM 2320 B-97	Total Alkalinity	1.1	I	mg CaCO3/L	P309623	
	SM 2540 C-97	TDS	115		mg/L	P310094	
	SM 2540 D-97	TSS	3	I	mg/L	P309998	
	SM 4500 F-C-97	Fluoride	0.045	I	mg F/L	P309626	
1822932	EPA 350.1 Rev. 2.0	Ammonia-N	0.027		mg N/L	P309589	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.2		mg N/L	P309785	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.009	I	mg N/L	P310020	
	EPA 365.1 Rev. 2.0	Total-P	0.030		mg P/L	P309792	
	SM 5310 B-00	Organic Carbon	58		mg C/L	P310047	
1822935	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.013		mg P/L	P309518	

Sample Location: STMARY SPRONG90

Collection Date/Time: 08/09/2016 11:00

Field ID: 19010021

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1822930	EPA 180.1 Rev. 2.0	Turbidity	4.1		NTU	P309512	
	EPA 300.0 Rev. 2.1	Chloride	22		mg Cl/L	P309940	
		Sulfate	10		mg SO4/L	P309943	
	SM 2120 B	Color (true)	87		PCU	P309524	
	SM 2320 B-97	Total Alkalinity	47		mg CaCO3/L	P309623	

Field ID: 19010021

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1822930	SM 2540 C-97	TDS	140		mg/L	P310094	
	SM 2540 D-97	TSS	3	I	mg/L	P309998	
	SM 4500 F-C-97	Fluoride	0.19		mg F/L	P309626	
1822933	EPA 350.1 Rev. 2.0	Ammonia-N	0.025		mg N/L	P309589	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.44		mg N/L	P309785	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.48		mg N/L	P309614	
	EPA 365.1 Rev. 2.0	Total-P	0.52		mg P/L	P309792	
	SM 5310 B-00	Organic Carbon	8.3		mg C/L	P310048	
1822936	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.45		mg P/L	P309518	

Sample Location: EQUIPMENT BLANK

Collection Date/Time: 08/09/2016 10:03

Field ID: BLANK

Matrix: W-EQPMT-BK

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1822937	EPA 300.0 Rev. 2.1	Chloride	0.20	U	mg Cl/L	P309940	
		Sulfate	0.20	U	mg SO4/L	P309943	
	SM 2540 C-97	TDS	15	U	mg/L	P310093	
1822938	SM 2540 D-97	TSS	2	U	mg/L	P309997	
	SM 4500 F-C-97	Fluoride	0.035	U	mg F/L	P309626	
	EPA 350.1 Rev. 2.0	Ammonia-N	0.002	U	mg N/L	P309589	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.080	U	mg N/L	P309785	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P309614	
1822939	EPA 365.1 Rev. 2.0	Total-P	0.002	U	mg P/L	P309792	
	SM 5310 B-00	Organic Carbon	0.50	U	mg C/L	P310048	
	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P309518	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P309512

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P309940

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P309943

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P309589

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P309785

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P309614

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P310020

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P309792

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P309518

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P309523

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2120 B
Batch ID: P309524

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P309623

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Quality Assurance Report Method Blank Results

Reference Method: SM 2540 C-97
Batch ID: P310093

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 C-97
Batch ID: P310094

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P309997

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P309998

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P309626

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P310047

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Reference Method: SM 5310 B-00
Batch ID: P310048

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P309940

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	97.7		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P309943

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
-----------	---------	---------	-----------	----------------

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P309943

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	94.3		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P309589

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	96.6		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P309785

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	107		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P309614

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	103		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P310020

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	101		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P309792

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	98.2		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P309518

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	104		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P309623

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	103		P	95 - 108

Reference Method: SM 4500 F-C-97
Batch ID: P309626

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	97.4		P	92 - 103

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 5310 B-00
Batch ID: P310047

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	96.0		P	85 - 115

Reference Method: SM 5310 B-00
Batch ID: P310048

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	97.5		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P309940

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1822928	Chloride	96.2		P	90 - 110
1822955	Chloride	96.8		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P309943

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1822928	Sulfate	93.8		P	90 - 110
1822955	Sulfate	93.8		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P309589

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1822924	Ammonia-N	94.8	95.8	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P309785

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1822763	Kjeldahl Nitrogen	105	102	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P309614

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1822924	NO2NO3-N	92.8	94.8	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P310020

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1823209	NO2NO3-N	96.0	97.0	P/P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P309792

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1822931	Total-P	99.4	101	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P309518

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1822819	O-Phosphate-P	100	102	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P309626

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1822753	Fluoride	98.9	99.6	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P310047

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1822503	Organic Carbon	102		P	85 - 115

Reference Method: SM 5310 B-00
Batch ID: P310048

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1822933	Organic Carbon	101	102	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P309512

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1822928	Turbidity	1.19	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P309940

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1822928	Chloride	0.0429	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P309943

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1822928	Sulfate	0.733	Sample	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P309589

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1822924	Ammonia-N	0.982	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P309785

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1822763	Kjeldahl Nitrogen	1.98	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P309614

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1822924	NO2NO3-N	2.02	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P310020

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1823209	NO2NO3-N	0.608	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P309792

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1822931	Total-P	1.12	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P309518

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1822819	O-Phosphate-P	1.27	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P309523

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1822753	Color (true)	1.31	Sample	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P309524

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1822973	Color (true)	0.270	Sample	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2320 B-97
Batch ID: P309623

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1822753	Total Alkalinity	0.684	Sample	P	0 - 4.8

Reference Method: SM 2540 C-97
Batch ID: P310093

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1822753	TDS	4.19	Sample	P	0 - 10

Reference Method: SM 2540 C-97
Batch ID: P310094

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1822817	TDS	5.54	Sample	P	0 - 10

Reference Method: SM 2540 D-97
Batch ID: P309998

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1822817	TSS	5.71	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P309626

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1822753	Fluoride	0.463	Spike	P	0 - 3.0

Reference Method: SM 5310 B-00
Batch ID: P310047

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1822503	Organic Carbon	0.0967	Spike	P	0 - 20

Reference Method: SM 5310 B-00
Batch ID: P310048

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1822933	Organic Carbon	1.05	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Quality Assurance Report Calibration Verification

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A71054

Included Lab Sample IDs: 1822928, 1822929, 1822930

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.5	100	P/P	95 - 105

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A71057

Included Lab Sample IDs: 1822934, 1822935, 1822936, 1822939

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	100	99.3	P/P	90 - 110
O-Phosphate-P	101	99.8	P/P	90 - 110
O-Phosphate-P	101	100	P/P	90 - 110
O-Phosphate-P	99.8	100	P/P	90 - 110

Reference Method: SM 2120 B

Run ID: A71058

Included Lab Sample IDs: 1822928, 1822929, 1822930

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	97.8	98.0	P/P	90 - 115
Color (true)	98.0	97.8	P/P	90 - 115
Color (true)	99.4	98.7	P/P	90 - 115

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A71077

Included Lab Sample IDs: 1822931, 1822933, 1822938

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	100	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A71081

Included Lab Sample IDs: 1822931, 1822932, 1822933, 1822938

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	97.5	97.3	P/P	90 - 110
Ammonia-N	98.0	97.5	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A71090

Included Lab Sample IDs: 1822928, 1822929, 1822930

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	103	100	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A71090

Included Lab Sample IDs: 1822928, 1822929, 1822930, 1822937

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	99.5	100	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A71177

Included Lab Sample IDs: 1822928, 1822929, 1822930, 1822937

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	101	99.8	P/P	90 - 110
Chloride	99.8	101	P/P	90 - 110
Sulfate	96.2	96.5	P/P	90 - 110
Sulfate	98.4	96.2	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A71179

Included Lab Sample IDs: 1822931, 1822933

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	101	99.7	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A71192

Included Lab Sample IDs: 1822931, 1822932, 1822933, 1822938

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	100	104	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A71219

Included Lab Sample IDs: 1822932, 1822938

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	98.5	99.0	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A71229

Included Lab Sample IDs: 1822932

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	99.5	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A71236

Included Lab Sample IDs: 1822931, 1822932, 1822933, 1822938

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	96.1	97.1	P/P	90 - 110
Organic Carbon	97.1	97.2	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery	Precision	MS
				LCS	SMP

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		Precision	
					LCS	MS
EPA 180.1 Rev. 2.0	Turbidity				1.19	
EPA 300.0 Rev. 2.1	Chloride	97.7	96.2	96.8	0.0429	
	Sulfate	94.3	93.8	93.8	0.733	
EPA 350.1 Rev. 2.0	Ammonia-N	96.6	94.8	95.8		0.982
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	107	105	102		1.98
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	103	92.8	94.8		2.02
	NO ₂ NO ₃ -N	101	96.0	97.0		0.608
EPA 365.1 Rev. 2.0	Total-P	98.2	99.4	101		1.12
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	104	100	102		1.27
SM 2120 B	Color (true)				1.31	
	Color (true)				0.270	
SM 2320 B-97	Total Alkalinity	103			0.684	
SM 2540 C-97	TDS				4.19	
	TDS				5.54	
SM 2540 D-97	TSS				5.71	
SM 4500 F-C-97	Fluoride	97.4	98.9	99.6		0.463
SM 5310 B-00	Organic Carbon	96.0	102			0.0967
	Organic Carbon	97.5	101	102		1.05

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1822928, 1822929, 1822930
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1822928, 1822929, 1822930, 1822937
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	1822928, 1822929, 1822930, 1822937
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1822931, 1822932, 1822933, 1822938
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1822931, 1822932, 1822933, 1822938
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1822931, 1822932, 1822933, 1822938
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1822931, 1822932, 1822933, 1822938
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1822934, 1822935, 1822936, 1822939
SM 2120 B / E31780	True Color measured at 450 nm	1822928, 1822929, 1822930
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	1822928, 1822929, 1822930
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1822928, 1822929, 1822930, 1822937
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1822928, 1822929, 1822930, 1822937
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1822928, 1822929, 1822930, 1822937
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1822931, 1822932, 1822933, 1822938

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	08/10/2016			08/10/2016 14:56	Sima Feizi	1822928, 1822929, 1822930
EPA 300.0 Rev. 2.1	08/10/2016			08/17/2016	Ping Hua	1822928, 1822929, 1822930, 1822937
EPA 350.1 Rev. 2.0	08/10/2016			08/11/2016	Joseph Suarez	1822931, 1822932, 1822933, 1822938
EPA 351.2 Rev. 2.0	08/10/2016	08/16/2016	Sofia Elnemr	08/17/2016	Christopher Armour	1822931, 1822932, 1822933, 1822938
EPA 353.2 Rev. 2.0	08/10/2016			08/11/2016	Beverly Y. Sanders	1822931, 1822933, 1822938
	08/10/2016			08/19/2016	Beverly Y. Sanders	1822932
EPA 365.1 Rev. 2.0	08/10/2016	08/16/2016	Vijayalakshmi Reddy	08/17/2016	Lipi Saha	1822931, 1822933
	08/10/2016	08/16/2016	Vijayalakshmi Reddy	08/19/2016	Lipi Saha	1822932, 1822938
EPA 365.1 Rev. 2.0 dissolved	08/10/2016			08/10/2016 16:07	Julia Storbeck	1822939
	08/10/2016			08/10/2016 16:27	Julia Storbeck	1822934
	08/10/2016			08/10/2016 16:41	Julia Storbeck	1822935
	08/10/2016			08/10/2016 16:51	Julia Storbeck	1822936
SM 2120 B	08/10/2016			08/10/2016 17:25	Rochelle Y Jackson	1822928
	08/10/2016			08/10/2016 17:29	Rochelle Y Jackson	1822930
	08/10/2016			08/10/2016 17:57	Rochelle Y Jackson	1822929
SM 2320 B-97	08/10/2016			08/11/2016	Dale L. Simmons	1822928, 1822929, 1822930
SM 2540 C-97	08/10/2016			08/11/2016	Yijie Li	1822928, 1822929, 1822930, 1822937
SM 2540 D-97	08/10/2016			08/11/2016	Yijie Li	1822928, 1822929, 1822930, 1822937
SM 4500 F-C-97	08/10/2016			08/11/2016	Dale L. Simmons	1822928, 1822929, 1822930, 1822937
SM 5310 B-00	08/10/2016			08/19/2016	Sofia Elnemr	1822931, 1822932, 1822933, 1822938