

Chemical Analysis Report

NE-DIST-2016-10-12-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780


Event Description: **Alachua Lakes**
Request ID: **RQ-2016-10-24-14**
Customer: **NE-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
8800 Baymeadows Way
Suite 100
Jacksonville, FL 32256
Attn: Nicole Frederick

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Thekkekalathil Chandrasekhar, PhD, QA Officer
Phone (850) 245-8085

Certified by: Liang Lin, Environmental Administrator

Date Certified: 04-NOV-2016 10:46

A handwritten signature in black ink, appearing to read "Liang Lin", is written on a light-colored, textured background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Moon Lake

Collection Date/Time: 10/11/2016 12:40

Field ID: G1NE0197 10112016

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1840176	EPA 180.1 Rev. 2.0	Turbidity	1.4		NTU	P313040	
	EPA 300.0 Rev. 2.1	Chloride	14		mg Cl/L	P313677	
		Sulfate	13		mg SO4/L	P313680	
	SM 2120 B	Color (true)	39		PCU	P313028	
	SM 2320 B-97	Total Alkalinity	26		mg CaCO3/L	P313458	
	SM 2540 C-97	TDS	84		mg/L	P313716	
	SM 2540 D-97	TSS	2	I	mg/L	P313837	
	SM 4500 F-C-97	Fluoride	0.24		mg F/L	P313462	
1840179	EPA 350.1 Rev. 2.0	Ammonia-N	0.54		mg N/L	P314071	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.3		mg N/L	P314205	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.046		mg N/L	P313379	
	EPA 365.1 Rev. 2.0	Total-P	0.25		mg P/L	P313923	
	SM 5310 B-00	Organic Carbon	9.0		mg C/L	P313264	
1840182	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.19		mg P/L	P313035	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 351.2 Rev. 2.0: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

EPA 365.1 Rev. 2.0 dissolved: O-Phosphate-P: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Sample Location: Lake Wauberg at Center

Collection Date/Time: 10/11/2016 10:52

Field ID: G1NE0114 10112016

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1840177	EPA 180.1 Rev. 2.0	Turbidity	65	A	NTU	P313040	
	EPA 300.0 Rev. 2.1	Chloride	13		mg Cl/L	P313677	
		Sulfate	0.20	U	mg SO4/L	P313680	
	SM 2120 B	Color (true)	16		PCU	P313028	
	SM 2320 B-97	Total Alkalinity	21		mg CaCO3/L	P313458	
	SM 2540 C-97	TDS	79		mg/L	P313716	
	SM 2540 D-97	TSS	64		mg/L	P313837	
	SM 4500 F-C-97	Fluoride	0.16		mg F/L	P313462	
1840180	EPA 350.1 Rev. 2.0	Ammonia-N	0.01		mg N/L	P314071	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	4.5		mg N/L	P314207	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P313379	
	EPA 365.1 Rev. 2.0	Total-P	0.22		mg P/L	P313923	
	SM 5310 B-00	Organic Carbon	17		mg C/L	P313264	
1840183	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.006	I	mg P/L	P313035	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 351.2 Rev. 2.0: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

EPA 365.1 Rev. 2.0 dissolved: O-Phosphate-P: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Sample Location: Hammock Lake

Collection Date/Time: 10/11/2016 13:25

Field ID: G1NE0198 10112016

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1840178	EPA 180.1 Rev. 2.0	Turbidity	18		NTU	P313040	
	EPA 300.0 Rev. 2.1	Chloride	34		mg Cl/L	P313677	
		Sulfate	6.1		mg SO4/L	P313680	
	SM 2120 B	Color (true)	30		PCU	P313028	
	SM 2320 B-97	Total Alkalinity	83		mg CaCO3/L	P313458	
	SM 2540 C-97	TDS	149		mg/L	P313716	
	SM 2540 D-97	TSS	20	I	mg/L	P313837	
	SM 4500 F-C-97	Fluoride	0.28		mg F/L	P313462	
1840181	EPA 350.1 Rev. 2.0	Ammonia-N	0.009		mg N/L	P314071	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	2.0		mg N/L	P314205	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P313379	
	EPA 365.1 Rev. 2.0	Total-P	0.18		mg P/L	P313923	
	SM 5310 B-00	Organic Carbon	11		mg C/L	P313264	
1840184	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.01	I	mg P/L	P313035	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample. The PQL was elevated due to sample matrix interference.

EPA 351.2 Rev. 2.0: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

EPA 365.1 Rev. 2.0 dissolved: O-Phosphate-P: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P313040

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P313677

Component	Result	Code	Units
Chloride	0.20	I	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P313680

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P314071

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P314205

Component	Result	Code	Units
-----------	--------	------	-------

Quality Assurance Report Method Blank Results

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P314205

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P314207

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P313379

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P313923

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P313035

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P313028

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P313458

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P313716

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P313837

Component	Result	Code	Units
TSS	2	U	mg/L

Quality Assurance Report Method Blank Results

Reference Method: SM 4500 F-C-97
Batch ID: P313462

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P313264

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P313677

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	102		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P313680

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	97.5		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P314071

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	104		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P314205

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	103		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P314207

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	105		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P313379

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO ₂ NO ₃ -N	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P313923

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
-----------	---------	---------	-----------	----------------

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P313923

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	97.2		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P313035

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	101		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P313458

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	103		P	95 - 108

Reference Method: SM 4500 F-C-97
Batch ID: P313462

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	97.3		P	92 - 103

Reference Method: SM 5310 B-00
Batch ID: P313264

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	96.2		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P313677

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1840069	Chloride	102		P	90 - 110
1840176	Chloride	101		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P313680

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1840069	Sulfate	97.7		P	90 - 110
1840176	Sulfate	96.0		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P314071

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1840134	Ammonia-N	101	102	P/P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P313379

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1840179	NO2NO3-N	104	103	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P313923

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1840179	Total-P	99.6	97.8	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P313462

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1840154	Fluoride	99.5	96.9	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P313264

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1840080	Organic Carbon	102		P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P313040

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1840069	Turbidity	1.65	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P313677

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1840069	Chloride	0.503	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P313680

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1840069	Sulfate	3.14	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P314071

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1840134	Ammonia-N	1.49	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P314205

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1840179	Kjeldahl Nitrogen	1.93	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P314207

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1839367	Kjeldahl Nitrogen	9.83	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P313379

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1840179	NO2NO3-N	0.791	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P313923

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1840179	Total-P	1.11	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P313035

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1840182	O-Phosphate-P	0.558	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P313028

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1840155	Color (true)	5.32	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P313458

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1840154	Total Alkalinity	2.03	Sample	P	0 - 4.8

Reference Method: SM 2540 C-97
Batch ID: P313716

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1839526	TDS	3.01	Sample	P	0 - 10

Quality Assurance Report Precision

Reference Method: SM 4500 F-C-97
Batch ID: P313462

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1840154	Fluoride	2.45	Spike	P	0 - 3.0

Reference Method: SM 5310 B-00
Batch ID: P313264

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1840080	Organic Carbon	1.21	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: SM 2120 B
Run ID: A72183
Included Lab Sample IDs: 1840176, 1840177, 1840178

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	102	102	P/P	85 - 115

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Run ID: A72185
Included Lab Sample IDs: 1840182, 1840183, 1840184

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	99.6	99.7	P/P	90 - 110
O-Phosphate-P	99.9	97.8	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0
Run ID: A72186
Included Lab Sample IDs: 1840176, 1840177, 1840178

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	100	P/P	95 - 105

Reference Method: SM 5310 B-00
Run ID: A72250
Included Lab Sample IDs: 1840179, 1840180, 1840181

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	97.2	96.8	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Run ID: A72281
Included Lab Sample IDs: 1840179, 1840180, 1840181

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	101	100	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A72305

Included Lab Sample IDs: 1840176, 1840177, 1840178

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	101	102	P/P	90 - 110
Chloride	102	102	P/P	90 - 110
Sulfate	98.3	99.2	P/P	90 - 110
Sulfate	99.2	98.6	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A72309

Included Lab Sample IDs: 1840176, 1840177, 1840178

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	100	103	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A72309

Included Lab Sample IDs: 1840176, 1840177, 1840178

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	101	99.5	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A72474

Included Lab Sample IDs: 1840179, 1840180, 1840181

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	101	99.8	P/P	90 - 110
Total-P	99.8	99.9	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A72480

Included Lab Sample IDs: 1840179, 1840180, 1840181

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	100	99.0	P/P	90 - 110
Ammonia-N	99.7	100	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A72551

Included Lab Sample IDs: 1840179, 1840180, 1840181

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	103	104	P/P	90 - 110
Kjeldahl Nitrogen	105	103	P/P	90 - 110
Kjeldahl Nitrogen	99.3	101	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements. Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		Precision	
					LCS	MS
EPA 180.1 Rev. 2.0	Turbidity				1.65	
EPA 300.0 Rev. 2.1	Chloride	102	102	101	0.503	
	Sulfate	97.5	97.7	96.0	3.14	
EPA 350.1 Rev. 2.0	Ammonia-N	104	101	102		1.49
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	103				1.93
	Kjeldahl Nitrogen	105				9.83
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	102	104	103		0.791
EPA 365.1 Rev. 2.0	Total-P	97.2	99.6	97.8		1.11
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	101				0.558
SM 2120 B	Color (true)				5.32	
SM 2320 B-97	Total Alkalinity	103			2.03	
SM 2540 C-97	TDS				3.01	
SM 4500 F-C-97	Fluoride	97.3	99.5	96.9		2.45
SM 5310 B-00	Organic Carbon	96.2	102			1.21

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1840176, 1840177, 1840178
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1840176, 1840177, 1840178
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	1840176, 1840177, 1840178
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1840179, 1840180, 1840181
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1840179, 1840180, 1840181
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1840179, 1840180, 1840181
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1840179, 1840180, 1840181
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1840182, 1840183, 1840184
SM 2120 B / E31780	True Color measured at 450 nm	1840176, 1840177, 1840178
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	1840176, 1840177, 1840178
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1840176, 1840177, 1840178
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1840176, 1840177, 1840178
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1840176, 1840177, 1840178
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1840179, 1840180, 1840181

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	10/12/2016			10/12/2016 15:09	Sima Feizi	1840176, 1840177, 1840178
EPA 300.0 Rev. 2.1	10/12/2016			10/21/2016	Julia Storbeck	1840176, 1840177, 1840178
EPA 350.1 Rev. 2.0	10/12/2016			10/27/2016	Sofia Elnemr	1840179, 1840180, 1840181
EPA 351.2 Rev. 2.0	10/12/2016	10/31/2016	Adrian Shelby	11/01/2016	Joseph Suarez	1840179, 1840180, 1840181
EPA 353.2 Rev. 2.0	10/12/2016			10/18/2016	Beverly Y. Sanders	1840179, 1840180, 1840181
EPA 365.1 Rev. 2.0	10/12/2016	10/26/2016	Vijayalakshmi Reddy	10/27/2016	Lipi Saha	1840179, 1840180, 1840181
EPA 365.1 Rev. 2.0 dissolved	10/12/2016			10/12/2016 15:05	Julia Storbeck	1840182
	10/12/2016			10/12/2016 15:33	Julia Storbeck	1840183, 1840184
SM 2120 B	10/12/2016			10/12/2016 15:32	Jared I. Manley	1840176
	10/12/2016			10/12/2016 15:33	Jared I. Manley	1840177
	10/12/2016			10/12/2016 15:34	Jared I. Manley	1840178
SM 2320 B-97	10/12/2016			10/19/2016	Dale L. Simmons	1840176, 1840177, 1840178
SM 2540 C-97	10/12/2016			10/15/2016	Yijie Li	1840176, 1840177, 1840178
SM 2540 D-97	10/12/2016			10/15/2016	Yijie Li	1840176, 1840177, 1840178
SM 4500 F-C-97	10/12/2016			10/19/2016	Dale L. Simmons	1840176, 1840177, 1840178
SM 5310 B-00	10/12/2016			10/14/2016	Julie Michelle Frank	1840179, 1840180, 1840181